
Wheelchair seating —

Part 3:

Determination of static, impact, and repetitive load strengths for postural support devices

Sièges de fauteuils roulants —

*Partie 3: Détermination de la résistance aux charges statiques,
dynamiques et cycliques pour les dispositifs de maintien de la posture*



This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Principle	5
4.1 Overview	5
4.1.1 General	5
4.1.2 Repetitive load	5
4.1.3 Static load	5
4.1.4 Impact load	5
4.2 PSD Sample	5
5 Requirements	5
6 Test report requirements	6
Annex A (normative) Method for repetitive load testing of PSDs	9
Annex B (normative) Method for static load testing of PSDs	17
Annex C (normative) Method for impact testing of PSDs	28
Annex D (normative) Mounting fixtures for the testing of postural support devices	32
Annex E (normative) PSD Loading Pads	36
Bibliography	41

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 173, *Assistive products*, Subcommittee SC 1, *Wheelchairs*.

This third edition cancels and replaces the second edition (ISO 16840-3:2014), which has been technically revised.

The main changes are as follows:

- the structure of the document has been updated;
- test results have been added;
- pass/fail requirements have been established.

A list of all parts in the ISO 16840 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 16840-3:2022 incorporates the following corrections:

- Missing list items were added to [6.10](#) and [B.5.8](#);
- Text clarified in [A.5.3](#).

Introduction

Postural support devices (PSDs), constructed as additional components to wheelchair seating or as wheelchair seating in its own right, are widely available and used extensively by people with disabilities. The selection or prescription of the most appropriate PSD is intended to be, where appropriate, partially dependent on knowledge of the PSD's ability to withstand static, impact, and repetitive loads. This document specifies test methods and requirements for the strength of PSDs as manufactured, which are designed to identify likely points of failure by breaking, yielding, or loosening of components - it is not intended to predict the long-term useful life. The useful life of a device depends upon many variables of use, aging, and environment: the way in which it is installed, the frequency of use and forces to which it is subjected, abrasion points, vibration and fatigue, cleaning and periodic maintenance, and temperature, humidity and UV exposure.

The tests involve mounting the PSD to a rigid test fixture to simulate mounting on a wheelchair. Rigid test fixtures are used to provide a worst-case situation by minimizing shock-absorption that can come, for example, from the damping effects of flex in the wheelchair frame, and also to make these tests repeatable by removing the variable of wheelchair type. Repetitive, static, and impact loads are then applied, as appropriate, according to the type of PSD, to determine if the minimum strength requirements are met.

When a series of strength/impact tests are performed on a PSD, the same sample PSD is used throughout and the tests conducted in series, from least stringent [lowest forces] to most stringent [highest forces]. In this manner, the PSD will be subjected to lower forces, which would typically be more frequently encountered in daily use, before being subjected to the higher forces that pose a greater risk of failure. If the sample PSD fails in a less stringent test, there is generally no reason to conduct more stringent tests until the PSD has been redesigned. Individual tests can be conducted using a unique sample PSD for each test, but this will not provide the same level of assurance about performance.

Some of the tests represented in this document are derived from ISO 7176-8. Many of the pass/fail criteria, test principles and test equipment are the same for this document as for ISO 7176-8.

Parts of this document are continuing to be developed so that future revisions can include the results of work in the following areas:

- further development of the test forces based on clinical data in order to determine actual impact, static, and repetitive forces that PSDs are subjected to;
- the collection of further data on the most common failures experienced in actual use of PSDs is ongoing;
- addressing any additional unaddressed PSD testing needs, including gaps as currently identified in [Table 1](#).

Wheelchair seating —

Part 3:

Determination of static, impact, and repetitive load strengths for postural support devices

1 Scope

This document specifies requirements for static, impact, and repetitive load strengths for postural support devices (PSDs) with associated attachment hardware intended for use with an undefined wheelchair seating system. It specifies the test methods for determining whether the minimum performance requirements have been met to release a product into use. It also specifies requirements for disclosure of the test results. Not all tests apply to all PSDs.

This document includes sets of tests for these particular types of PSDs listed in [Table 1](#):

Table 1 — Index of tests

PSD	Repetitive	Static	Impact
Seat cushion supporting structure	X		X
Back cushion supporting structure	X	X	X
Pelvic positioning	X	X	
Anterior trunk support	X	X	
Lateral support		X	
Medial knee support		X	
Head support	X	X	
Lower arm positioning devices	^a	^a	
Foot support		^b	^b
NOTE 1 Seat Cushion Supporting Structure is the system upon which the seat cushion is mounted.			
NOTE 2 Back Cushion Supporting Structure is the system upon which the back support is mounted.			
^a Repetitive and static tests are under development for non-integrated, lower arm positioning devices. Arm support static test protocols are defined in ISO 7176-8 for supports which are integrated within the wheelchair.			
^b Static and impact foot support test protocols are defined in ISO 7176-8 for supports which are integrated within the wheelchair.			

This document is also applicable to other seating systems.

The test methods can be used to verify the manufacturer's claims that a product meets the requirements of this document. This document does not apply to PSDs that are designed to fail under certain static, dynamic, or repetitive loads.

This document does not apply to the strength of PSDs under crash conditions in a motor vehicle.

This document does not evaluate long-term useful life.

NOTE 1 ISO 16840-4 provides crash test methods and requirements for wheelchair seating systems when used as part of a wheelchair seat in a motor vehicle.

NOTE 2 For user masses greater or less than those specified in this document, appropriate extrapolation of test apparatus dimensions, mounting point separation, forces, etc. can be carried out, and the test parameters noted in the test report.

NOTE 3 Rigid surrogate test fixtures are utilized to provide a standardized test method, and consequently this document does not involve a test of a PSD on a particular wheelchair.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 898-7, *Mechanical properties of fasteners — Part 7: Torsional test and minimum torques for bolts and screws with nominal diameters 1 mm to 10 mm*

ISO 7176-8:2014, *Wheelchairs — Part 8: Requirements and test methods for static, impact and fatigue strengths*

ISO 7176-26, *Wheelchairs — Part 26: Vocabulary*

ISO 16840-2, *Wheelchair seating — Part 2: Determination of physical and mechanical characteristics of seat cushions intended to manage tissue integrity*

JIS K 7312, *Physical testing methods for molded products of thermosetting polyurethane elastomers*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7176-26 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1 active support surface

postural support device that is power assisted to change its position or support surface shape

EXAMPLE An alternating pressure seat cushion or an electronically operated back support surface that reclines.

3.2 integrated lateral trunk support

lateral support that has a continuous structure with a back support

Note 1 to entry: See [Figure 1](#).