

**Health informatics - Device interoperability -
Part 10415: Personal health device
communication - Device specialization -
Weighing scale (ISO/IEEE 11073- 10415:2022)**

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO/IEEE 11073-10415:2022 sisaldab Euroopa standardi EN ISO/IEEE 11073-10415:2022 ingliskeelset teksti.	This Estonian standard EVS-EN ISO/IEEE 11073-10415:2022 consists of the English text of the European standard EN ISO/IEEE 11073-10415:2022.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 21.12.2022.	Date of Availability of the European standard is 21.12.2022.
Standard on kättesaadav Eesti Standardimis-ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.40

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis-ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis-ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Health informatics - Device interoperability - Part 10415:
Personal health device communication - Device
specialization - Weighing scale (ISO/IEEE 11073-
10415:2022)

Informatique de santé - Interopérabilité des dispositifs
- Partie 10415: Communication entre dispositifs de
santé personnels - Spécialisation des dispositifs -
Plateau de balance (ISO/IEEE 11073-10415:2022)

Medizinische Informatik - Kommunikation von Geräten
für die persönliche Gesundheit - Teil 10415:
Gerätespezifikation - Waage (ISO/IEEE 11073-
10415:2022)

This European Standard was approved by CEN on 4 December 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO/IEEE 11073-10415:2022) has been prepared by Technical Committee ISO/TC 215 "Health informatics" in collaboration with Technical Committee CEN/TC 251 "Health informatics" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2023, and conflicting national standards shall be withdrawn at the latest by June 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11073-10415:2011.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO/IEEE 11073-10415:2022 has been approved by CEN as EN ISO/IEEE 11073-10415:2022 without any modification.

Introduction

This introduction is not part of IEEE Std 11073-10415-2019, Health informatics—Personal health device communication—Part 10415: Device specialization—Weighing scale.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. This document uses the optimized framework created in IEEE Std 11073-20601^a and describes a specific, interoperable communication approach for weighing scales. These standards align with, and draw on the existing clinically focused standards to provide support for communication of data from clinical or personal health devices.

^a Information on normative references can be found in Clause 2.

Contents

1. Overview	12
1.1 Scope	12
1.2 Purpose	12
1.3 Context	12
1.4 Word usage	13
2. Normative references	13
3. Definitions, acronyms, and abbreviations	14
3.1 Definitions	14
3.2 Acronyms and abbreviations	14
4. Introduction to ISO/IEEE 11073 personal health devices	15
4.1 General	15
4.2 Introduction to IEEE 11073-20601 modeling constructs	15
4.3 Compliance with other standards	16
5. Weighing scale device concepts and modalities	16
5.1 General	16
5.2 Body weight	16
5.3 Body height	17
5.4 Body mass index	17
6. Weighing scale domain information model	17
6.1 Overview	17
6.2 Class extensions	17
6.3 Object instance diagram	17
6.4 Types of configuration	19
6.5 Medical device system object	20
6.6 Numeric objects	23
6.7 Real-time sample array objects	28
6.8 Enumeration objects	28
6.9 PM-store objects	28
6.10 Scanner objects	28
6.11 Class extension objects	28
6.12 Weighing scale information model extensibility rules	28
7. Weighing scale service model	29
7.1 General	29
7.2 Object access services	29
7.3 Object access event report services	31
8. Weighing scale communication model	31
8.1 Overview	31
8.2 Communications characteristics	31

8.3 Association procedure	32
8.4 Configuring procedure.....	34
8.5 Operating procedure	35
8.6 Time synchronization	36
9. Test associations.....	36
9.1 General	36
9.2 Behavior with standard configuration.....	36
9.3 Behavior with extended configurations	36
10. Conformance	37
10.1 Applicability	37
10.2 Conformance specification	37
10.3 Levels of conformance	37
10.4 Implementation conformance statements	38
Annex A (informative) Bibliography	43
Annex B (normative) Any additional ASN.1 definitions	44
Annex C (normative) Allocation of identifiers.....	45
Annex D (informative) Message sequence examples.....	46
Annex E (informative) Protocol data unit examples	48
Annex F (informative) Revision history.....	58

Health informatics—Personal health device communication

Part 10415: Device specialization— Weighing scale

1. Overview

1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth weighing scale devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards, including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth weighing scales.

1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes). Interoperability is the key to growing the potential market for these devices and to enabling people to be better informed participants in the management of their health.

1.3 Context

See IEEE Std 11073-20601™ for an overview of the environment within which this standard is written.¹

This document, IEEE Std 11073-10415, defines the device specialization for the weighing scale, being a specific agent type, and it provides a description of the device concepts, its capabilities, and its implementation according to this standard.

This standard is based on IEEE Std 11073-20601, which in turn draws information from both ISO/IEEE 11073-10201:2004 [B7] and ISO/IEEE 11073-20101:2004 [B8].² The medical device encoding rules (MDER) used within this standard are fully described in IEEE Std 11073-20601.

¹ Information on normative references can be found in Clause 2.

² The numbers in brackets correspond to the numbers of the bibliography in Annex A.

IEEE Std 11073-10415-2019
Health Informatics—Personal health device communication
Part 10415: Device specialization—Weighing scale

This standard defines specialized nomenclature codes that will be collected in future versions of IEEE Std 11073-10101. Between this standard, IEEE Std 11073-10101, IEEE Std 11073-20601, and IEEE Std 11073-104zz, all required nomenclature codes for implementation are documented. New codes may be defined in newer versions/revisions of each of these documents. In the case of a conflict, where one term code has been assigned to two separate semantic concepts with different RefIDs, in general the oldest definition that is in actual use should take precedence. The same policy applies when one RefID has two different code values assigned in different specifications. The resolution of such conflicts will be determined through joint action by the responsible working groups and other stakeholders, and any corrective action published as corrigenda.

NOTE— In this standard, the term *IEEE Std 11073-104zz* is used to refer to the collection of device specialization standards that utilize IEEE Std 11073-20601, where *zz* can be any number from 01 to 99, inclusive.³

1.4 Word usage

The word *shall* indicates mandatory requirements strictly to be followed in order to conform to the standard and from which no deviation is permitted (shall equals is required to).^{4,5}

The word *should* indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others; or that a certain course of action is preferred but not necessarily required (should equals is recommended that).

The word *may* is used to indicate a course of action permissible within the limits of the standard (may equals is permitted to).

The word *can* is used for statements of possibility and capability, whether material, physical, or causal (can equals is able to).

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used, so that each referenced document is cited in text and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-20601™, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol.^{6,7}

IEEE Std 11073-10101™, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.

See Annex A for all informative material referenced by this standard.

³ Notes in text, tables, and figures are given for information only and do not contain requirements needed to implement the standard.

⁴ The use of the word *must* is deprecated and cannot be used when stating mandatory requirements; *must* is used only to describe unavoidable situations.

⁵ The use of *will* is deprecated and cannot be used when stating mandatory requirements; *will* is used only in statements of fact.

⁶ The IEEE standards or products referred to in this clause are trademarks of The Institute of Electrical and Electronics Engineers, Inc.

⁷ IEEE publications are available from The Institute of Electrical and Electronics Engineers (<http://standards.ieee.org/>).