

INTERNATIONAL ISO/IEEE STANDARD 11073-10407

Second edition
2022-12

Health informatics — Device interoperability —

Part 10407: Personal health device communication — Device specialization — Blood pressure monitor

Informatique de santé — Interopérabilité des dispositifs —

*Partie 10407: Communication entre dispositifs de santé personnels —
Spécialisation des dispositifs — Moniteur de pression sanguine*



IEEE

Reference number
ISO/IEEE 11073-10407:2022(E)

© IEEE 2020

This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© IEEE 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from IEEE at the address below.

Institute of Electrical and Electronics Engineers, Inc
3 Park Avenue, New York
NY 10016-5997, USA

Email: stds.ipr@ieee.org
Website: www.ieee.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see www.iso.org/directives).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO/IEEE 11073-10407 was prepared by the *IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society* (as IEEE Std 11073-10407-2020) and drafted in accordance with its editorial rules. It was adopted, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

This second edition cancels and replaces the first edition (ISO/IEEE 11073-10407:2010), which has been technically revised.

A list of all parts in the ISO/IEEE 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Health informatics—Personal health device communication

Part 10407: Device specialization— Blood pressure monitor

Developed by the

IEEE 11073™ Standards Committee
of the
IEEE Engineering in Medicine and Biology Society

Approved 30 January 2020

IEEE SA Standards Board

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth blood pressure monitor devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth blood pressure monitors.

Keywords: blood pressure monitor, IEEE 11073-10407™, medical device communication, personal health devices

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2020 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 30 April 2020. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-6460-4 STD24062
Print: ISBN 978-1-5044-6461-1 STDPD24062

IEEE prohibits discrimination, harassment, and bullying.

For more information, visit <https://www.ieee.org/about/corporate/governance/p9-26.html>.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading “Important Notices and Disclaimers Concerning IEEE Standards Documents.” They can also be obtained on request from IEEE or viewed at <https://standards.ieee.org/ipr/disclaimers.html>.

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (“IEEE SA”) Standards Board. IEEE (“the Institute”) develops its standards through a consensus development process, approved by the American National Standards Institute (“ANSI”), which brings together volunteers representing varied viewpoints and interests to achieve the final product. IEEE Standards are documents developed through scientific, academic, and industry-based technical working groups. Volunteers in IEEE working groups are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE Standards do not guarantee or ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers and users of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; non-infringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied “AS IS” and “WITH ALL FAULTS.”

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under US and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every 10 years. When a document is more than 10 years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit IEEE Xplore at <https://ieeexplore.ieee.org/> or contact IEEE at the address listed previously. For more information about the IEEE SA or IEEE's standards development process, visit the IEEE SA Website at <https://standards.ieee.org>.

Errata

Errata, if any, for IEEE standards can be accessed via <https://standards.ieee.org/standard/index.html>. Search for standard number and year of approval to access the web page of the published standard. Errata links are located under the Additional Resources Details section. Errata are also available in IEEE Xplore: <https://ieeexplore.ieee.org/browse/standards/collection/ieee/>. Users are encouraged to periodically check for errata.

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE SA Website at <https://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patent Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

Participants

At the time this IEEE standard was completed, the Personal Health Device Working Group had the following membership:

Daidi Zhong, Co-Chair
Michael J. Kirwan, Co-Chair

| | | |
|------------------------|---------------------------|---------------------|
| Karsten Aalders | Cory Condek | Robert Hall |
| Charles R. Abbruscato | Todd H. Cooper | Shu Han |
| Nabil Abujbara | David Cornejo | Nathaniel Hamming |
| Maher Abuzaid | Douglas Coup | Rickey L. Hampton |
| James Agnew | Nigel Cox | Sten Hanke |
| Manfred Aigner | Hans Crommenacker | Aki Harma |
| Jorge Alberola | Tomio Crosley | Jordan Hartmann |
| David Aparisi | Allen Curtis | Kai Hassing |
| Lawrence Arne | Jesús Daniel Trigo | Wolfgang Heck |
| Diego B. Arquillo | David Davenport | Nathaniel Heintzman |
| Serafin Arroyo | Russell Davis | Charles Henderson |
| Muhammad Asim | Sushil K. Deka | Jun-Ho Her |
| Kit August | Ciro de la Vega | Helen B. Hernandez |
| Doug Baird | Pedro de-las-Heras-Quiros | Timothy L. Hirou |
| David Baker | Jim Dello Stritto | Allen Hobbs |
| Anindya Bakshi | Kent Dicks | Alex Holland |
| Ananth Balasubramanian | Hyoungdo Do | Arto Holopainen |
| Sunlee Bang | Jonathan Dougherty | Kris Holtzclaw |
| M. Jonathan Barkley | Xiaolian Duan | Xinyi Hong |
| Gilberto Barrón | Sourav Dutta | Robert Hoy |
| David Bean | Jakob Ehrensvarð | Di Hu |
| John Bell | Fredrik Einberg | Anne Huang |
| Olivia Bellamou-Huet | Javier Escayola Calvo | Ron Huby |
| Rudy Belliardi | Mark Estes | David Hughes |
| Daniel Bernstein | Leonardo Estevez | Robert D. Hughes |
| George A. Bertos | Bosco T. Fernandes | Jiyoung Huh |
| Chris Biernacki | Christoph Fischer | Hugh Hunter |
| Ola Björnsne | Morten Flintrup | Philip O. Isaacson |
| Thomas Blackadar | Joseph W. Forler | Atsushi Ito |
| Thomas Bluethner | Russell Foster | Michael Jaffe |
| Douglas P. Bogia | Eric Freudenthal | Praduman Jain |
| Xavier Boniface | Matthias Frohner | Hu Jin |
| Shannon Boucousis | Ken Fuchs | Danny Jochelson |
| Julius Broma | Jing Gao | Akiyoshi Kabe |
| Lyle G. Bullock, Jr. | Marcus Garbe | Steve Kahle |
| Bernard Burg | John Garguilo | Tomio Kamioka |
| Chris Burns | Rick Geimer | James J. Kang |
| Jeremy Byford-Rew | Igor Gejdos | Kei Kariya |
| Satya Calloji | Ferenc Gerbovics | Andy Kaschl |
| Carole C. Carey | Alan Godfrey | Junzo Kashiwara |
| Craig Carlson | Nicolae Goga | Colin Kennedy |
| Santiago Carot-Nemesio | Julian Goldman | Ralph Kent |
| Randy W. Carroll | Raul Gonzalez Gomez | Laurie M. Kermes |
| Seungchul Chae | Chris Gough | Ahmad Kheirandish |
| Peggy Chien | Channa Gowda | Junhyung Kim |
| David Chiu | Charles M. Gropper | Minho Kim |
| Jinyong Choi | Amit Gupta | Min-Joon Kim |
| Chia-Chin Chong | Jeff Guttmacher | Taekon Kim |
| Saeed A. Choudhary | Rasmus Haahr | Tetsuya Kimura |
| Jinhan Chung | Christian Habermann | Alfred Kloos |
| John A. Cogan | Michael Hagerty | Jeongmee Koh |
| John T. Collins | Jerry Hahn | Jean-Marc Koller |

| | | |
|-------------------------------|------------------------|----------------------------|
| John Koon | Begonya Otal | Nicholas Steblay |
| Patty Krantz | Marco Paleari | Lars Steubesand |
| Raymond Krasinski | Bud Panjwani | John (Ivo) Stivorice |
| Alexander Kraus | Carl Pantiskas | Raymond A. Strickland |
| Ramesh Krishna | Harry P. Pappas | Chandrasekaran Subramaniam |
| Geoffrey Kruse | Hanna Park | Hermann Suominen |
| Falko Kuester | Jong-Tae Park | Lee Surprenant |
| Rafael Lajara | Myungeun Park | Ravi Swami |
| Pierre Landau | Soojun Park | Ray Sweidan |
| Jaechul Lee | Phillip E. Pash | Na Tang |
| Jong-Muk Lee | Tong-Bi Pei | Haruyuki Tatsumi |
| Kyong Ho Lee | Soren Petersen | Isabel Tejero |
| Rami Lee | James Petisce | Tom Thompson |
| Sungkee Lee | Peter Piction | Jonas Tirén |
| Woojae Lee | Michael Pliskin | Janet Traub |
| Qiong Li | Varshney Prabodh | Gary Tschautscher |
| Xiangchen Li | Jeff Price | Masato Tsuchida |
| Yingsong Li | Harald Prinzhorn | Ken Tubman |
| Zhuofang Li | Harry Qiu | Akib Uddin |
| Patrick Lichter | Tanzilur Rahman | Sunil Unadkat |
| Lin Lin | Phillip Raymond | Fabio Urbani |
| Jisoon Lim | Terrie Reed | Philipp Urbauer |
| Joon-Ho Lim | Barry Reinhold | Laura Vanzago |
| Liang Liu | Brian Reinhold | Alpo Värri |
| Xiaoming Liu | Melvin I. Reynolds | Andrei Vasilateanu |
| Wei-Jung Lo | John G. Rhoads | Dalimar Velez |
| Charles Lowe | Jeffrey S. Robbins | Martha Velez |
| Don Ludolph | Chris Roberts | Rudi Voon |
| Christian Lusztick | Moskowitz Robert | Barry Vornbrock |
| Bob MacWilliams | Stefan Robert | Isobel Walker |
| Srikkanth Madhurbootheswaran | Scott M. Robertson | David Wang |
| Miriam L. Makhoul | Timothy Robertson | Linling Wang |
| Romain Marmot | David Rosales | Jerry P. Wang |
| Sandra Martinez | Bill Saltzstein | Yao Wang |
| Miguel Martínez de Espronceda | Giovanna Sannino | Yi Wang |
| Cámara | Jose A. Santos-Cadenas | Steve Warren |
| Peter Mayhew | Stefan Sauermann | Fujio Watanabe |
| Jim McCain | John Sawyer | Toru Watsuji |
| László Meleg | Alois Schloegl | Kathleen Wible |
| Alexander Mense | Paul S. Schluter | Paul Williamson |
| Behnaz Minaei | Mark G. Schnell | Jan Wittenber |
| Jinsei Miyazaki | Richard A. Schrenker | Jia-Rong Wu |
| Erik Moll | Antonio Scorpiniti | Will Wykeham |
| Darr Moore | KwangSeok Seo | Ariton Xhafa |
| Chris Morel | Riccardo Serafin | Ricky Yang |
| Carsten Mueglitz | Sid Shaw | Melanie S. Yeung |
| Soundharya Nagasubramanian | Frank Shen | Qiang Yin |
| Alex Neefus | Min Shih | Done-Sik Yoo |
| Trong-Nghia Nguyen-Dobinsky | Mazen Shihabi | Zhi Yu |
| Michael E. Nidd | Redmond Shouldice | Jianchao Zeng |
| Jim Niswander | Sternly K. Simon | Jason Zhang |
| Hiroaki Niwamoto | Marjorie Skubic | Jie Zhao |
| Thomas Norgall | Robert Smith | Thomas Zhao |
| Yoshiteru Nozoe | Ivan Soh | Yuanhong Zhong |
| Abraham Ofek | Motoki Sone | Miha Zoubek |
| Brett Olive | Emily Sopensky | Szymon Zyskoter |
| | Rajagopalan Srinivasan | |

The following members of the individual Standards Association balloting group voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Robert Aiello
Bjoern Andersen
Lyle G. Bullock, Jr.
Keith Chow
Malcolm Clarke
David Fuschi
Randall Groves
Nathaniel Hamming

Werner Hoelzl
Noriyuki Ikeuchi
Atsushi Ito
Raj Jain
Piotr Karocki
Raymond Krasinski
H. Moll

Iulian Profir
Beth Pumo
Stefan Schlichting
Janek Schumann
Walter Struppler
Oren Yuen
Janusz Zalewski
Daidi Zhong

When the IEEE SA Standards Board approved this standard on 30 January 2020, it had the following membership:

Gary Hoffman, *Chair*
Vacant Position, *Vice Chair*
Jean-Philippe Faure, *Past Chair*
Konstantinos Karachalios, *Secretary*

Ted Burse
J. Travis Griffith
Grace Gu
Guido R. Hiertz
Joseph L. Koepfinger*
John D. Kulick
David J. Law

Howard Li
Dong Liu
Kevin Lu
Paul Nikolich
Damir Novosel
Jon Walter Rosdahl

Dorothy Stanley
Mehmet Ulema
Lei Wang
Sha Wei
Philip B. Winston
Daidi Zhong
Jingyi Zhou

*Member Emeritus

Introduction

This introduction is not part of IEEE Std 11073-10407-2020, Health informatics—Personal health device communication—Part 10407: Device specialization—Blood pressure monitor.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. This document uses the optimized framework created in IEEE Std 11073-20601™-2019 and describes a specific, interoperable communication approach for blood pressure monitors.^a These standards align with and draw on the existing clinically focused standards to provide support for communication of data from personal health devices.

^a Information on references can be found in Clause 2.

Contents

| | |
|--|----|
| 1. Overview | 12 |
| 1.1 Scope | 12 |
| 1.2 Purpose | 12 |
| 1.3 Word usage | 12 |
| 1.4 Context | 13 |
| 2. Normative references..... | 13 |
| 3. Definitions, acronyms, and abbreviations | 14 |
| 3.1 Definitions | 14 |
| 3.2 Acronyms and abbreviations | 15 |
| 4. Introduction to ISO/IEEE 11073 personal health devices | 15 |
| 4.1 General | 15 |
| 4.2 Introduction to IEEE 11073-20601 modeling constructs | 16 |
| 4.3 Compliance with other standards..... | 17 |
| 5. Blood pressure monitor device concepts and modalities..... | 17 |
| 5.1 General | 17 |
| 5.2 Systolic and diastolic pressure | 18 |
| 5.3 Mean arterial pressure..... | 18 |
| 5.4 Pulse rate | 18 |
| 5.5 Blood pressure measurement status..... | 18 |
| 6. Blood pressure monitor domain information model..... | 18 |
| 6.1 Overview | 18 |
| 6.2 Class extensions..... | 18 |
| 6.3 Object instance diagram | 18 |
| 6.4 Types of configuration..... | 20 |
| 6.5 Medical device system object..... | 20 |
| 6.6 Numeric objects..... | 25 |
| 6.7 Real-time sample array objects..... | 30 |
| 6.8 Enumeration objects | 34 |
| 6.9 PM-store objects..... | 35 |
| 6.10 Scanner objects..... | 35 |
| 6.11 Class extension objects..... | 36 |
| 6.12 Blood pressure monitor information model extensibility rules..... | 36 |
| 7. Blood pressure monitor service model | 36 |
| 7.1 General | 36 |
| 7.2 Object access services..... | 36 |
| 7.3 Object access event report services | 38 |
| 8. Blood pressure monitor communication model..... | 38 |
| 8.1 Overview | 38 |
| 8.2 Communication characteristics..... | 38 |
| 8.3 Association procedure | 39 |
| 8.4 Configuring procedure..... | 41 |
| 8.5 Operating procedure | 43 |
| 8.6 Time synchronization | 43 |

| | |
|---|----|
| 9. Test associations | 44 |
| 9.1 General | 44 |
| 9.2 Behavior with standard configuration | 44 |
| 9.3 Behavior with extended configurations | 44 |
| 10. Conformance | 44 |
| 10.1 Applicability | 44 |
| 10.2 Conformance specification | 45 |
| 10.3 Levels of conformance | 45 |
| 10.4 Implementation conformance statements | 45 |
| Annex A (informative) Bibliography | 51 |
| Annex B (normative) Additional ASN.1 definitions | 52 |
| B.1 Device and sensor status bit mapping | 52 |
| Annex C (normative) Allocation of identifiers | 53 |
| Annex D (informative) Message sequence examples | 55 |
| Annex E (informative) Protocol data unit examples | 57 |
| E.1 General | 57 |
| E.2 Association information exchange | 57 |
| E.3 Configuration information exchange | 60 |
| E.4 GET MDS attributes service | 63 |
| E.5 Data reporting | 65 |
| E.6 Disassociation | 66 |
| Annex F (informative) Revision history | 67 |

Health informatics—Personal health device communication

Part 10407: Device specialization— Blood pressure monitor

1. Overview

1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth blood pressure monitor devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth blood pressure monitors.

1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes). Interoperability is the key to growing the potential market for these devices and to enabling people to be better informed participants in the management of their health.

1.3 Word usage

The word *shall* indicates mandatory requirements strictly to be followed in order to conform to the standard and from which no deviation is permitted (*shall* equals *is required to*).^{1,2}

¹ The use of the word *must* is deprecated and cannot be used when stating mandatory requirements; *must* is used only to describe unavoidable situations.

² The use of *will* is deprecated and cannot be used when stating mandatory requirements; *will* is used only in statements of fact.

IEEE P11073-10407/D12
Health informatics—Personal health device communication
Part 10407: Device specialization—Blood pressure monitor

The word *should* indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others; or that a certain course of action is preferred but not necessarily required (*should* equals *is recommended that*).

The word *may* is used to indicate a course of action permissible within the limits of the standard (*may* equals *is permitted to*).

The word *can* is used for statements of possibility and capability, whether material, physical, or causal (*can* equals *is able to*).

1.4 Context

See IEEE Std 11073-20601-2019TM for an overview of the environment within which this standard is written.³

This document, IEEE Std 11073-10407, defines the device specialization for the blood pressure monitor, being a specific agent type, and provides a description of the device concepts, its capabilities, and its implementation according to this standard.

This standard is based on IEEE Std 11073-20601-2019, which in turn draws information from both ISO/IEEE 11073-10201:2004 [B6] and ISO/IEEE 11073-20101:2004 [B7].⁴ The medical device encoding rules (MDERs) used within this standard are fully described in IEEE Std 11073-20601-2019.

This standard defines specialized nomenclature codes that will be collected in future revisions of IEEE Std 11073-10101. Between this standard, IEEE Std 11073-10101-2019, IEEE Std 11073-20601-2019, and other IEEE Std 11073-104xx, all required nomenclature codes for implementation are documented. New codes may be defined in newer versions / revisions of each of these documents. In the case of a conflict, where one term code has been assigned to two separate semantic concepts with different RefIDs, in general the oldest definition that is in actual use should take precedence. The same policy applies when one RefID has two different code values assigned in different specifications. The resolution of such conflicts will be determined through joint action by the responsible working groups and other stakeholders, and any corrective actions will be published as corrigenda.

NOTE—In this standard, IEEE Std 11073-104zz is used to refer to the collection of device specialization standards that utilize IEEE Std 11073-20601-2019, where zz can be any number from 01 to 99, inclusive.⁵

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used; therefore, each referenced document is cited in text, and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-10101TM-2019, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.^{6,7}

³ Information on references can be found in Clause 2.

⁴ The numbers in brackets correspond to the numbers of the bibliography in Annex A.

⁵ Notes in text, tables, and figures are given for information only and do not contain requirements needed to implement the standard.

⁶ The IEEE standards or products referred to in this clause are trademarks of The Institute of Electrical and Electronics Engineers, Inc.

⁷ IEEE publications are available from The Institute of Electrical and Electronics Engineers (<https://standards.ieee.org/>).

IEEE Std 11073-20601™-2019, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol.

See Annex A for all informative material referenced by this standard.

3. Definitions, acronyms, and abbreviations

3.1 Definitions

For the purposes of this document, the following terms and definitions apply. The *IEEE Standards Dictionary Online* should be consulted for terms not defined in this clause.⁹

agent: A node that collects and transmits personal health data to an associated manager.

blood pressure: The cyclic pressure (i.e., amount of force applied over a given area divided by the size of this area) exerted by blood against the walls of blood vessels. Noninvasive blood pressure measurement is typically performed at the brachial artery (arm) or radial artery (wrist). There are usually two numbers reported for blood pressure, and with the home monitors, a third number is typically available. The first, and higher, number is produced by the contraction of the heart (See: systolic pressure). The second, lower number is produced by relaxation of the heart (See: diastolic pressure). The third number is the mean arterial pressure.

class: In object-oriented modeling, a class describes the attributes, methods, and events that objects instantiated from the class utilize.

compute engine: *See:* **manager**.

device: A term used to refer to a physical apparatus implementing either an agent or a manager role.

diastolic pressure: This is minimum pressure achieved during the cardiac cycle. It is typically the second and the lower of the readings given as the blood pressure.

handle: An unsigned 16-bit number that is locally unique and identifies one of the object instances within an agent.

manager: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

mean arterial pressure: value of the integral of one cycle of the blood pressure curve divided by the period between successive heart beats.

object: In object-oriented modeling, a particular instantiation of a class. The instantiation realizes attributes, methods, and events from the class.

obj-handle: *See:* **handle**.

personal health device: A device used in personal health applications.

personal telehealth device: *See:* **personal health device**.

⁹ *IEEE Standards Dictionary Online* is available at <https://dictionary.ieee.org>.