
**Conformity assessment — General
requirements for the competence of
proficiency testing providers**

*Évaluation de la conformité — Exigences générales concernant la
compétence des organisateurs d'essais d'aptitude*



This document is a preview generated by EUS



COPYRIGHT PROTECTED DOCUMENT

© ISO/IEC 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	3
4.1 Impartiality	3
4.2 Confidentiality	4
5 Structural requirements	4
6 Resource requirements	5
6.1 General	5
6.2 Personnel	5
6.3 Facilities and environmental conditions	6
6.4 Externally provided products and services	7
7 Process requirements	8
7.1 Establishing, contracting and communicating the PT scheme objectives	8
7.1.1 Review of requests, tenders and contracts	8
7.1.2 PT scheme communication	8
7.2 Design and planning of a PT scheme	9
7.2.1 General	9
7.2.2 Statistical design	10
7.2.3 Determination of assigned values	11
7.3 Production and distribution of PT items	11
7.3.1 Production of PT items	11
7.3.2 Homogeneity and stability assessment of PT items	12
7.3.3 Handling and storage of PT items	12
7.3.4 Packaging, labelling and distribution of PT items	12
7.3.5 Instructions for participants	13
7.4 Evaluation and reporting of PT scheme results	14
7.4.1 Data analysis	14
7.4.2 Evaluation of performance	14
7.4.3 PT reports	15
7.5 Control of the PT scheme process	16
7.5.1 Technical records	16
7.5.2 Control of data and information management	16
7.5.3 Surveillance of the processes	17
7.5.4 Nonconforming work	17
7.6 Handling of complaints	18
7.7 Handling of appeals	19
8 Management system requirements	19
8.1 General requirements	19
8.2 Management system documentation	20
8.3 Control of management system documents	20
8.4 Control of records	20
8.5 Actions to address risks and opportunities	21
8.6 Improvement	21
8.7 Corrective actions	22
8.8 Internal audits	22
8.9 Management reviews	23
Annex A (informative) Types of PT schemes	24

Annex B (informative) Statistical methods for PT	28
Bibliography	36

This document is a preview generated by EVS

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

ISO and IEC draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO and IEC take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO and IEC had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents and <https://patents.iec.ch>. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared by the ISO Committee on Conformity Assessment (CASCO), in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/CLC/JTC 1, *Criteria for conformity assessment bodies*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO/IEC 17043:2010), which has been technically revised.

The main changes are as follows:

- harmonization with ISO/IEC 17025:2017, including technical requirements and structure;
- harmonization with ISO 13528:2022 in terms of terminology;
- incorporation of requirements from ISO/CASCO PROC 33;
- inclusion of the requirement that testing activities, calibration activities and proficiency testing item production conform to the relevant requirements of appropriate ISO conformity assessment standards;
- deletion of Annex C and revision of Annexes A and B.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

Introduction

Proficiency testing (PT) is widely recognized as an essential tool for demonstrating the competence of conformity assessment bodies. PT can provide evidence of competence and it can be an indicator of an underlying or emerging problem. This document is intended to promote confidence in the operations of PT providers. It contains requirements for PT providers to enable them to demonstrate that they operate competently and can generate valid evaluations of participant performance.

PT involves the use of interlaboratory comparisons for the evaluation of laboratory performance. The definition of “interlaboratory comparison” (see 3.4) broadens the use of both the terms “laboratories” and “measurements or tests” for the purposes of this document to include all types of conformity assessment bodies and their activities, respectively. The term “method” as used in this document can be considered synonymous with the term “measurement procedure” as defined in ISO/IEC Guide 99.

There are many different purposes for interlaboratory comparisons, which can be addressed by PT schemes, including but not limited to:

- a) evaluation of the performance of laboratories for specific measurements, tests, calibrations, examinations, inspections or sampling;
- b) identification of problems in laboratories that, for example, can be related to measurement or test methods, effectiveness of training and supervision of personnel, or calibration of equipment;
- c) establishment of the effectiveness of measurement or test methods and the comparability of measurement or test results;
- d) provision of additional confidence to users of measurement or test results;
- e) identification of differences in measurement or test results;
- f) education of participating laboratories based on the outcomes of such comparisons;
- g) validation of measurement uncertainty claims.

For the following types of interlaboratory comparisons, the term PT does not usually apply because laboratory competence must be established in advance, in order to ensure the validity of measurements or tests as well as the metrological traceability of assigned values:

- h) evaluation of the performance characteristics of a measurement or test method (often described as collaborative trials);
- i) assignment of values to reference materials;
- j) support for statements of the equivalence of measurements of National Metrology Institutes (NMIs), or their Designated Institutes (DIs) through “key and supplementary comparisons”, conducted on behalf of the International Bureau of Weights and Measures (BIPM) and associated Regional Metrology Organizations (RMOs).

It is recognized that interlaboratory comparisons for purposes h), i) and j) can contribute to independent demonstrations of laboratory competence. The requirements of this document can be applied to many of the technical planning and operational activities for these interlaboratory comparisons.

This document also requires PT providers to plan and implement actions to address risks and opportunities, based on their experience. Addressing both risks and opportunities establishes a basis for increasing the effectiveness of the management system, achieving improved results and preventing negative events. The PT provider is responsible for deciding which risks and opportunities to address.

The need for ongoing confidence in laboratory performance is essential not only for laboratories and their customers but also for other interested parties, such as regulators, accreditation bodies and other organizations that specify requirements for laboratories. Most of the requirements in this document apply to those evolving areas, especially regarding management, planning and design, personnel,

assuring validity of results and performance evaluations, confidentiality and other aspects, as appropriate.

This document is intended to provide a consistent basis for all interested parties to determine the competence of organizations that provide PT.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Further details can be found in the ISO/IEC Directives, Part 2.

Conformity assessment — General requirements for the competence of proficiency testing providers

1 Scope

This document specifies general requirements for the competence and impartiality of proficiency testing (PT) providers and consistent operation of all proficiency testing schemes. This document can be used as a basis for specific technical requirements for particular fields of application.

Users of proficiency testing schemes, regulatory authorities, organizations and schemes using peer-assessment, accreditation bodies and others can use these requirements in confirming or recognizing the competence of proficiency testing providers.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 17000, *Conformity assessment — Vocabulary and general principles*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 17034, *General requirements for the competence of reference material producers*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC 17000 and ISO/IEC Guide 99 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

assigned value

value attributed to a particular property or characteristic of a *proficiency testing item* (3.8)

[SOURCE: ISO 13528:2022, 3.3, modified — The words "or characteristic" have been added and the word "test" has been replaced with "testing".]

3.2

consensus value

value derived from a collection of results in an *interlaboratory comparison* (3.4)

Note 1 to entry: The phrase "consensus value" is typically used to describe estimates of location and dispersion derived from *participant* (3.6) results in a round of a *proficiency testing scheme* (3.11), but may also be used to refer to values derived from results of a specified subset of such results or, for example, from a number of expert laboratories.

[SOURCE: ISO 13528:2022, 3.11.]