

Water quality - Application of inductively coupled
plasma mass spectrometry (ICP-MS) - Part 1: General
requirements (ISO 17294-1:2024)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 17294-1:2024 sisaldab Euroopa standardi EN ISO 17294-1:2024 ingliskeelset teksti. Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas. Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 03.04.2024. Standard on kättesaadav Eesti Standardimis-ja Akrediteerimiskeskusest.	This Estonian standard EVS-EN ISO 17294-1:2024 consists of the English text of the European standard EN ISO 17294-1:2024. This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation. Date of Availability of the European standard is 03.04.2024. The standard is available from the Estonian Centre for Standardisation and Accreditation.
---	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.060.50

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis-ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis-ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Water quality - Application of inductively coupled plasma
mass spectrometry (ICP-MS) - Part 1: General
requirements (ISO 17294-1:2024)

Qualité de l'eau - Application de la spectrométrie de
masse avec plasma à couplage inductif (ICP- MS) -
Partie 1: Exigences générales (ISO 17294-1:2024)

Wasserbeschaffenheit - Anwendung der induktiv
gekoppelten Plasma-Massenspektrometrie (ICP-MS) -
Teil 1: Allgemeine Anforderungen (ISO 17294-1:2024)

This European Standard was approved by CEN on 20 January 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 17294-1:2024) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2024, and conflicting national standards shall be withdrawn at the latest by October 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 17294-1:2006.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 17294-1:2024 has been approved by CEN as EN ISO 17294-1:2024 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Apparatus	4
5.1 General	4
5.2 Sample introduction	4
5.2.1 General	4
5.2.2 Pump	5
5.2.3 Nebulizer	5
5.2.4 Spray chamber	6
5.2.5 Other systems	6
5.3 Torch and plasma	6
5.4 Gas and gas control	7
5.5 Generator	7
5.6 Transfer of the ions to the mass spectrometer	7
5.7 Mass spectrometer	8
5.7.1 General	8
5.7.2 Lens system	8
5.7.3 Collision or reaction cell	8
5.7.4 Analyser	8
5.7.5 Detector	9
5.7.6 Alternative mass spectrometers and types of instruments	10
5.8 Signal processing and instrument control	11
6 Interferences by concomitant elements	11
6.1 General	11
6.2 Spectral interferences	11
6.2.1 General	11
6.2.2 Possible elimination strategies for polyatomic ion interferences	12
6.3 Non-spectral interferences	13
6.3.1 General	13
6.3.2 Interferences in the nebulization process	13
6.3.3 Interferences in the plasma	13
6.3.4 Interferences in the interface or lens area	14
6.3.5 Possible elimination strategies for non-spectral interferences (matrix effects)	14
7 Adjustment of the apparatus	17
7.1 General	17
7.2 Tuning the apparatus	17
7.2.1 General	17
7.2.2 Alignment of the plasma	18
7.2.3 Mass calibration	18
7.2.4 Resolution	18
7.2.5 Detector	18
7.3 Verification of instrument performance criteria	19
8 Preparatory steps	19
8.1 General	19
8.2 Choice of isotopes	20
8.3 Choice of instrumental settings	20
8.4 Choice of integration time	21
8.5 Choice of reference elements — Internal standards	22

8.6	Linearity and working range.....	22
8.7	Composition of calibration solutions.....	23
8.8	Method development for cool plasma conditions.....	23
8.9	Determination of the method performance.....	23
8.9.1	General.....	23
8.9.2	Instrument detection limit.....	24
8.9.3	Method detection limit.....	24
8.9.4	Precision of the method.....	24
9	Procedure.....	24
9.1	General.....	24
9.2	Calibration.....	24
9.3	Necessary solutions.....	25
9.4	Measurement.....	25
Annex A (informative)	Spectral interferences, choice of isotopes and instrument detection limits for quadrupole ICP-MS instruments.....	27
Bibliography		32

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 230, *Water analysis*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 17294-1:2004), which has been technically revised.

The main changes are as follows:

- scope has been revised to align with ISO 17294-2;
- text has been revised to reflect currently available instruments used in routine daily practice in many laboratories;
- [Clauses 5](#) and [6](#) have been revised to reflect the state-of-the-art equipment used to measure elements according to ISO 17294-2;
- abbreviated terms in [Clause 9](#) have been revised to align with common terms used in other standards;
- [Table A.1](#) has been updated.

A list of all parts in the ISO 17294 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Since the last edition of this document, new developments in metal analysis with inductively coupled plasma mass spectrometry (ICP-MS) have taken place. The use of the collision or reaction cell (CRC) technology in quadrupole ICP-MS and triple quadrupole ICP-MS has increased in laboratories. For this reason, this document has been revised and new items have been added.

The intention for the revision of this document was to focus on the instrumentation currently available and in use for determining elements according to ISO 17294-2 in daily practice in laboratories. The consequence of this starting point is that the use of correction formulae has been moved to [Annex A](#) because of its reduced importance in modern instrumentation. Many principles also apply for high-resolution or accurate mass instrumentation, although they are not described in detail in this document.

Water quality — Application of inductively coupled plasma mass spectrometry (ICP-MS) —

Part 1: General requirements

1 Scope

This document specifies the principles of inductively coupled plasma mass spectrometry (ICP-MS) and provides general requirements for the use of this technique to determine elements in water, digests of sludges and sediments (e.g. digests of water as described in ISO 15587-1 or ISO 15587-2). Generally, the measurement is carried out in water, but gases, vapours or fine particulate matter can be introduced too. This document applies to the use of ICP-MS for aqueous solution analysis.

The ultimate determination of the elements is described in a separate International Standard for each series of elements and matrix. The individual clauses of this document refer the user to these guidelines for the basic principles of the method and the configuration of the instrument.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO Guide 33, *Reference materials — Good practice in using reference materials*

ISO 5725-1, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*

ISO 6206, *Chemical products for industrial use — Sampling — Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO Guide 33, ISO 5725-1, ISO 6206 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

analyte

element(s) to be determined

3.2

blank calibration solution

solution prepared in the same way as the *calibration solution* (3.3) but leaving out the *analyte* (3.1)