

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-62: Particular requirements for the basic safety and essential performance
of high intensity therapeutic ultrasound (HITU) equipment**

**Appareils électromédicaux –
Partie 2-62: Exigences particulières pour la sécurité de base et les performances
essentielle des appareils ultrasonores thérapeutiques de haute intensité (HITU)**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2013 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

Useful links:

IEC publications search - www.iec.ch/searchpub

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Liens utiles:

Recherche de publications CEI - www.iec.ch/searchpub

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-62: Particular requirements for the basic safety and essential performance
of high intensity therapeutic ultrasound (HITU) equipment**

**Appareils électromédicaux –
Partie 2-62: Exigences particulières pour la sécurité de base et les performances
essentielles des appareils ultrasonores thérapeutiques de haute intensité (HITU)**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

XB

ICS 11.040.01; 17.140.50

ISBN 978-2-8322-0865-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
201.1 Scope, object and related standards.....	7
201.2 Normative references	9
201.3 Terms and definitions	9
201.4 General requirements.....	21
201.5 General requirements for testing of ME EQUIPMENT	22
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	23
201.7 ME EQUIPMENT identification, marking and documents	23
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	25
201.9 Protection against mechanical hazards of ME EQUIPMENT and ME SYSTEMS	25
201.10 Protection against unwanted and excessive radiation HAZARDS	25
201.11 Protection against excessive temperatures and other HAZARDS	28
201.12 Accuracy of controls and instruments and protection against hazardous outputs	28
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	30
201.14 Programmable ELECTRICAL MEDICAL SYSTEMS (PEMS).....	30
201.15 Construction of ME EQUIPMENT	30
201.16 ME systems.....	30
201.17 * Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	30
202 Electromagnetic compatibility – Requirements and tests	30
Annexes	33
Annex AA (informative) Particular guidance and rationale	34
Annex BB (informative) Targeting	38
Annex CC (informative) HITU – specific risks.....	41
Annex DD (informative) Determining regions of HITU fields for measurement	46
Annex EE (informative) Guidance in classification according to CISPR 11	57
Annex FF (informative) Notes on using a saline or water bath for EMI testing	58
Bibliography.....	61
Figure 201.101 – Schematic diagram showing the relationship between the various defined surfaces and distances for an ULTRASONIC TRANSDUCER with water stand-off distance when applied to a PATIENT. [IEC 61157 Ed2]	20
Figure 201.102 – Parameters for describing a focusing transducer of a known geometry	20
Figure 201.103 – Example set-up for the measurement of the unwanted ultrasound radiation on the side-wall (the handle) of the transducer	27
Figure DD.1 – Illustration of target, intermediate (shaded or yellow) region and safe regions defined by boundaries 1 and 2.	46
Figure DD.2 – Exposure time vs temperature increase above 37 °C for three different bioeffects threshold exposures shown as solid curves.	47
Figure DD.3 – Two-layer model with target	51
Figure DD.4 – TEMPORAL-AVERAGE INTENSITY (in dB) corrected for absorption vs transverse dimension in the focal plane	54

Figure DD.5 – TEMPORAL-AVERAGE INTENSITY (in dB) vs axial distance z (mm) for a beam from a spherical focusing transducer with a radius of 20 mm and a geometric focal length of 40 mm at 1 MHz.....	55
Figure DD.6 – Overlapping multiple exposure regions in a target region depicted by the dark ellipse	56
Figure FF.1 – Representing the patient or operator impedance.	58
Figure FF.2 – Possible setup for artificial hand for HITU equipment.	59
Figure FF.3 – Showing copper band in saline.	60
Table 201.101 – List of symbols & abbreviations.....	21
Table 201.102 – Distributed ESSENTIAL PERFORMANCE requirements	22
Table CC.1 – Hazards related to image to focus misalignment.....	41
Table CC.2 – Hazards related to use of HITU device by unskilled or untrained personnel or reasonably foreseeable misuse	41
Table CC.3 – Hazards arising from improper acoustic energy	42
Table CC.4 – Lack of, or inadequate, specification for maintenance including inadequate specification of post-maintenance functional checks	43
Table CC.5 – Miscellaneous hazards	43
Table CC.6 – Data transfer errors	43
Table CC.7 – HITU transducer failure	44
Table CC.8 – Generator failure	44
Table CC.9 – Cooling system failure	44
Table CC.10 – Software gets stuck in endless loop.....	44
Table CC.11 – Wrong calculations by computer	45

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-62: Particular requirements for the basic safety and essential performance of high intensity therapeutic ultrasound (HITU) equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any OPERATOR.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-62 has been prepared by IEC subcommittee 62D: [Therapy equipment] Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice. It has been prepared in close co-operation with TC 87 (Ultrasonics).

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1069/FDIS	62D/1076/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

In this particular standard, safety requirements additional to those in the general standard are specified for HIGH INTENSITY THERAPEUTIC ULTRASOUND (HITU) EQUIPMENT.

This particular standard takes into account IEC 62555 and IEC/TS 62556.

The requirements are followed by specifications for the relevant tests.

A rationale for the more important requirements, where appropriate, is given in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the particular standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However this annex does not form part of the requirements of this standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-62: Particular requirements for the basic safety and essential performance of high intensity therapeutic ultrasound (HITU) equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Addition:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HIGH INTENSITY THERAPEUTIC ULTRASOUND EQUIPMENT as defined in 201.3.218, hereafter referred to as ME EQUIPMENT.

This International Standard adds or replaces clauses listed in the IEC 60601-1 that are specific for HIGH INTENSITY THERAPEUTIC ULTRASOUND EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE 1 See also 4.2 of the general standard.

NOTE 2 As, in HITU fields, the acoustic waveform is expected to be extremely distorted due to non-linear propagation effects, the ultrasonic measurements are to be made under quasi linear conditions and then extrapolated following procedures given in IEC/TS 62556. See also IEC/TS 61949

This standard can also be applied to:

- therapeutic equipment for thrombolysis through exposure to high-intensity therapeutic ultrasound;
- therapeutic equipment for the treatment of occluding feeding vessels through exposure to high-intensity focused ultrasound;
- equipment intended to be used for relieving cancer pain due to bone metastases.

This particular standard does not apply to:

- ULTRASOUND EQUIPMENT intended to be used for physiotherapy (use: IEC 60601-2-5 [1]²⁾ and IEC 61689);
- ULTRASOUND EQUIPMENT intended to be used for lithotripsy (use: IEC 60601-2-36 [2]);
- ULTRASOUND EQUIPMENT intended to be used for dedicated hyperthermia devices;
- ULTRASOUND EQUIPMENT intended to be used for phacoemulsification.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

²⁾ Numbers in square brackets refer to the Bibliography.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for HIGH INTENSITY THERAPEUTIC ULTRASOUND (HITU) EQUIPMENT [as defined in 201.3.218.]

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2007 applies as modified in Clause 202. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this particular standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

NOTE Informative references [3,4,5,6,7,8,9,10] are listed in the bibliography beginning on page 61.

Replacement:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

Addition:

IEC 61689:2013, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz*

IEC/TS 61949, *Ultrasonics – Field characterization – In-situ exposure estimation in finite amplitude ultrasonic beams*

IEC 62127-1, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields up to 40 MHz*

IEC 62359, *Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields*

IEC 62555³⁾, *Ultrasonics – Power measurement – High intensity therapeutic ultrasound (HITU) transducers and systems*

IEC/TS 62556⁴⁾, *Ultrasonics – Field characterization – Specification and measurement of field parameters for high intensity therapeutic ultrasound (HITU) transducers and systems*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1, IEC 62359, IEC 62127-1 and IEC 61689, as well as the following additional terms and definitions apply:

3) To be published.

4) To be published.