

PAIKSETE SAASTEALLIKATE HEITED
Lämmastikoksiidide massikontsentratsiooni
määramine
Automaatmõõteseadmete suutlikkusnäitajad

Stationary source emissions
Determination of the mass concentration of nitrogen
oxides
Performance characteristics of automated measuring
systems

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

Käesolev Eesti standard EVS-ISO 10849:2006 "Paiksete saasteallikate heited. Lämmastikoksiidide massikontsentratsiooni määramine. Automaatmõõteseadmete suutlikkusnäitajad" sisaldab rahvusvahelise standardi ISO 10849:1996 "Stationary source emissions - Determination of the mass concentration of nitrogen oxides - Performance characteristics of automated measuring systems" identset ingliskeelset teksti.	This Estonian Standard EVS-ISO 10849:2006 consists of the identical English text of the International Standard ISO 10849:1996 "Stationary source emissions - Determination of the mass concentration of nitrogen oxides - Performance characteristics of automated measuring systems".
Standardi avaldamise korraldas Eesti Standardikeskus.	Estonian standard is published by the Estonian Centre for Standardisation.
Standard EVS-ISO 10849:2006 on kinnitatud Eesti Standardikeskuse 22.12.2006 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teataja 2007. aasta jaanuarikuu numbris.	This standard is ratified with the order of Estonian Centre for Standardisation dated 22.12.2006 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from Estonian Centre for Standardisation.

Käsitlusala

Käesoleva standardiga täpsustatakse paiksete saasteallikate (nt põletusseadmete) juures kasutatavate lämmastikoksiidide automaatmõõtesüsteemi (AMS-i) põhikonstruktsiooni ja peamisi suutlikkusnäitajaid. Samuti kirjeldatakse võtteid suutlikkus näitajate määramiseks.

Lisaks kirjeldatakse meetodeid ja seadmeid NO või NO_x (NO + NO₂) määramiseks suitsugaasides, sh proovivõtusüsteemi ja proovigaasi tasakaalustussüsteemi. Diämmastikoksiidi (N₂O) käesolevas standardis kirjeldatud meetoditega määrata ei saa. Toodud suutlikkusnäitajad kehtivad kogu mõõtesüsteemi kohta proovivõtuseadmest analüsaatorini.

Käesolev standard kirjeldab gaasi väljavõtuga (ekstraktiivse) ja mitteekstraktiivse (saasteallikasiseste (in situ)) süsteeme erinevate analüsaatoritega, mille töö põhineb näiteks järgmistel meetoditel:

- kemoluminestsents;
- dispersioonita infrapunaspektroskoopia;
- dispersioonita ultraviolettspektroskoopia;
- diferentsiaalne optiline absorptsioonspektromeetria.

Märkus 1. Kirjeldatud meetodeid kasutavad ja käesoleva standardi nõuetele vastavad seadmed on kaubanduslikult kättesaadavad.

ICS 13.040.40 Püsiallikate heitmed**Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele**

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse poolt antud kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10849 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 1, *Stationary source emissions*.

Annex A forms an integral part of this International Standard. Annexes B and C are for information only.

© ISO 1996

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Introduction

Nitrogen oxides are produced during most combustion processes. In fossil fuel combustion, nitrogen oxides are produced from nitrogen combined in the fuel and from the oxidation of nitrogen in the air used for combustion. The quantity of nitrogen oxides produced depends upon the nitrogen content of the fuel, the boiler design, the burner design and the boiler operating conditions.

In flue gases from conventional combustion systems, the nitrogen oxides consist of approximately 95 % nitrogen monoxide (NO). The remaining oxide is predominantly nitrogen dioxide (NO₂) formed from the oxidation of NO when the flue gas temperature decreases. These two oxides (NO + NO₂) are generally designated as NO_x. It should be noted that in other processes the ratio of NO to NO₂ may be different and other nitrogen oxides may be present.

There are numerous ways of determining nitrogen oxides in the gases of combustion plants, both wet chemical/analytical methods and instrumental techniques.

This document is a preview generated by EVS

Stationary source emissions — Determination of the mass concentration of nitrogen oxides — Performance characteristics of automated measuring systems

1 Scope

This International Standard specifies the fundamental structure and the most important performance characteristics of automated measuring systems for oxides of nitrogen to be used on stationary source emissions, for example combustion plants. The procedures to determine the performance characteristics are also specified. Furthermore, it describes methods and equipment to determine NO or NO_x (NO + NO₂) in flue gases including the sampling system and sample gas conditioning system. Dinitrogen monoxide (N₂O) is not determined by the methods described in this International Standard. The given performance characteristics refer to the complete measuring system, from sampling to analyser.

This International Standard describes extractive and non-extractive systems in connection with a range of analysers that operate using, for example, the following principles:

- chemiluminescence;
- non-dispersive infrared spectroscopy;
- non-dispersive ultraviolet spectroscopy;
- differential optical absorption spectrometry.

NOTE 1 Commercial devices using the described techniques, that meet the requirements of this International Standard, are available.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publi-

cation, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 6879:1995, *Air quality — Performance characteristics and related concepts for air quality measuring methods*.

ISO 7996:1985, *Ambient air — Determination of the mass concentration of nitrogen oxides — Chemiluminescence method*.

ISO 8096:1992, *Stationary source emissions — Determination of concentration and mass flow rate of particulate material in gas-carrying ducts — Manual gravimetric method*.

ISO 9169:1994, *Air quality — Determination of performance characteristics of measurement methods*.

ISO 10396:1993, *Stationary source emissions — Sampling for the automated determination of gas concentrations*.

3 Definitions

For the purposes of this International Standard, the following definitions apply.

3.1 automated measuring system (AMS): System that may be attached to a chimney to continuously measure and record the mass concentration of nitrogen oxides passing through the chimney.

3.2 analyser: Analytical part in an extractive AMS.