

Water quality - Determination of total cyanide and free cyanide using flow analysis (FIA and CFA) - Part 1: Method using flow injection analysis (FIA) (ISO 14403-1:2012)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 14403-1:2012 sisaldab Euroopa standardi EN ISO 14403-1:2012 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 14403-1:2012 consists of the English text of the European standard EN ISO 14403-1:2012.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 15.07.2012.	Date of Availability of the European standard is 15.07.2012.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.060.50

Võtmesõnad: chemical analysis and testin, chemical analysis and testing, cyanides, determination of content, flow analysis, phenol, quality, testing, water, water quality,

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:
Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

ICS 13.060.50

English Version

Water quality - Determination of total cyanide and free cyanide
using flow analysis (FIA and CFA) - Part 1: Method using flow
injection analysis (FIA) (ISO 14403-1:2012)

Qualité de l'eau - Dosage des cyanures totaux et des
cyanures libres par analyse en flux continu (FIA et CFA) -
Partie 1: Méthode par analyse avec injection de flux (FIA)
(ISO 14403-1:2012)

Wasserbeschaffenheit - Bestimmung von Gesamtcyanid
und freiem Cyanid mittels Fließanalytik (FIA und CFA) - Teil
1: Verfahren mittels Fließinjektionsanalyse (FIA) (ISO
14403-1:2012)

This European Standard was approved by CEN on 13 July 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (EN ISO 14403-1:2012) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2013, and conflicting national standards shall be withdrawn at the latest by January 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 14403-1:2012 has been approved by CEN as a EN ISO 14403-1:2012 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Interferences	2
4.1 Interferences by oxidizing agents	2
4.2 Interferences by sulfide, sulfite, nitrite, and carbonyl compounds	2
4.3 Other interferences	2
5 Principle	2
5.1 Determination of total cyanide	2
5.2 Determination of free cyanide	3
6 Reagents	3
7 Apparatus	5
8 Sampling and sample preparation	6
9 Procedure	7
9.1 Flow system set-up	7
9.2 Reagent blank measurement	7
9.3 Checking the suitability of the flow system	7
9.4 Calibration	8
9.5 Sample measurement	8
10 Calculation	9
11 Expression of results	9
12 Test report	9
Annex A (informative) Examples of flow systems	10
Annex B (normative) Determination of the real cyanide concentration in the potassium cyanide solution (6.18.1)	11
Annex C (informative) Performance data	12
Bibliography	14

Introduction

Methods using flow analysis automate wet chemical procedures and are particularly suitable for the processing of many analytes in water in large series of samples at a high frequency of analysis.

Analysis can be performed by flow injection analysis (FIA) or continuous flow analysis (CFA). Both methods share the feature of an automatic introduction of the sample into a flow system (manifold) in which analytes in the sample react with reagent solutions on their way through the manifold. Sample preparation may be integrated in the manifold. The reaction product is measured in a flow detector (e.g. flow photometer).

See the foreword for a list of parts of this International Standard.

It should be investigated whether and to what extent particular problems require the specification of additional marginal conditions.

Water quality — Determination of total cyanide and free cyanide using flow analysis (FIA and CFA) —

Part 1: Method using flow injection analysis (FIA)

WARNING — Persons using this part of ISO 14403 should be familiar with normal laboratory practice. This part of ISO 14403 does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this part of ISO 14403 be carried out by suitably trained staff.

1 Scope

This part of ISO 14403 specifies methods for the determination of cyanide in various types of water (such as ground, drinking, surface, leachate, and waste water) with cyanide concentrations from 2 µg/l to 500 µg/l expressed as cyanide ions in the undiluted sample. The range of application can be changed by varying the operation conditions, e.g. by diluting the original sample or using a different injection volume.

In this part of ISO 14403, a suitable mass concentration range from 20 µg/l to 200 µg/l is described.

Seawater can be analysed with possible changes in sensitivity and adaptation of the reagent and calibration solutions to the salinity of the samples.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical and laboratory use — Specification and test methods*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

ISO 8466-2, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 2: Calibration strategy for non-linear second-order calibration functions*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

free cyanide

easily liberatable cyanide

sum of cyanide ions and the cyanide bound in weak metal cyanide complexes that liberate HCN at pH 3,8

3.2

total cyanide

free cyanide (3.1), and in addition stronger metal–cyanide complex compounds, with the exception of cyanide bound in gold, platinum, cobalt, ruthenium, and rhodium complexes from which recovery can be partial