

Automotive fuels - Determination of manganese and iron content in unleaded petrol - Inductively coupled plasma optical emission spectrometry (ICP OES) method

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 16136:2015 sisaldab Euroopa standardi EN 16136:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 16136:2015 consists of the English text of the European standard EN 16136:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 11.02.2015.	Date of Availability of the European standard is 11.02.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 71.080.60, 75.160.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Automotive fuels - Determination of manganese and iron content
in unleaded petrol - Inductively coupled plasma optical emission
spectrometry (ICP OES) method**

Carburants pour automobiles - Détermination des teneurs
en fer et en manganèse dans les essences sans plomb -
Méthode spectrométrique optique par plasma à couplage
inductif (ICP OES)

Kraftstoffe für Kraftfahrzeuge - Bestimmung des Gehaltes
an Mangan und Eisen in unverbleitem Ottokraftstoff -
Optische Emissionsspektrometrie mit induktiv gekoppeltem
Plasma (ICP OES)

This European Standard was approved by CEN on 12 December 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle.....	4
4 Reagents.....	4
5 Apparatus	5
6 Sampling.....	6
7 Preparation of solutions	6
7.1 General.....	6
7.2 Preparation of the internal standard solution.....	7
7.3 Preparation of the manganese intermediate solution.....	7
7.4 Preparation of the iron intermediate solution.....	7
7.5 Preparation of the calibration solutions.....	7
7.6 Preparation of quality control solution.....	7
8 Calibration	8
8.1 General.....	8
8.2 Calibration of the ICP OES spectrometer.....	8
8.3 Procedure A.....	8
8.4 Procedure B.....	9
8.5 Check of calibration	9
9 Sample analysis	10
9.1 Sample solution preparation	10
9.2 Sample solution measurement.....	10
10 Calculation.....	10
11 Expression of results	11
12 Precision	11
12.1 General.....	11
12.2 Repeatability, r	11
12.3 Reproducibility, R	11
13 Test report	11
Bibliography.....	13

Foreword

This document (EN 16136:2015) has been prepared by Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2015, and conflicting national standards shall be withdrawn at the latest by August 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 16136:2011.

The major updates are the lowering of the manganese content to allow a specification setting of 2 mg/l of manganese in line with the FQD requirement per 2014-01-01, and the introduction in the scope of determination of iron content, which can be added into petrol as ferrocene.

This document answers requirements originating from the amended Fuels Quality Directive (FQD, [2]).

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies a method based on inductively coupled plasma optical emission spectrometry (ICP OES) for the determination of manganese content from about 0,5 mg/l to about 7,5 mg/l and of iron content from about 1,4 mg/l to about 6,0 mg/l in unleaded petrol containing up to 3,7 % (m/m) oxygen.

WARNING — The use of this European Standard may involve hazardous materials, operations and equipment. This European Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this European Standard to establish appropriate safety and health practices and to determine the applicability of regulatory limitations prior to use.

NOTE 1 Manganese as MMT and iron as ferrocene are added to petrol to increase anti-knock properties.

NOTE 2 Solutions of MMT in petrol are unstable when exposed to light. Low and erratic results are expected if petrol samples are exposed to light prior the analysis.

Iron and manganese contents higher than 6,0 mg/l and 7,5 mg/l respectively may be measured after preliminary dilution of the sample with a suitable solvent. However, the precision has not been established for such a procedure. Further work regarding automotive ethanol (E85) fuel is on-going in CEN.

NOTE 3 For the purposes of this European Standard, the terms “% (m/m)” and “% (V/V)” are used to represent the mass fraction (μ) and the volume fraction (φ) of a material respectively.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 1042, *Laboratory glassware — One-mark volumetric flasks (ISO 1042)*

EN ISO 3170, *Petroleum liquids — Manual sampling (ISO 3170)*

EN ISO 3171, *Petroleum liquids — Automatic pipeline sampling (ISO 3171)*

EN ISO 3675, *Crude petroleum and liquid petroleum products — Laboratory determination of density — Hydrometer method (ISO 3675)*

EN ISO 12185, *Crude petroleum and petroleum products — Determination of density — Oscillating U-tube method (ISO 12185)*

3 Principle

A petrol sample is diluted with a hydrocarbon solvent. The solution is introduced directly into the plasma of an ICP OES spectrometer. Iron and manganese contents are calculated by comparison with calibration solutions prepared from suitable iron and manganese compounds. An internal standard is employed to correct viscosity and vapour pressure effects.

4 Reagents

Unless specified otherwise, only chemicals which are known to have a high degree of purity shall be used.

Some ready-made commercial multi-element Standard solutions may be used instead of the single element Standard solution (4.4 and 4.5).