

Water quality - Enumeration of Escherichia coli and coliform bacteria - Part 2: Most probable number method (ISO 9308-2:2012)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 9308-2:2014 sisaldab Euroopa standardi EN ISO 9308-2:2014 inglisekeelset teksti.	This Estonian standard EVS-EN ISO 9308-2:2014 consists of the English text of the European standard EN ISO 9308-2:2014.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 16.04.2014.	Date of Availability of the European standard is 16.04.2014.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 07.100.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:
Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

ICS 07.100.20

English Version

Water quality - Enumeration of Escherichia coli and coliform bacteria - Part 2: Most probable number method (ISO 9308-2:2012)

Qualité de l'eau - Dénombrement des Escherichia coli et des bactéries coliformes - Partie 2: Méthode du nombre le plus probable (ISO 9308-2:2012)

Wasserbeschaffenheit - Zählung von Escherichia coli und coliformen Bakterien - Teil 2: Verfahren zur Bestimmung der wahrscheinlichsten Keimzahl (ISO 9308-2:2012)

This European Standard was approved by CEN on 11 April 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

The text of ISO 9308-2:2012 has been prepared by Technical Committee ISO/TC 147 "Water quality" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 9308-2:2014 by Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2014, and conflicting national standards shall be withdrawn at the latest by October 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 9308-2:2012 has been approved by CEN as EN ISO 9308-2:2014 without any modification.

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	2
5 Apparatus and glassware	2
6 Culture media and reagents	3
7 Sampling	3
8 Procedure	3
9 Expression of results	4
10 Test report	4
11 Quality assurance	4
Annex A (informative) Further microbiological information on coliform bacteria	5
Annex B (normative) The Quanti-Tray ⁵ Sealer and calculation of results	6
Annex C (informative) Composition of the Colilert-18 medium	42
Annex D (informative) Validation of Colilert-18/Quanti-Tray ⁸ for the enumeration of <i>E.coli</i> and coliform bacteria from water	44
Bibliography	46

Introduction

The presence and extent of faecal pollution is an important factor in assessing the quality of a body of water and the risk to human health from infection. Examination of water samples for the presence of *Escherichia coli* (*E. coli*), which normally inhabits the bowel of man and other warm-blooded animals, provides an indication of such pollution. Examination for coliform bacteria can be more difficult to interpret because some coliform bacteria live in soil and surface fresh water and are not always intestinal. Therefore, the presence of coliform bacteria, although not a proof of faecal contamination, may indicate a failure in treatment or ingress of water into the distribution system.

The International Organization for Standardization (ISO) draws attention to the fact that it is claimed that compliance with this document may involve the use of patents concerning Colilert-18 and Quanti-Tray and Quanti-Tray 2000 given in this document.

ISO takes no position concerning the evidence, validity and scope of these patent rights.

The holder of this patent right has assured the ISO that he/she is willing to negotiate licences either free of charge or under reasonable and non-discriminatory terms and conditions with applicants throughout the world. In this respect, the statement of the holder of this patent right is registered with ISO. Information may be obtained from:

IDEXX Laboratories, Inc.
One IDEXX Drive
Westbrook, Maine 04092 USA

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights other than those identified above. ISO shall not be held responsible for identifying any or all such patent rights.

ISO (<http://www.iso.org/patents>) and IEC (<http://patents.iec.ch>) maintain on-line databases of patents relevant to their standards. Users are encouraged to consult the databases for the most up to date information concerning patents.

Water quality — Enumeration of *Escherichia coli* and coliform bacteria —

Part 2: Most probable number method

WARNING – Persons using this part of ISO 9308 should be familiar with normal laboratory practice. This International Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT – It is absolutely essential that tests conducted in accordance with this part of ISO 9308 be carried out by suitably qualified staff.

1 Scope

This part of ISO 9308 specifies a method for the enumeration of *E. coli* and coliform bacteria in water. The method is based on the growth of target organisms in a liquid medium and calculation of the “Most Probable Number” (MPN) of organisms by reference to MPN tables. This method can be applied to all types of water, including those containing an appreciable amount of suspended matter and high background counts of heterotrophic bacteria. However, it must not be used for the enumeration of coliform bacteria in marine water. When using for the enumeration of *E. coli* in marine waters, a 1→10 dilution in sterile water is typically required, although the method has been shown to work well with some marine waters that have a lower than normal concentration of salts. In the absence of data to support the use of the method without dilution, a 1→10 dilution is used.

This method relies upon the detection of *E. coli* based upon expression of the enzyme β -D-glucuronidase and consequently does not detect many of the enterohaemorrhagic strains of *E. coli*, which do not typically express this enzyme. Additionally, there are a small number of other *E. coli* strains that do not express β -D-glucuronidase.

The choice of tests used in the detection and confirmation of the coliform group of bacteria, including *E. coli*, can be regarded as part of a continuous sequence. The extent of confirmation with a particular sample depends partly on the nature of the water and partly on the reasons for the examination. The test described in this part of ISO 9308 provides a confirmed result with no requirement for further confirmation of positive wells.

NOTE While this method describes the use of an enumeration device that is commercially available, the medium described here can also be used in a standard MPN format.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8199, *Water quality — General guide to the enumeration of micro-organisms by culture*

ISO/IEC Guide 2:2004, *Standardization and related activities — General vocabulary*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in ISO/IEC Guide 2 and the following apply.

- 3.1**
coliform bacterium
member of the *Enterobacteriaceae* that express the enzyme β -D-galactosidase
- 3.2**
Escherichia coli
member of the *Enterobacteriaceae* that expresses both β -D-galactosidase and β -D-glucuronidase enzymes

4 Principle

A snap pack of dehydrated medium is added to a sample of water (100 ml), or a dilution of a sample made up to 100 ml. Sample plus medium is gently shaken to ensure adequate mixing and to afford dissolution of the medium. The sample plus medium is then aseptically poured into a Quanti-Tray¹⁾ or Quanti-Tray/2000¹⁾ to enumerate up to 201 organisms or 2 419 organisms per 100 ml, respectively. Trays are sealed with a Quanti-Tray¹⁾ Sealer and then incubated at $(36 \pm 2) ^\circ\text{C}$ for 18 h to 22 h.

After incubation, sample wells that have a yellow colour of equal or greater intensity than that of the comparator wells are considered positive for coliform bacteria. Yellow wells that also exhibit any degree of fluorescence are considered positive for *E. coli*.

By means of statistical tables, or a simple computer program, the most probable number (MPN) of coliform bacteria and *E. coli* in 100 ml of the sample can be determined.

NOTE The yellow colouration can be seen with the naked eye and results from the cleavage of ortho-nitrophenol galactoside by the enzyme β -D-galactosidase. The fluorescence is demonstrable under ultraviolet light (365 nm) and originates from the cleavage of the molecule 4-methylumbelliferyl glucuronide (MUG) by the enzyme β -D-glucuronidase to produce the fluorescent compound methyl umbelliferone.

5 Apparatus and glassware

Use microbiological laboratory equipment and, in particular, the following:

5.1 Apparatus for sterilization by steam (autoclave)

Apparatus and glassware not supplied sterile shall be sterilized according to the instructions given in ISO 8199.

5.2 Hot air oven, for dry heat sterilization.

5.3 Incubator, thermostatically controlled at $(36 \pm 2) ^\circ\text{C}$.

5.4 Quanti-Tray¹⁾ sealer.

5.5 Sterile wide mouthed vessels of at least 110 ml.

1) Quanti-Tray is a trademark or registered trademark of IDEXX Laboratories, Inc. or its affiliates in the United States and/or other countries. This information is given for the convenience of users of this part of ISO 9308 and does not constitute an endorsement by ISO of this product.