

INTERNATIONAL
STANDARD

ISO/ASTM
51401

Third edition
2013-11-15

Practice for use of a dichromate dosimetry system

Pratique de l'utilisation d'un système dosimétrique au dichromate



Reference number
ISO/ASTM 51401:2013(E)

© ISO/ASTM International 2013

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor ASTM International accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and ASTM members. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat or ASTM International at the addresses given below.

© ISO/ASTM International 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

ASTM International, 100 Barr Harbor Drive, PO Box C700,
West Conshohocken, PA 19428-2959, USA
Tel. +610 832 9634
Fax +610 832 9635
E-mail khooper@astm.org
Web www.astm.org

Published in Switzerland

Contents	Page
1 Scope	1
2 Referenced documents	1
3 Terminology	2
4 Significance and use	2
5 Effect of influence quantities	2
6 Interferences	2
7 Apparatus	3
8 Reagents	3
9 Preparation of dosimeters	3
10 Calibration of the dosimetry system	4
11 Application of dosimetry system	5
12 Minimum documentation requirements	6
13 Measurement uncertainty	6
14 Keywords	6
Figure 1	
Relative response of dichromate dosimeter as a function of irradiation temperature. A fit of the data using Eq 1 yields fit parameters as follows: $b_0 = 1.021; b_1 = -6.259 \times 10^{-5}; b_2 = 1.806$	3
Figure 2	
Response of high-range dosimeter in terms of ΔA as a function of absorbed dose to water. A least-squares third order polynomial fit (see Eq 4) of the data yields fit parameters as follows: $b_0 = 7.515 \times 10^{-4}; b_1 = 1.745 \times 10^{-2}; b_2 = 3.485 \times 10^{-6}; b_3 = -2.765 \times 10^{-7}$	5
Figure 3	
Response of the low-range dosimeter in terms of ΔA as a function of absorbed dose to water. A least-squares second order polynomial fit (see Eq 4) of the data yields fit parameters as follows: $b_0 = -1.162 \times 10^{-3}; b_1 = 1.200 \times 10^{-1}; b_2 = -3.398 \times 10^{-4}$	5
Table 1 Effect of irradiation temperature on dosimeter response	2
Table 2 Typical dichromate calibration data	5

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

A pilot project between ISO and ASTM International has been formed to develop and maintain a group of ISO/ASTM radiation processing dosimetry standards. Under this pilot project, ASTM Subcommittee E61, Radiation Processing, is responsible for the development and maintenance of these dosimetry standards with unrestricted participation and input from appropriate ISO member bodies.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. Neither ISO nor ASTM International shall be held responsible for identifying any or all such patent rights.

International Standard ISO/ASTM 51401 was developed by ASTM Committee E61, Radiation Processing, through Subcommittee E61.02, on Dosimetry Systems, and by Technical Committee ISO/TC 85, Nuclear energy, nuclear technologies and radiological protection.

This third edition cancels and replaces the second edition (ISO/ASTM 51401:2003), which has been technically revised.



Standard Practice for Use of a Dichromate Dosimetry System¹

This standard is issued under the fixed designation ISO/ASTM 51401; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

1. Scope

1.1 This practice covers the preparation, testing, and procedure for using the acidic aqueous silver dichromate dosimetry system to measure absorbed dose to water when exposed to ionizing radiation. The system consists of a dosimeter and appropriate analytical instrumentation. For simplicity, the system will be referred to as the dichromate system. The dichromate dosimeter is classified as a type I dosimeter on the basis of the effect of influence quantities. The dichromate system may be used as either a reference standard dosimetry system or a routine dosimetry system.

1.2 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ISO/ASTM Practice 52628 for the dichromate dosimetry system. It is intended to be read in conjunction with ISO/ASTM Practice 52628.

1.3 This practice describes the spectrophotometric analysis procedures for the dichromate system.

1.4 This practice applies only to gamma radiation, X-radiation/bremsstrahlung, and high energy electrons.

1.5 This practice applies provided the following conditions are satisfied:

1.5.1 The absorbed dose range is from 2×10^3 to 5×10^4 Gy.

1.5.2 The absorbed dose rate does not exceed 600 Gy/pulse (12.5 pulses per second), or does not exceed an equivalent dose rate of 7.5 kGy/s from continuous sources (1).²

1.5.3 For radionuclide gamma sources, the initial photon energy shall be greater than 0.6 MeV. For bremsstrahlung photons, the initial energy of the electrons used to produce the bremsstrahlung photons shall be equal to or greater than 2 MeV. For electron beams, the initial electron energy shall be greater than 8 MeV.

¹ This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved Sept. 14, 2013. Published November 2013. Originally published as ASTM E 1401 – 91. ASTM E 1401 – 96¹ was adopted by ISO in 1998 with the intermediate designation ISO 15561:1998(E). The present International Standard ISO/ASTM 51401:2013(E) replaces ISO 15561 and is a major revision of the last previous edition ISO/ASTM 51401:2003(E).

² The boldface numbers in parentheses refer to the bibliography at the end of this practice.

NOTE 1—The lower energy limits given are appropriate for a cylindrical dosimeter ampoule of 12 mm diameter. Corrections for displacement effects and dose gradient across the ampoule may be required for electron beams (2). The dichromate system may be used at lower energies by employing thinner (in the beam direction) dosimeter containers (see ICRU Report 35).

1.5.4 The irradiation temperature of the dosimeter shall be above 0°C and should be below 80°C.

NOTE 2—The temperature coefficient of dosimeter response is known only in the range of 5 to 50°C (see 5.2). Use outside this range requires determination of the temperature coefficient.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. Specific precautionary statements are given in 9.3.*

2. Referenced documents

2.1 ASTM Standards:³

E170 Terminology Relating to Radiation Measurements and Dosimetry

E178 Practice for Dealing With Outlying Observations

E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers

E666 Practice for Calculating Absorbed Dose From Gamma or X Radiation

E668 Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose in Radiation-Hardness Testing of Electronic Devices

E925 Practice for Monitoring the Calibration of Ultraviolet-Visible Spectrophotometers whose Spectral Bandwidth does not Exceed 2 nm

E958 Practice for Estimation of the Spectral Bandwidth of Ultraviolet-Visible Spectrophotometers

2.2 ISO/ASTM Standards:³

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51707 Guide for Estimating Uncertainties in Dosimetry for Radiation Processing

52628 Practice for Dosimetry in Radiation Processing

³ For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.



52701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

2.3 ISO/IEC Standards:⁴

17025 General Requirements for the Competence of Testing and Calibration Laboratories

2.4 Joint Committee for Guides in Metrology (JCGM) Reports:⁵

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement

2.5 International Commission on Radiation Units and Measurements (ICRU) Reports:⁶

ICRU Report 35 Radiation Dosimetry: Electrons With Initial Energies Between 1 and 50 MeV

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

3. Terminology

3.1 Definitions:

3.1.1 approved laboratory—laboratory that is a recognized national metrology institute; or has been formally accredited to ISO/IEC 17025; or has a quality system consistent with the requirements of ISO/IEC 17025.

3.1.1.1 Discussion—A recognized national metrology institute or other calibration laboratory accredited to ISO/IEC 17025 should be used in order to ensure traceability to a national or international standard. A calibration certificate provided by a laboratory not having formal recognition or accreditation will not necessarily be proof of traceability to a national or international standard.

3.1.2 reference standard dosimetry system—dosimetry system, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.3 type I dosimeter—dosimeter of high metrological quality, the response of which is affected by individual influence quantities in a well-defined way that can be expressed in terms of independent correction factors.

3.2 Definitions of other terms used in this practice that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E170. Definitions in E170 are compatible with ICRU Report 85a; that document, therefore, may be used as an alternative reference.

4. Significance and use

4.1 The dichromate system provides a reliable means for measuring absorbed dose to water. It is based on a process of

reduction of dichromate ions to chromic ions in acidic aqueous solution by ionizing radiation.

4.2 The dosimeter is a solution containing silver and dichromate ions in perchloric acid in an appropriate container such as a sealed glass ampoule. The solution indicates absorbed dose by a change (decrease) in optical absorbance at a specified wavelength(s) ((3), ICRU Report 80). A calibrated spectrophotometer is used to measure the absorbance.

5. Effect of influence quantities

5.1 Guidance on the determination of the performance characteristics of dosimeters and dosimetry systems can be found in ASTM Guide 52701. The relevant influence quantities that need to be considered when using the dichromate dosimetry system are given below.

5.2 The dosimeter response has a temperature dependence during irradiation that is approximately equal to −0.2 % per degree Celsius between 25 and 50°C. At temperatures below 25°C, the dependence is smaller. The dosimeter response between 5 and 50°C is shown in Table 1, where the response at a given temperature is tabulated relative to the response at 25°C (4,5).

5.2.1 The data in Table 1 may be fitted with an appropriate formula for convenience of interpolation as follows:

R_t = b_0 + b_1 t^{b_2} (1)

where:

R_t = dosimeter response at temperature t relative to that at 25°C.

The curve generated from the fitted data is shown in Fig. 1.

5.3 No effect of ambient light (even direct sunlight) has been observed on dichromate solutions in glass ampoules (6).

5.4 The dosimeter response is dependent on the type and energy of the radiation employed. For example, the response in high energy (10 MeV) electron beams is reported to be approximately 3 % lower than the response in cobalt-60 radiation (2).

5.5 Provided the dosimeter solution is prepared as described in this document, and steps are taken to avoid contamination, the dosimeter solution stored, or sealed, in glass vessels (for example, ampoules) is stable for several years before and after irradiation.

6. Interferences

6.1 The dichromate dosimetric solution response is sensitive to impurities, particularly organic impurities. Even in trace quantities, impurities can cause a detectable change in the observed response (6). For high accuracy results, organic

TABLE 1 Effect of irradiation temperature on dosimeter response

Temperature, °C	Relative Response	Temperature, °C	Relative Response
5	1.020	30	0.992
10	1.017	35	0.983
15	1.013	40	0.972
20	1.007	45	0.960
25	1.000	50	0.948

⁴ Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, CP 56, CH-1211 Geneva 20, Switzerland, <http://www.iso.org>.

⁵ Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

⁶ Available from the International Commission on Radiation Units and Measurements (ICRU), 7910 Woodmont Ave., Bethesda, MD 20814, U.S.A.