
Compressed air —

Part 3:

Test methods for measurement of humidity

Air comprimé —

Partie 3: Méthodes d'essai pour mesurer le taux d'humidité



Contents

1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Units	1
5 Selection guide and available methods	2
6 Sampling techniques	2
7 Measurement methods	3
8 Evaluation of test results	5
9 Conversions from non-standard humidity units to standard format and vice versa	5
10 Uncertainty	6
11 Expression of results	6
12 Test report	6
Annex A (informative) Example of compressed air humidity statement	7
Annex B (informative) Calculation of vapour pressure	8
Annex C (informative) Preferred methods of humidity measurement	10
Annex D (informative) Non-preferred methods of humidity measurement	13
Bibliography	14

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 8573-3 was prepared by Technical Committee ISO/TC 118 *Compressors, pneumatic tools and pneumatic machines*, Subcommittee SC 4, *Quality of compressed air*.

ISO 8573 consists of the following parts, under the general title *Compressed air*

- *Part 1: Contaminants and quality classes*
- *Part 2: Test methods for aerosol oil content*
- *Part 3: Test methods for measurement of humidity*
- *Part 4: Test methods for solid particle content*
- *Part 5: Test methods for oil vapour and organic solvent content*
- *Part 6: Test methods for gaseous contaminant content*
- *Part 7: Test methods for viable microbiological contaminant content*

Annexes A, B, C and D are for information only.

Introduction

This part of ISO 8573 is one in a series of International Standards (planned or published) with the aim of harmonizing air contamination measurements. It is also intended to be used for reference when stating air purity class according to ISO 8573-1.

This document is a preview generated by EVS

Compressed air —

Part 3:

Test methods for measurement of humidity

1 Scope

This part of ISO 8573 provides guidance on selection from the available suitable methods for measurement of humidity in compressed air and specifies the limitations of the various methods.

It does not provide methods for measurement of water content in states other than vapour.

This part of ISO 8573 specifies sampling techniques, measurement, evaluation, uncertainty considerations and reporting for the air contamination parameter humidity.

It gives guidance for the conversion of humidity statements to the standard format.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 8573. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 8573 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest editions of the normative documents referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3857-1, *Compressors, pneumatic tools and machines — Vocabulary — Part 1: General*.

ISO 5598, *Fluid power systems and components — Vocabulary*.

ISO 7183:1986, *Compressed air dryers — Specifications and testing*.

ISO 8573-1, *Compressed air — Part 1: Contaminants and purity classes*.

3 Terms and definitions

For the purposes of this part of ISO 8573, the terms and definitions given in ISO 3857-1 and ISO 5598 and the specific humidity terms and definitions given in ISO 7183 apply.

4 Units

For the purposes of this part of ISO 8573, the following non-preferred SI units are used:

1 bar = 100 000 Pa

NOTE Bar(e) is used to indicate effective pressure above atmospheric.

1 l (litre) = 0,001 m³