
Reference neutron radiations —

Part 1:

Characteristics and methods of production

Rayonnements neutroniques de référence —

Partie 1: Caractéristiques et méthodes de production



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope	1
2 Normative references	2
3 Tests and definitions	2
4 Reference radiations for the calibration of neutron-measuring devices	5
5 Reference radiations for the determination of the response of neutron-measuring devices as a function of neutron energy	9
Annex A (normative) Tabular and graphical representation of the neutron spectra for radionuclide sources	13
Annex B (informative) Angular source strength characteristics of two radionuclide neutron sources	20
Annex C (normative) Conventional thermal neutron fluence rate	22
Bibliography	23

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 8529 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 8529-1 was prepared by Technical Committee ISO/TC 85, *Nuclear energy*, Subcommittee SC 2, *Radiation protection*.

ISO 8529 consists of the following parts, under the general title *Reference neutron radiations*:

- *Part 1: Characteristics and methods of production*
- *Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*
- *Part 3: Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence*

Annexes A and C form a normative part of this part of ISO 8529. Annex B is for information only.

Introduction

This part of ISO 8529 supersedes ISO 8529:1989. It is the first of a set of three International Standards concerning the calibration of dosimeters and dose-rate meters for neutron radiation for protection purposes. It describes the characteristics and methods of production of the reference neutron radiations to be used for calibrations. ISO 8529-2 describes fundamentals related to the physical quantities characterizing the radiation field and calibration procedures in general terms, with emphasis on active dose-rate meters and the use of radionuclide sources. ISO 8529-3 deals with dosimeters for area and individual monitoring, describing the respective procedures for calibrating and determining the response in terms of the International Commission on Radiation Units and Measurements (ICRU) operational quantities. Conversion coefficients for converting neutron fluence into these operational quantities are provided in ISO 8529-3.

This document is a preview generated by EVS

This document is a preview generated by EVS

Reference neutron radiations —

Part 1:

Characteristics and methods of production

1 Scope

This part of ISO 8529 specifies the reference neutron radiations, in the energy range from thermal up to 20 MeV, for calibrating neutron-measuring devices used for radiation protection purposes and for determining their response as a function of neutron energy. Reference radiations are given for neutron fluence rates of up to $1 \times 10^9 \text{ m}^{-2} \cdot \text{s}^{-1}$, corresponding, at a neutron energy of 1 MeV, to dose-equivalent rates of up to $100 \text{ mSv} \cdot \text{h}^{-1}$.

This part of ISO 8529 is concerned only with the methods of producing and characterizing the neutron reference radiations. The procedures for applying these radiations for calibrations are described in ISO 8529-2 and ISO 8529-3.

The reference radiations specified are the following:

- neutrons from radionuclide sources, including neutrons from sources in a moderator;
- neutrons produced by nuclear reactions with charged particles from accelerators;
- neutrons from reactors.

In view of the methods of production and use of them, these reference radiations are divided, for the purposes of this part of ISO 8529, into the following two separate sections.

- In clause 4, radionuclide neutron sources with wide spectra are specified for the calibration of neutron-measuring devices. These sources should be used by laboratories engaged in the routine calibration of neutron-measuring devices, the particular design of which has already been type tested.
- In clause 5, accelerator-produced monoenergetic neutrons and reactor-produced neutrons with wide or quasi monoenergetic spectra are specified for determining the response of neutron-measuring devices as a function of neutron energy. Since these reference radiations are produced at specialized and well equipped laboratories, only the minimum of experimental detail is given.

For the conversion of neutron fluence into the quantities recommended for radiation protection purposes, conversion coefficients have been calculated based on the spectra presented in normative annex A and using the fluence-to-dose-equivalent conversion coefficients as a function of neutron energy as given in ICRP Publication 74 and ICRU Report 57.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 8529. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 8529 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 8529-2:2000 *Reference neutron radiations — Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field.*

ISO 8529-3:1998, *Reference neutron radiations — Part 3: Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence.*

ICRP Publication 74, *Conversion Coefficients for use in Radiological Protection against External Radiation, Annals of the ICRP*, Vol. 26, No.3/4 (1996).

ICRU Report 33:1980, *Radiation Quantities and Units.*

ICRU Report 51:1993, *Quantities and Units in Radiation Protection Dosimetry.*

ICRU Report 57:1998, *Conversion Coefficients for use in Radiological Protection Against External Radiation.*

3 Tests and definitions

For the purposes of this part of ISO 8529, the terms and definitions given in ICRU Reports 33 and 51 and the following apply.

3.1

neutron fluence

Φ

quotient of dN by da , where dN is the number of neutrons incident on a sphere of cross-sectional area da

$$\Phi = \frac{dN}{da}$$

NOTE The unit of the neutron fluence is m^{-2} ; a frequently used unit is cm^{-2} .

3.2

neutron fluence rate

neutron flux density

ϕ

quotient of $d\Phi$ by dt , where $d\Phi$ is the increment of **neutron fluence** (3.1) in the time interval dt

$$\phi = \frac{d\Phi}{dt} = \frac{d^2N}{da \cdot dt}$$

NOTE The unit of the neutron fluence rate is $m^{-2} \cdot s^{-1}$.

3.3

spectral neutron fluence

energy distribution of the neutron fluence

Φ_E

quotient of $d\Phi$ by dE , where $d\Phi$ is the increment of neutron fluence in the energy interval between E and $E + dE$