

INTERNATIONAL
STANDARD

ISO/ASTM
52628

First edition
2013-11-15

Practice for dosimetry in radiation processing

Pratique standard pour dosimétrie au traitement par irradiation



Reference number
ISO/ASTM 52628:2013(E)

© ISO/ASTM International 2013

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor ASTM International accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and ASTM members. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat or ASTM International at the addresses given below.

© ISO/ASTM International 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

ASTM International, 100 Barr Harbor Drive, PO Box C700,
West Conshohocken, PA 19428-2959, USA
Tel. +610 832 9634
Fax +610 832 9635
E-mail khooper@astm.org
Web www.astm.org

Published in Switzerland

Contents	Page
1 Scope	2
2 Reference documents	2
3 Terminology	3
4 Significance and use	4
5 Dosimetry system requirements	4
6 Classification	4
7 Guidance	6
8 Keywords.....	8
Annex.....	9
Table 1 Examples of type I dosimeters	5
Table 2 Examples of type II dosimeters	6
Table 3 General dosimetry requirements for all radiation applications	6
Table 4 Dosimetry requirements for specific radiation applications	6
Table 5 Guidance for dosimetry in specific radiation applications	7
Table 6 Guidance on absorbed-dose mapping and mathematical methods.....	7
Table A1.1 Summary of characteristics of dosimeters described in ASTM and ISO/ASTM radiation processing standards	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

A pilot project between ISO and ASTM International has been formed to develop and maintain a group of ISO/ASTM radiation processing dosimetry standards. Under this pilot project, ASTM Committee E61, Radiation Processing, is responsible for the development and maintenance of these dosimetry standards with unrestricted participation and input from appropriate ISO member bodies.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. Neither ISO nor ASTM International shall be held responsible for identifying any or all such patent rights.

International Standard ISO/ASTM 52628 was developed by ASTM Committee E61, Radiation Processing, through Subcommittee E61.01, on Dosimetry, and by Technical Committee ISO/TC 85, Nuclear energy, nuclear technologies and radiological protection.

This first edition of ISO/ASTM 52628 cancels and replaces ASTM E2628-09 ϵ 1.



Standard Practice for Dosimetry in Radiation Processing¹

This standard is issued under the fixed designation ISO/ASTM 52628; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

INTRODUCTION

The use of ionizing radiation for the treatment of commercial products such as the sterilization of medical devices, the reduction of microbial contamination in food or the modification of polymers is referred to as radiation processing. The types of radiation used may be gamma radiation (typically from cobalt-60 sources), X-radiation or accelerated electrons.

It is necessary to ensure that the specified absorbed dose is applied in each of the radiation processing applications. The absorbed dose must be measured, and measurement systems have been developed for this purpose. Much of the development of these systems rests on the early development of dosimetry systems for personnel radiation protection and for medical treatment. However, the absorbed doses used in radiation processing are generally higher, ranging from ~10 Gy up to 100 kGy or more and new dosimetry systems have been developed for measurements of these doses.

Note that the terms “dose” and “absorbed dose” are used interchangeably in this standard (see 3.1.1).

The dose measurements required in radiation processing concern characterization of radiation facilities in installation qualification (IQ) and operational qualification (OQ), measurement of dose distribution in irradiated products in performance qualification (PQ) and routine monitoring of the irradiation process.

The literature is abundant with articles on dosimeters for radiation processing, and guidelines and standards have been written by several organizations (the International Atomic Energy Agency (IAEA) and the International Commission on Radiation Units and Measurements (ICRU), for example) for the operation of the dosimetry systems and for their use in the characterization and validation of the radiation processing applications. In particular, ICRU Report 80 provides information on the scientific basis and historical development of many of the systems in current use.

ASTM Subcommittee E10.01 on Radiation Processing: Dosimetry and Applications was formed in 1984 initially with the scope of developing standards for food irradiation, but its scope was widened to include all radiation processing applications. The subcommittee, now Committee E61, has under its jurisdiction approximately 30 standard practices and standard guides, collectively known as the E61 standards on radiation processing. A number of these standards have been published as ISO/ASTM standards, thereby ensuring a wider international acceptance. These practices and guides describe the dosimetry systems most commonly used in radiation processing, and the dose measurements that are required in the validation and routine monitoring of the radiation processes. A current list of the E61 standards on radiation processing is given in 2.1 and 2.2.

The development, validation and routine control of a radiation process comprises a number of activities, most of which rely on the ability to measure the delivered dose accurately. It is therefore necessary that dose is measured with traceability to national, or international, standards, and the uncertainty is known, including the effect of influence quantities. The E61 standards on radiation processing dosimetry serve to fulfill these requirements.

The practices describing dosimetry systems have several common attributes, and there is a need to have one general standard that can act as a common reference and that can be used as a basis for the selection of dosimetry systems for defined tasks. ISO/ASTM Practice 52628 serves this purpose. It outlines general requirements for the calibration and use of dosimetry systems and for the estimation of measurement uncertainties. Details relating to each dosimetry system are found in the respective standards and each of these refer to ISO/ASTM Practice 52628 for the general requirements.



1. Scope

1.1 This practice describes the basic requirements that apply when making absorbed dose measurements in accordance with the ASTM E61 series of dosimetry standards. In addition, it provides guidance on the selection of dosimetry systems and directs the user to other standards that provide specific information on individual dosimetry systems, calibration methods, uncertainty estimation and radiation processing applications.

1.2 This practice applies to dosimetry for radiation processing applications using electrons or photons (gamma- or X-radiation).

1.3 This practice addresses the minimum requirements of a *measurement management system*, but does not include general quality system requirements.

1.4 This practice does not address personnel dosimetry or medical dosimetry.

1.5 This practice does not apply to *primary standard dosimetry systems*.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Reference documents

2.1 ASTM Standards:²

- E170 Terminology Relating to Radiation Measurements and Dosimetry
- E1026 Practice for Using the Fricke Dosimetry System
- E2232 Guide for Selection and Use of Mathematical Methods for Calculating Absorbed Dose in Radiation Processing Applications
- E2303 Guide for Absorbed-Dose Mapping in Radiation Processing Facilities
- E2304 Practice for Use of a LiF Photo-Fluorescent Film Dosimetry System
- E2381 Guide for Dosimetry In Radiation Processing of Fluidized Beds and Fluid Streams
- E2449 Guide for Irradiation of Pre-packaged Processed Meat and Poultry Products to Control Pathogens and Other Microorganisms
- F1355 Guide for Irradiation of Fresh Agricultural Produce as a Phytosanitary Treatment
- F1356 Practice for Irradiation of Fresh and Frozen Red Meat and Poultry to Control Pathogens and Other Microorganisms

F1736 Guide for Irradiation of Finfish and Aquatic Invertebrates Used as Food to Control Pathogens and Spoilage Microorganisms

F1885 Guide for Irradiation of Dried Spices, Herbs, and Vegetable Seasonings to Control Pathogens and Other Microorganisms

2.2 ISO/ASTM Standards:²

- 51205 Practice for Use of a Ceric-Cerous Sulfate Dosimetry System
- 51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing
- 51275 Practice for Use of a Radiochromic Film Dosimetry System
- 51276 Practice for Use of a Polymethylmethacrylate Dosimetry System
- 51310 Practice for Use of a Radiochromic Optical Waveguide Dosimetry System
- 51401 Practice for Use of a Dichromatic Dosimetry System
- 51431 Practice for Dosimetry in Electron Beam and X-Ray (Bremsstrahlung) Irradiation Facilities for Food Processing
- 51538 Practice for Use of the Ethanol-Chlorobenzene Dosimetry System
- 51540 Practice for Use of a Radiochromic Liquid Dosimetry System
- 51607 Practice for Use of the Alanine-EPR Dosimetry System
- 51608 Practice for Dosimetry in an X-Ray (Bremsstrahlung) Facility for Radiation Processing
- 51631 Practice for Use of Calorimetric Dosimetry Systems for Electron Beam Dose Measurements and Routine Dosimeter Calibration
- 51649 Practice for Dosimetry in an Electron-Beam Facility for Radiation Processing at Energies Between 300 keV and 25 MeV
- 51650 Practice for Use of a Cellulose Triacetate Dosimetry System
- 51702 Practice for Dosimetry in a Gamma Facility for Radiation Processing
- 51707 Guide for Estimating Uncertainties in Dosimetry for Radiation Processing
- 51818 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies Between 80 and 300 keV
- 51900 Guide for Dosimetry in Radiation Research on Food and Agricultural Products
- 51939 Practice for Blood Irradiation Dosimetry
- 51940 Guide for Dosimetry for Sterile Insect Release Programs
- 51956 Practice for Thermoluminescence-Dosimetry (TLD) Systems for Radiation Processing
- 52116 Practice for Dosimetry for a Self-Contained Dry-Storage Gamma Irradiator
- 52701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

¹ This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.01 on Dosimetry, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved July 20, 2013. Published November 2013. Originally published as ASTM E2628–09. Last previous ASTM edition E2628–09^{e1}. The present International Standard ISO/ASTM 52628–2013(E) replaces ASTM E2628–09^{e1}.

² For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.



2.3 ISO Standards:³

ISO 11137-1 Sterilization of health care products – Radiation – Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices

ISO 11137-3 Sterilization of health care products – Radiation – Part 3: Guidance on dosimetric aspects

ISO 10012 Measurement managements systems – Requirements for measurement processes and measuring equipment

ISO/IEC 17025 General requirements for the competence of testing and calibration laboratories

2.4 International Commission on Radiation Units and Measurements (ICRU) Reports:⁴

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

2.5 Joint Committee for Guides in Metrology (JCGM) Reports:

JCGM 100:2008, GUM, 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement⁵

JCGM 100:2008, VIM, International vocabulary of metrology – basis and general concepts and associated terms⁶

3. Terminology

3.1 Definitions:

3.1.1 *absorbed dose (D)*—quotient of $d\bar{\epsilon}$ by dm , where $d\bar{\epsilon}$ is the mean energy imparted by ionizing radiation to matter of mass dm , thus

$$D = d\bar{\epsilon}/dm$$

ICRU 85a

3.1.1.1 *Discussion*—The SI unit of absorbed dose is the gray (Gy), where 1 gray is equivalent to the absorption of 1 joule per kilogram of the specified material (1 Gy = 1 J/kg).

3.1.2 *accredited dosimetry calibration laboratory*—dosimetry laboratory with formal recognition by an accrediting organization that the dosimetry laboratory is competent to carry out specific activities which lead to the calibration or calibration verification of dosimetry systems in accordance with documented requirements of the accrediting organization.

3.1.3 *calibration*—set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or a reference material, and the corresponding values realized by standards.

3.1.4 *dosimeter*—device that, when irradiated, exhibits a quantifiable change that can be related to absorbed dose in a given material using appropriate measurement instruments and procedures.

3.1.5 *dosimeter/dosimetry system characterization*—determination of performance characteristics, such as dose range, reproducibility and the effect of influence quantities, for a dosimeter/dosimetry system under defined test conditions.

3.1.6 *dosimeter response*—reproducible, quantifiable effect produced in the dosimeter by ionizing radiation.

3.1.6.1 *Discussion*—The dosimeter response value, obtained from one or more measurements, is used in the estimation of the derived absorbed dose. The response value may be obtained from such measurements as optical absorbance, thickness, mass peak-to-peak distance in EPR spectra, or electropotential between solutions.

3.1.7 *dosimetry*—measurement of absorbed dose by the use of a dosimetry system.

3.1.8 *dosimetry system*—system used for measuring absorbed dose, consisting of dosimeters, measurement instruments and their associated reference standards, and procedures for the system's use.

3.1.9 *influence quantity*—quantity that, in a direct measurement, does not affect the quantity that is actually measured, but affects the relation between the indication and the measurement result.

VIM

3.1.10 *measurement management system*—set of interrelated or interacting elements necessary to achieve metrological confirmation and continual control of measurement processes.

3.1.11 *primary standard dosimetry system*—dosimetry system that is designated or widely acknowledged as having the highest metrological qualities and whose value is accepted without reference to other standards of the same quantity.

3.1.12 *radiation processing*—intentional irradiation of products or materials to preserve, modify or improve their characteristics.

3.1.13 *reference standard dosimetry system*—dosimetry system, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.14 *reference standard radiation field*—calibrated radiation field, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.15 *response function*—mathematical representation of the relationship between dosimeter response and absorbed dose for a given dosimetry system.

3.1.16 *routine dosimetry system*—dosimetry system calibrated against a reference standard dosimetry system and used for routine absorbed dose measurements, including dose mapping and process monitoring.

3.1.17 *traceability*—property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties.

³ Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, Case postale 56, CH-1211, Geneva 20, Switzerland, <http://www.iso.ch>.

⁴ Available from the International Commission on Radiation Units and Measurements, 7910 Woodmont Ave, Suite 800, Bethesda, MD 20815, USA.

⁵ Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

⁶ Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM/WG 2). Available free of charge at the BIPM website (<http://www.bipm.org>).