

This document is a Preview generated by EVS

Meditiiniseadmete steriliseerimine. Niiske kuumusega steriliseerimise valideerimine ja rutiinkontroll

Sterilization of health care products - Moist heat - Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

| | |
|--|---|
| <p>Käesolev Eesti standard EVS-EN ISO 17665-1:2006 sisaldab Euroopa standardi EN ISO 17665-1:2006 ingliskeelset teksti.</p> | <p>This Estonian standard EVS-EN ISO 17665-1:2006 consists of the English text of the European standard EN ISO 17665-1:2006.</p> |
| <p>Standard on kinnitatud Eesti Standardikeskuse 20.09.2006 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.</p> | <p>This standard is ratified with the order of Estonian Centre for Standardisation dated 20.09.2006 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.</p> |
| <p>Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on .</p> | <p>Date of Availability of the European standard text .</p> |
| <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p> | <p>The standard is available from Estonian standardisation organisation.</p> |

ICS 11.080.01

Võtmesõnad: inspekteerimine, kvalifikatsioon, meditsiiniseadmed, määratlused, sterilisaatorid, steriliseerimine, veeaur

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

English Version

**Sterilization of health care products - Moist heat - Part 1:
Requirements for the development, validation and routine
control of a sterilization process for medical devices (ISO 17665-
1:2006)**

Stérilisation des produits de santé - Chaleur humide -
Partie 1: Exigences pour le développement, la validation et
le contrôle de routine d'un procédé de stérilisation des
dispositifs médicaux (ISO 17665-1:2006)

Sterilisation von Produkten für die Gesundheitsfürsorge -
Feuchte Hitze - Teil 1: Anforderungen an die Entwicklung,
Validierung und Lenkung der Anwendung eines
Sterilisationsverfahrens für Medizinprodukte (ISO 17665-
1:2006)

This European Standard was approved by CEN on 14 July 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

This document (EN ISO 17665-1:2006) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2007, and conflicting national standards shall be withdrawn at the latest by August 2009.

This document supersedes EN 554:1994.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 17665-1:2006 has been approved by CEN as EN ISO 17665-1:2006 without any modifications.

ANNEX ZA

(informative)

Relationship between this European Standard and the Essential Requirements of EU Directives 90/385/EEC, 93/42/EEC and 98/79/EEC

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive EU Directives 90/385/EEC, 93/42/EEC and 98/79/EC.

Once this European Standard is cited in the Official Journal of the European Communities under that Directive and has been implemented as a national standard in at least one Member State, compliance with the normative clauses of this standard, as shown in Table ZA.1, confers, within the limits of the scope of this standard, a presumption of conformity with the relevant Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 – Correspondence between this European Standard and Directives 90/385/EEC, 93/42/EEC and 98/79/EC

| Clause(s)/Sub-clause(s) of this EN | Essential Requirements (ERs) of Directive 90/385/EEC | Essential Requirements (ERs) of Directive 93/42/EEC | Essential Requirements (ERs) of Directive 98/79/EC | Qualifying remarks/ Notes |
|------------------------------------|--|---|--|---------------------------|
| 4, 5, 6, 7, 8, 9, 10, 11, 12 | 7 | 8.3 | B.2.3 | In part |
| 4, 5, 6, 7, 8, 9, 10, 11, 12 | | 8.4 | B.2.4 | In part |

WARNING: Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this European Standard.

**Sterilization of health care products —
Moist heat —**

Part 1:
**Requirements for the development,
validation and routine control of a
sterilization process for medical devices**

Stérilisation des produits de santé — Chaleur humide —

*Partie 1: Exigences pour le développement, la validation et le contrôle
de routine d'un procédé de stérilisation des dispositifs médicaux*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

| | |
|--|-----------|
| Foreword..... | v |
| Introduction..... | vi |
| 1 Scope..... | 1 |
| 1.1 Inclusions..... | 1 |
| 1.2 Exclusions..... | 1 |
| 2 Normative references..... | 2 |
| 3 Terms and definitions..... | 3 |
| 4 Quality management system elements..... | 10 |
| 4.1 Documentation..... | 10 |
| 4.2 Management responsibility..... | 10 |
| 4.3 Product realization..... | 10 |
| 4.4 Measurement, analysis and improvement — Control of non-conforming product..... | 10 |
| 5 Sterilizing agent characterization..... | 11 |
| 5.1 Sterilizing agent..... | 11 |
| 5.2 Microbicidal effectiveness..... | 11 |
| 5.3 Materials effects..... | 11 |
| 5.4 Environmental consideration..... | 11 |
| 6 Process and equipment characterization..... | 11 |
| 6.1 Process..... | 11 |
| 6.1.1 General..... | 11 |
| 6.1.2 Saturated steam processes..... | 12 |
| 6.1.3 Contained product processes..... | 12 |
| 6.2 Equipment..... | 13 |
| 7 Product definition..... | 14 |
| 8 Process definition..... | 15 |
| 9 Validation..... | 17 |
| 9.1 General..... | 17 |
| 9.2 Installation qualification (IQ)..... | 17 |
| 9.2.1 Equipment..... | 17 |
| 9.2.2 Installation..... | 17 |
| 9.2.3 Function..... | 17 |
| 9.3 Operational qualification (OQ)..... | 18 |
| 9.4 Performance qualification (PQ)..... | 18 |
| 9.5 Review and approval of validation..... | 19 |
| 10 Routine monitoring and control..... | 20 |
| 11 Product release from sterilization..... | 21 |
| 12 Maintaining process effectiveness..... | 21 |
| 12.1 Demonstration of continued effectiveness..... | 21 |
| 12.2 Recalibration..... | 21 |
| 12.3 Maintenance of equipment..... | 21 |
| 12.4 Requalification..... | 22 |
| 12.5 Assessment of change..... | 22 |
| Annex A (informative) Guidance..... | 23 |
| Annex B (informative) Process definition based on inactivation of the microbial population in its natural state (bioburden-based method)..... | 27 |

| | |
|---|-----------|
| Annex C (informative) Process definition based on the inactivation of a reference microorganism and a knowledge of bioburden on product items to be sterilized (combined bioburden/biological indicator based method)..... | 28 |
| Annex D (informative) Conservative process definition based on inactivation of reference microorganisms (overkill method)..... | 29 |
| Annex E (informative) Operating cycles..... | 31 |
| Bibliography | 36 |

document is a preview generated by EVS

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 17665-1 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This first edition of ISO 17665-1 cancels and replaces ISO 11134:1994 and ISO 13683:1997 both of which have been technically revised.

ISO 17665 consists of the following parts, under the general title *Sterilization of health care products — Moist heat*:

- *Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*
- *Part 2: Guidance on the application of ISO 17665-1*

Introduction

A sterile medical device is one which is free of viable microorganisms. International standards that specify requirements for validation and routine control of sterilization processes require, when it is necessary to supply a sterile medical device, that adventitious microbiological contamination of a medical device prior to sterilization be minimized. Even so, medical devices produced under standard manufacturing conditions in accordance with the requirements for quality management systems (see, for example, ISO 13485) may, prior to sterilization, have microorganisms on them, albeit in low numbers. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize medical devices generally can best be described by an exponential relationship between the number of microorganisms surviving and the extent of treatment with the sterilizing agent; inevitably this means that there is always a finite probability that a microorganism may survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the organisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization processing cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product item.

ISO 17665 describes requirements that, if met, will provide a moist heat sterilization process intended to sterilize medical devices, which has appropriate microbicidal activity. Furthermore, compliance with the requirements ensures this activity is both reliable and reproducible so that predictions can be made, with reasonable confidence, that there is a low level of probability of there being a viable microorganism present on product after sterilization. Specification of this probability is a matter for regulatory authorities and may vary from country to country (see, for example, EN 556-1 and ANSI/AAMI ST67).

Generic requirements of the quality management system for design and development, production, installation and servicing are given in ISO 9001 and particular requirements for quality management systems for medical device production are given in ISO 13485. The standards for quality management systems recognise that, for certain processes used in manufacturing, the effectiveness of the process cannot be fully verified by subsequent inspection and testing of the product. Sterilization is an example of such a process. For this reason, sterilization processes are validated for use, the performance of the sterilization process is monitored routinely and the equipment is maintained.

Exposure to a properly validated, accurately controlled sterilization process is not the only factor associated with the provision of reliable assurance that the product is sterile and, in this regard, suitable for its intended use. Attention is therefore given to a number of factors including:

- a) the microbiological status of incoming raw materials and/or components;
- b) the validation and routine control of any cleaning and disinfection procedures used on the product;
- c) the control of the environment in which the product is manufactured, assembled and packaged;
- d) the control of equipment and processes;
- e) the control of personnel and their hygiene;
- f) the manner and materials in which the product is packaged;
- g) the conditions under which product is stored.

The type of contamination on a product to be sterilized varies and this has an impact upon the effectiveness of a sterilization process. It is preferable that products that have been used in a health care setting and that are being presented for resterilization in accordance with the manufacturer's instructions (see ISO 17664) be regarded as special cases. There is the potential for such products to possess a wide range of contaminating microorganisms and residual inorganic and/or organic contamination in spite of the application of a cleaning process. Hence, particular attention has to be given to the validation and control of the cleaning and disinfection processes used during reprocessing.

This part of ISO 17665 describes the requirements for ensuring that the activities associated with the process of moist heat sterilization are performed properly. These activities are described in documented work programmes designed to demonstrate that the moist heat sterilization process will consistently yield sterile products on treatment with process variables falling within the predetermined limits.

The requirements are the normative parts of this part of ISO 17665 with which compliance is claimed. The guidance given in the informative annexes is not normative and is not provided as a checklist for auditors. The guidance provides explanations and methods that are regarded as being suitable means for complying with the requirements. Methods other than those given in the guidance may be used, if they are effective in achieving compliance with the requirements of this part of ISO 17665.

The development, validation and routine control of a sterilization process comprise a number of discrete but interrelated activities, e.g. calibration, maintenance, product definition, process definition, installation qualification, operational qualification and performance qualification. While the activities required by this part of ISO 17665 have been grouped together and are presented in a particular order, this part of ISO 17665 does not require that the activities be performed in the order that they are presented. The activities required are not necessarily sequential, as the programme of development and validation may be iterative. It is possible that performing these different activities will involve a number of separate individuals and/or organizations, each of whom undertake one or more of these activities. This part of ISO 17665 does not specify the particular individuals or organizations to carry out the activities.

Sterilization of health care products — Moist heat —

Part 1:

Requirements for the development, validation and routine control of a sterilization process for medical devices

1 Scope

1.1 Inclusions

1.1.1 This part of ISO 17665 specifies requirements for the development, validation and routine control of a moist heat sterilization process for medical devices.

NOTE Although the scope of this part of ISO 17665 is limited to medical devices, it specifies requirements and provides guidance that may be applicable to other health care products.

1.1.2 Moist heat sterilization processes covered by this part of ISO 17665 include but are not limited to:

- a) saturated steam venting systems;
- b) saturated steam active air removal systems;
- c) air steam mixtures;
- d) water spray;
- e) water immersion.

NOTE See also Annex E.

1.2 Exclusions

1.2.1 This part of ISO 17665 does not specify requirements for development, validation, and routine control of a process for inactivating the causative agents of spongiform encephalopathies such as scrapie, bovine spongiform encephalopathy and Creutzfeldt-Jakob disease. Specific recommendations have been produced in particular countries for the processing of materials potentially contaminated with these agents.

NOTE See also ISO 22442-1, ISO 22442-2 and ISO 22442-3.

1.2.2 This part of ISO 17665 does not apply to those sterilization processes that are based on a combination of moist heat with other biocidal agents (e.g. formaldehyde) as the sterilizing agent.

1.2.3 This part of ISO 17665 does not detail a specified requirement for designating a medical device as “sterile.”

NOTE Attention is drawn to national or regional requirements for designating medical devices as “sterile.” See, for example, EN 556-1 or ANSI/AAMI ST67.

1.2.4 This part of ISO 17665 does not specify a quality management system for the control of all stages of production of medical devices.

NOTE It is not a requirement of this part of ISO 17665 to have a complete quality management system during manufacture, but the elements of a quality management system that are the minimum necessary to control the sterilization process are normatively referenced at appropriate places in the text (see, in particular, Clause 4). Attention is drawn to the standards for quality management systems (see ISO 13485) that control all stages of production of medical devices, including the sterilization process. Regional and national regulations for the provision of medical devices might require implementation of a complete quality management system and the assessment of that system by a third party.

1.2.5 This part of ISO 17665 does not specify requirements for occupational safety associated with the design and operation of moist heat sterilization facilities.

NOTE Requirements for operational safety are specified in IEC 61010-2-040. Additionally, safety regulations exist in some countries.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10012, *Measurement management systems — Requirements for measurement processes and measuring equipment*

ISO 11138-1, *Sterilization of health care products — Biological indicators — Part 1: General Requirements*

ISO 11138-3, *Sterilization of health care products — Biological indicators — Part 3: Biological indicators for moist heat sterilization processes*

ISO 11140-1, *Sterilization of health care products — Chemical indicators — Part 1: General requirements*

ISO 11140-3, *Sterilization of health care products — Chemical indicators — Part 3: Class 2 indicator systems for use in the Bowie and Dick steam penetration test*

ISO 11140-4, *Sterilization of health care products — Chemical indicators — Part 4: Class 2 indicators as an alternative to Bowie and Dick test for detection of steam penetration*

ISO 11140-5, *Sterilization of health care products — Chemical indicators — Part 5: Class 2 indicators for Bowie and Dick air removal test sheets and packs*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 11737-1, *Sterilization of medical devices — Microbiological methods — Part 1: Determination of a population of microorganisms on products*

ISO 11737-2, *Sterilization of medical devices — Microbiological methods — Part 2: Tests of sterility performed in the validation of a sterilization process*

ISO 13485:2003, *Medical devices — Quality management systems — Requirements for regulatory purposes*

ISO 17664, *Sterilization of medical devices — Information to be provided by the manufacturer for the processing of resterilizable medical devices*