

SOOJUSARVESTID. OSA 1: ÜLDNÕUDED

Heat meters - Part 1: General requirements

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 1434-1:2015 sisaldab Euroopa standardi EN 1434-1:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 1434-1:2015 consists of the English text of the European standard EN 1434-1:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 04.11.2015.	Date of Availability of the European standard is 04.11.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 17.200.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Heat meters - Part 1: General requirements

Compteurs d'énergie thermique - Partie 1:
Prescriptions générales

Wärmezähler - Teil 1: Allgemeine Anforderungen

This European Standard was approved by CEN on 5 September 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	4
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	6
4 Types of instruments	11
4.1 General.....	11
4.2 Complete instrument.....	11
4.3 Combined instrument.....	11
4.4 Hybrid instrument.....	11
4.5 Sub-assemblies of a heat meter, which is a combined instrument	11
4.5.1 General.....	11
4.5.2 Flow sensor	12
4.5.3 Temperature sensor pair	12
4.5.4 Calculator	12
4.6 Equipment under test (EUT)	12
5 Rated operating conditions	12
5.1 Limits of temperature range	12
5.2 Limits of temperature differences	12
5.3 Limits of flow rate	12
5.4 Limit of thermal power	13
5.5 Limits of working pressure (PS and $P_{\min}$).....	13
5.6 Nominal pressure (PN).....	13
5.7 Limits in ambient temperature.....	13
5.8 Limits in deviations in supply voltage	13
5.9 Maximum pressure loss.....	13
5.10 Specific requirements on registration devices.....	13
5.10.1 General.....	13
5.10.2 Suitability	13
5.10.3 Rated operated conditions	14
5.10.4 Indication	14
5.10.5 MPE.....	15
6 Technical characteristics	15
6.1 Materials and construction.....	15
6.2 Requirements outside the limiting values of the flow rate.....	16
6.3 Display.....	16
6.4 Protection against fraud	16
6.5 Supply voltage	17
6.6 Qualifying immersion depth of a temperature sensor	17
6.7 The influence on a temperature sensor pair caused by mounting in pockets.....	17
6.8 Reproducibility	17
6.9 Repeatability.....	17
6.10 Software.....	18
7 Specified working range.....	18
7.1 General.....	18

7.2	Temperature difference	18
7.3	Flow rate	18
8	Heat transmission formula	18
9	Metrological characteristics (Maximum Permissible Error, MPE)	19
9.1	General	19
9.2	Values of maximum permissible errors	19
9.2.1	Maximum permissible relative errors of complete heat meters	19
9.2.2	Maximum permissible relative error of sub-assemblies	19
9.3	Application of maximum permissible errors	20
10	Environmental classification	20
10.1	General	20
10.2	Environmental class A (Domestic use, indoor installations)	21
10.3	Environmental class B (Domestic use, outdoor installations)	21
10.4	Environmental class C (Industrial installations)	21
10.5	Mechanical classes M1 to M3	21
11	Heat meter specification	21
11.1	General	21
11.2	Flow sensor	21
11.3	Temperature sensor pair	23
11.4	Calculator	23
11.5	Complete meters	25
12	Information to be made available by the manufacturer or supplier	26
12.1	Installation instructions	26
12.2	Parameter setting instructions	27
12.3	Adjustment instructions	28
12.4	Maintenance instructions	28
12.5	Hints for disposal instructions	29
Annex A (normative)	Heat coefficient equations	30
Annex B (normative)	Flow conditioner package	32
Annex C (normative)	Fast response meters	34
Annex ZA (informative)	Relationship between this European Standard and the Essential Requirements of EU Directive 2004/22/EC, MID	35
Bibliography		37

European foreword

This document (EN 1434-1:2015) has been prepared by Technical Committee CEN/TC 176 “Heat meters”, the secretariat of which is held by SIS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2016, and conflicting national standards shall be withdrawn at the latest by May 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1434-1:2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

EN 1434, *Heat meters* consists of the following parts:

- *Part 1: General requirements*
- *Part 2: Constructional requirements*
- *Part 3: Data exchange and interfaces¹⁾*
- *Part 4: Pattern approval tests*
- *Part 5: Initial verification tests*
- *Part 6: Installation, commissioning, operational monitoring and maintenance*

In comparison to EN 1434-1:2007, the following changes have been made:

- special cases for combined cooling and heating meters are added;
- additional functionality for smart metering applications are added;
- metrological requirements for smart metering applications are added;
- definitions and requirements for the cooling meter are added;
- tariff meters are added;
- terms and definitions, requirements for registration devices and cooling meters are added;
- requirements for fast response meters are added (informative Annex C).

¹⁾ EN 1434-3 is maintained by CEN/TC 294.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

This document is a preview generated by EVS

1 Scope

This European Standard specifies the general requirements for heat meters. Heat meters are instruments intended for measuring the energy which in a heat-exchange circuit is absorbed (cooling) or given up (heating) by a liquid called the heat-conveying liquid. The heat meter indicates the quantity of heat in legal units.

Electrical safety requirements are not covered by this European Standard.

Pressure safety requirements are not covered by this European Standard.

Surface mounted temperature sensors are not covered by this European Standard.

This standard covers meters for closed systems only, where the differential pressure over the thermal load is limited.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1434-2:2015, *Heat meters — Part 2: Constructional requirements*

EN 1434-4:2015, *Heat meters — Part 4: Pattern approval test*

EN 60751, *Industrial platinum resistance thermometers and platinum temperature sensors (IEC 60751)*

EN 61010-1, *Safety requirements for electrical equipment for measurement, control and laboratory use — Part 1: General requirements (IEC 61010-1)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

response time

$\tau_{0,5}$

time interval between the instant when flow or temperature difference is subjected to a specified abrupt change and the instant when the response reaches 50 % of the step value

3.2

fast response meter

meter suitable for heat exchanging circuits with rapid dynamic variations in the exchanged heat

Note 1 to entry: See also Annex C.

3.3

rated voltage

U_n

voltage of the external power supply required to operate the heat meter, conventionally the voltage of the AC mains supply

3.4

rated operating conditions

conditions of use, giving the range of values of influence quantities, for which the metrological characteristics of the instrument are within the specified maximum permissible errors