

SOOJUSARVESTID. OSA 4: MUDELI TÜÜBIKATSED

Heat meters - Part 4: Pattern approval tests

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 1434-4:2015 sisaldab Euroopa standardi EN 1434-4:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 1434-4:2015 consists of the English text of the European standard EN 1434-4:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 04.11.2015.	Date of Availability of the European standard is 04.11.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 17.200.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Heat meters - Part 4: Pattern approval tests

Compteurs d'énergie thermique - Partie 4: Essais en
vue de l'approbation de modèle

Wärmezähler - Teil 4: Prüfungen für die
Bauartzulassung

This European Standard was approved by CEN on 5 September 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword	4
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
4 General	7
5 Requirements	7
6 Specification of operating conditions	7
6.1 Rated operating conditions.....	7
6.2 Reference conditions.....	7
6.3 Reference values for the measurand, RVM.....	8
7 Tests and measurements	8
7.1 General.....	8
7.2 Test programme.....	9
7.3 Uncertainty of test equipment and influences of EUT	11
7.4 Performance tests.....	11
7.5 Dry heat	16
7.6 Cold.....	17
7.7 Static deviations in supply voltage	18
7.8 Durability test.....	19
7.9 Damp heat.....	22
7.10 Short time mains voltage reduction	24
7.11 Electrical transients	24
7.12 Electromagnetic field.....	26
7.13 Electromagnetic field specifically caused by digital radio equipment.....	27
7.14 Radio frequency amplitude modulated	28
7.15 Electrostatic discharge.....	30
7.16 Static magnetic field (fraud protection)	30
7.17 Mains frequency magnetic field.....	31
7.18 Internal pressure	31
7.19 Pressure loss.....	31
7.20 Electromagnetic emission.....	32
7.21 24 h interruption in the mains power supply voltage.....	32
7.22 Flow disturbances.....	33
7.23 Vibration/mechanical shock.....	33
8 Documentation	34
Annex A (informative) Testing procedure for temperature sensor pairs with pockets and without pockets	35
A.1 Test set-up	35
A.2 Test sequence	37
A.3 Calculations.....	37
Annex B (informative) Checklist for type approvals of heat meters according to EN 1434	38

Annex C (informative) Criteria for a fully developed flow profile	52
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 2004/22/EC, MID.....	53
Bibliography	54

This document is a preview generated by EVS

European foreword

This document (EN 1434-4:2015) has been prepared by Technical Committee CEN/TC 176 “Heat meters”, the secretariat of which is held by SIS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2016, and conflicting national standards shall be withdrawn at the latest by May 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1434-4:2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

EN 1434, *Heat meters* consists of the following parts:

- *Part 1: General requirements*
- *Part 2: Constructional requirements*
- *Part 3: Data exchange and interfaces¹⁾*
- *Part 4: Pattern approval tests*
- *Part 5: Initial verification tests*
- *Part 6: Installation, commissioning, operational monitoring and maintenance*

In comparison to EN 1434-4:2007, the following changes have been made:

- metrological requirements for smart metering applications are added;
- additional functionalities for smart metering applications are added;
- cooling meters are added;
- influences of sensors are added;
- tests for cooling applications and for fast response meters are added;
- test for additional functionalities for smart metering applications, e.g. internal clock, external digital signal, absolute temperature are added;
- calculator with single temperature sensor are added;

¹⁾ EN 1434-3 is maintained by CEN/TC 294.

- test for communication interfaces, endurance test for flow sensors and accelerated durability test are added;
- electromagnetic field caused by digital radio equipment;
- static magnetic field;
- test procedure for temperature sensor pairs with pockets and without pockets.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies pattern approval tests for heat meters. Heat meters are instruments intended for measuring the energy which in a heat-exchange circuit is absorbed (cooling) or given up (heating) by a liquid called the heat-conveying liquid. The heat meter indicates the quantity of heat in legal units.

Electrical safety requirements are not covered by this European Standard.

Pressure safety requirements are not covered by this European Standard.

Surface mounted temperature sensors are not covered by this European Standard.

This standard covers meters for closed systems only, where the differential pressure over the thermal load is limited.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1434-1:2015, *Heat meters — Part 1: General requirements*

EN 55022, *Information technology equipment — Radio disturbance characteristics — Limits and methods of measurement (CISPR 22:2008)*

EN 60068-2-1, *Environmental testing — Part 2-1: Tests - Test A: Cold (IEC 60068-2-1)*

EN 60068-2-2, *Environmental testing — Part 2-2: Tests - Test B: Dry heat (IEC 60068-2-2)*

EN 60068-2-30, *Environmental testing — Part 2-30: Tests - Test Db: Damp heat, cyclic (12 h + 12 h cycle) (IEC 60068-2-30)*

EN 60751:2008, *Industrial platinum resistance thermometers and platinum temperature sensors (IEC 60751:2008)*

EN 61000-4-2, *Electromagnetic compatibility (EMC) — Part 4-2: Testing and measurement techniques — Electrostatic discharge immunity test (IEC 61000-4-2)*

EN 61000-4-3, *Electromagnetic compatibility (EMC) — Part 4-3: Testing and measurement techniques — Radiated, radio-frequency, electromagnetic field immunity test (IEC 61000-4-3)*

EN 61000-4-4, *Electromagnetic compatibility (EMC) — Part 4-4: Testing and measurement techniques — Electrical fast transient/burst immunity test (IEC 61000-4-4)*

EN 61000-4-5, *Electromagnetic compatibility (EMC) — Part 4-5: Testing and measurement techniques — Section 5: Surge immunity test (IEC 61000-4-5) (IEC 61000-4-5)*

EN 61000-4-6:2014, *Electromagnetic compatibility (EMC) — Part 4-6: Testing and measurement techniques — Immunity to conducted disturbances, induced by radio-frequency fields (IEC 61000-4-6:2013)*

EN 61000-4-8, *Electromagnetic compatibility (EMC) — Part 4-8: Testing and measurement techniques — Power frequency magnetic field immunity test (IEC 61000-4-8)*

EN 61000-4-11, *Electromagnetic compatibility (EMC) — Part 4-11: Testing and measurement techniques — Voltage dips, short interruptions and voltage variations immunity tests (IEC 61000-4-11)*

EN ISO 4064-2, *Water meters for cold potable water and hot water — Part 2: Test methods (ISO 4064-2)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 1434-1:2015 apply.

4 General

The procedure shall ascertain that the pattern conforms to the metrological requirements of this European Standard. In addition to the checking of the documentation (Clause 7) and the comparison of the pattern with the metrological requirements of this European Standard, the tests in Clause 6 shall be performed.

It is recommended to use a checklist as in Annex B to report in a standardized way the result of the comparison between the patterns under approval with the essential requirements of this European Standard.

5 Requirements

Under normal operating conditions, the error of heat meters or their sub-assemblies shall not exceed the maximum permissible error, MPE specified in EN 1434-1.

When heat meters or their sub-assemblies are exposed to disturbances, significant faults shall not occur.

6 Specification of operating conditions

6.1 Rated operating conditions

The rated operating conditions are those given in Table 1.

Table 1 — Rated operating conditions

Environmental class	A	B	C
Ambient temperature in °C	+5 to +55	−25 to +55	+5 to +55
Relative humidity in %	< 93		
Mains supply voltage in V	195 V to 253 V		
Mains frequency	$f_{\text{nom}} \pm 2 \%$		
Battery voltage	The voltage of a battery in service under normal conditions		
Remote AC supply voltage	12 V to 36 V		
Remote DC supply voltage	12 V to 42 V		
Local external DC supply voltage	as specified by manufacturer		

6.2 Reference conditions

Range of ambient temperature: +15 °C to +35 °C