

**Milk - Quantitative determination of  
bacteriological quality - Guidance for  
establishing and verifying a conversion  
relationship between routine method results  
and anchor method results**

Milk - Quantitative determination of bacteriological  
quality - Guidance for establishing and verifying a  
conversion relationship between routine method  
results and anchor method results

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 21187:2005 sisaldab Euroopa standardi EN ISO 21187:2005 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 28.12.2005 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 21187:2005 consists of the English text of the European standard EN ISO 21187:2005.</p> <p>This document is endorsed on 28.12.2005 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p><b>Käsitlusala:</b> This International Standard gives guidelines for the establishment of a conversion relationship between the results of a routine method and an anchor method, and its verification for the quantitative determination of the bacteriological quality of milk.</p>	<p><b>Scope:</b> This International Standard gives guidelines for the establishment of a conversion relationship between the results of a routine method and an anchor method, and its verification for the quantitative determination of the bacteriological quality of milk.</p>
--	--

**ICS** 07.100.30, 67.100.01

**Võtmesõnad:**

English Version

Milk - Quantitative determination of bacteriological quality -  
Guidance for establishing and verifying a conversion relationship  
between routine method results and anchor method results (ISO  
21187:2004)

Lait - Mesure quantitative de la qualité bactériologique -  
Lignes directrices pour établir et vérifier une relation de  
conversion entre les résultats de la méthode de routine et  
les résultats de la méthode d'ancrage (ISO 21187:2004)

Milch - Quantitative Bestimmung der bakteriologischen  
Qualität - Leitfaden für die Erarbeitung einer  
Übertragungsbeziehung zwischen den Messwerten von  
Routine- und Bezugsverfahren sowie deren Verifizierung  
(ISO 21187:2004)

This European Standard was approved by CEN on 30 September 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

## Foreword

The text of ISO 21187:2004 has been prepared by Technical Committee ISO/TC 34 "Agricultural food products" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 21187:2005 by Technical Committee CEN/TC 302 "Milk and milk products - Methods of sampling and analysis", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2006, and conflicting national standards shall be withdrawn at the latest by April 2006.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

## Endorsement notice

The text of ISO 21187:2004 has been approved by CEN as EN ISO 21187:2005 without any modifications.

---

---

**Milk — Quantitative determination of  
bacteriological quality — Guidance for  
establishing and verifying a conversion  
relationship between routine method  
results and anchor method results**

*Lait — Mesure quantitative de la qualité bactériologique — Lignes  
directrices pour établir et vérifier une relation de conversion entre  
les résultats de la méthode de routine et les résultats de la méthode  
d'ancrage*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor the IDF accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IDF national committees. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat at the address given below.

© ISO and IDF 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO or IDF at the respective address below.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

International Dairy Federation  
Diamant Building • Boulevard Auguste Reyers 80 • B-1030 Brussels  
Tel. + 32 2 733 98 88  
Fax + 32 2 733 04 13  
E-mail [info@fil-idf.org](mailto:info@fil-idf.org)  
Web [www.fil-idf.org](http://www.fil-idf.org)

Published in Switzerland

## Contents

Page

Foreword.....	iv
Introduction .....	vi
1 Scope.....	1
2 Normative references .....	1
3 Terms and definitions.....	1
4 Principles .....	2
4.1 Prerequisites .....	2
4.2 Organizational set-up .....	2
5 Establishing a conversion relationship.....	3
5.1 Consideration of influencing factors and their consequences.....	3
5.2 Test samples .....	4
5.3 Pretreatment of test samples.....	5
5.4 Analysis.....	6
5.5 Calculation .....	7
5.6 Expression of results.....	7
6 Verification of a conversion relationship .....	8
6.1 Frequency of verification .....	8
6.2 Test samples .....	8
6.3 Pretreatment of test samples.....	8
6.4 Analysis.....	8
6.5 Calculation and verification of a conversion relationship.....	8
7 Test report.....	9
Annex A (informative) Number of samples for linear regression .....	10
Bibliography .....	13

## Foreword

**ISO (the International Organization for Standardization)** is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21187|IDF 196 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.



## Foreword

**IDF (the International Dairy Federation)** is a worldwide federation of the dairy sector with a National Committee in every member country. Every National Committee has the right to be represented on the IDF Standing Committees carrying out the technical work. IDF collaborates with ISO and AOAC International in the development of standard methods of analysis and sampling for milk and milk products.

Draft International Standards adopted by the Action Teams and Standing Committees are circulated to the National Committees for voting. Publication as an International Standard requires approval by at least 50 % of the National Committees casting a vote.

ISO 21187|IDF 196 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

All work was carried out by the Joint ISO/IDF/AOAC Action Team, *Routine analysis in quantitative microbiology*, of the Standing Committee on *Quality assurance, statistics of analytical data and sampling*, under the aegis of its project leader, Mr H.J.C.M. van den Bijgaart (NL).

## Introduction

Conversion in quantitative microbiology means expressing the result of a quantitative determination of the bacteriological status of a test sample as obtained with a routine method in units of another method, generally a reference or anchor method. Through this, quantitative results obtained with routine methods can be compared to values or limits that are stated in reference or anchor method units. For establishing and applying a conversion relationship, a number of prerequisites should be met. These are referred to in this International Standard, but are generally described elsewhere.

Although a considerable part of the applied principles for conversion coincides with those applied for the calibration of indirect or routine methods against a reference method, or by means of (certified) reference materials, it is stressed that the background and aims for applying conversion are different from those for calibration. Calibration involves the determination of the adjustment needed for each level of an analyte to closely approximate the true value of its concentration or number. However in quantitative microbiology, a true value in its strict sense cannot be established and is only defined by the method description applied. When applying routine methods in the quantitative determination of bacteriological quality, one is often dealing with different methodological principles and therefore also other units. Conversion is used to transfer results obtained with different methods to a common scale.

# Milk — Quantitative determination of bacteriological quality — Guidance for establishing and verifying a conversion relationship between routine method results and anchor method results

## 1 Scope

This International Standard gives guidelines for the establishment of a conversion relationship between the results of a routine method and an anchor method, and its verification for the quantitative determination of the bacteriological quality of milk.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3534-1, *Statistics — Vocabulary and symbols — Part 1: Probability and general statistical terms*

ISO 8196-1|IDF 128-1, *Milk — Definition and evaluation of the overall accuracy of indirect methods of milk analysis — Part 1: Analytical attributes of indirect methods*

ISO 8196-2|IDF 128-2, *Milk — Definition and evaluation of the overall accuracy of indirect methods of milk analysis — Part 2: Calibration and quality control in the dairy laboratory*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 3534-1, ISO 8196-1|IDF 128-1, ISO 8196-2|IDF 128-2 and the following apply.

### 3.1

**routine method**

**alternative method**

method of analysis allowing quantification of the bacteriological status of a test sample

NOTE 1 The method can be proprietary or non-commercial.

NOTE 2 The term “routine” or “alternative” in this International Standard refers to the entire method. It includes all aspects (such as sample pretreatment, materials and instruments) required for the execution of the method.

NOTE 3 The term “routine method” is used in this International Standard.