

**Statsionaarsed emissiooniallikad.
Kõikide esinevate gaaside
massikontsentratsiooni määramine.
Pideva leegi ionisatsioonidetektori
meetod**

Stationary source emissions - Determination of the mass concentration of total gaseous organic carbon at low concentrations in flue gases - Continuous flame ionisation detector method

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 12619:2000 sisaldab Euroopa standardi EN 12619:1999 ingliskeelset teksti. Käesolev dokument on jõustatud 11.01.2000 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes. Standard on kättesaadav Eesti standardiorganisatsioonist.	This Estonian standard EVS-EN 12619:2000 consists of the English text of the European standard EN 12619:1999. This document is endorsed on 11.01.2000 with the notification being published in the official publication of the Estonian national standardisation organisation. The standard is available from Estonian standardisation organisation.
--	---

Käsitlusala: This European Standard specifies a set of minimum performance requirements for an instrument using Flame Ionisation Detection (FID), together with procedures for its calibration and operation, for the continuous measurement of the mass concentration of total gaseous organic carbon (TOC) in stationary source combustion emissions.	Scope: This European Standard specifies a set of minimum performance requirements for an instrument using Flame Ionisation Detection (FID), together with procedures for its calibration and operation, for the continuous measurement of the mass concentration of total gaseous organic carbon (TOC) in stationary source combustion emissions.
--	--

ICS 13.040.40

Võtmesõnad:

ICS 13.030.40; 13.040.40

English version

Stationary source emissions

**Determination of the mass concentration of total gaseous
organic carbon at low concentrations in flue gases
Continuous flame ionisation detector method**

Emissions de sources fixes –
Détermination de la concentration
massique en carbone organique total
à de faibles concentrations dans les
effluents gazeux – Méthode du
détecteur continu à ionisation de
flamme

Emissionen aus stationären Quellen –
Bestimmung der Massenkonzentration
des gesamten gasförmigen
organisch gebundenen Kohlenstoffs
in geringen Konzentrationen in
Abgasen – Kontinuierliches Verfahren
unter Verwendung eines
Flammenionisationsdetektors

This European Standard was approved by CEN on 1999-05-06.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Contents

	Page
Foreword	3
1 Scope	3
2 Normative references	4
3 Terms and definitions	4
4 The principle of the technique	5
5 Apparatus and gases	7
6 Measurement procedure	9
Annex A (informative) Recommended minimum operational requirements for long term continuous monitoring applications	12
Annex B (informative) Measurement uncertainty and associated statistics	13
Annex C (normative) Determination of the performance characteristics of a FID	15
Annex D (informative) Safety measures	18
Annex E (informative) Calculation of total organic carbon mass concentration from volume concentrations	18