
**Aseptic processing of health care
products —**

Part 7:

**Alternative processes for medical devices
and combination products**

Traitement aseptique des produits de santé —

*Partie 7: Procédés alternatifs pour les dispositifs médicaux et les
produits de combinaison*



This document is a preview generated by EVS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Quality system elements	2
5 Aseptic process definition	2
5.1 General	2
5.2 Risk management	2
6 Manufacturing environment	3
7 Equipment	3
8 Personnel	3
9 Manufacture of the product	3
10 Process simulation	3
10.1 General	3
10.2 Media selection and growth support	3
10.3 Simulation procedures	3
10.4 Incubation and inspection of process simulation units	6
10.5 Initial performance qualification	6
10.6 Periodic performance requalification	6
10.7 Repeat of initial performance qualification	7
10.8 Documentation of process simulations	7
10.9 Disposition of filled product	7
11 Test for sterility	7
11.1 General	7
11.2 Investigation of positive units from tests for sterility	7
Annex A (informative) Risk assessment for aseptic processing — Quality risk management method	8
Annex B (informative) Selection of a sample for testing for microbial contamination	15
Annex C (informative) Testing options for process simulation	16
Bibliography	19

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 13408-7 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

ISO 13408 consists of the following parts, under the general title *Aseptic processing of health care products*:

- *Part 1: General requirements*
- *Part 2: Filtration*
- *Part 3: Lyophilization*
- *Part 4: Clean-in-place technologies*
- *Part 5: Sterilization in place*
- *Part 6: Isolator systems*
- *Part 7: Alternative processes for medical devices and combination products*

Introduction

ISO 13408 is the International Standard, published in a series of parts, for aseptic processing of health care products. Historically, sterile health care products that are aseptically produced have typically been liquids, powders or suspensions that cannot be terminally sterilized. More recently, medical devices and health care products have been developed that are combined with medicinal products, including biological and viable cells, that cannot be terminally sterilized.

The application of ISO 13408-1 to these medical devices and combination products can require the development of alternative approaches to process simulation. This part of ISO 13408 specifies requirements and provides guidance for developing such alternative approaches for the qualification of aseptic processes through process simulation of medical devices and combination products that meet the requirements of ISO 13408-1.

ISO 13408-1:2008, 10.1.2 permits the use of alternative process simulation approaches, based on particular medical devices or combination products, where the substitution in full with sterile liquid media might not be possible.

Medical devices and combination products that typically require aseptic processing might include, for example, the following.

- a) Medical devices that cannot be terminally sterilized and where the process simulation approach according to ISO 13408-1 cannot be applied:
 - bioprostheses (e.g. heart valves, vascular implants);
 - biodegradable implants (e.g. hernia meshes);
 - artificial and/or non-viable biologically based matrixes;
 - extracorporeal processing devices (e.g. immuno-adsorbers);
 - implantable osmotic pumps;
 - hermetically sealed electromechanical devices and partially enclosed electronic devices (e.g. invasive and non-invasive diagnostic devices).
- b) Combination products (including viable cell-based combination products):
 - implants coated with drug and/or biologically derived substances (e.g. drug-coated stents, carrier materials with protein, bone-graft material with growth factors, biodegradable drug-coated stents);
 - wound dressings (e.g. dressings with haemostatic agents, tissue sealants, or biologics);
 - transdermal or injectable delivery systems (e.g. drug-coated or biologics interstitial patches);
 - kits containing a biological or drug component (e.g. demineralized bone matrices).

For such products, a risk management strategy and method(s) can be used for the identification, evaluation and quantification (estimation) of contamination risks throughout the entire product/process life cycle. Environmental monitoring and microbiological studies can be performed on individual steps of the process to evaluate the effectiveness of contamination controls and risk mitigations. The design of the process simulation can then be driven by the results of the risk analysis. If the results of the process simulation are acceptable, this provides evidence that the aseptic process is in a state of contamination control (i.e. no extrinsic microbiological/microbial contamination has been introduced during the aseptic process).

This part of ISO 13408 should be read in conjunction with ISO 13408-1.

Within this International Standard, text that supplements ISO 13408-1 by providing additional requirements or guidance is identified by the prefix "Addition".

Aseptic processing of health care products —

Part 7:

Alternative processes for medical devices and combination products

1 Scope

This part of ISO 13408 specifies requirements and provides guidance on alternative approaches to process simulations for the qualification of the aseptic processing of medical devices and combination products that cannot be terminally sterilized and where the process simulation approach according to ISO 13408-1 cannot be applied.

This part of ISO 13408 describes how risk assessment can be used during the development of an aseptic process to design a process simulation study for medical devices and combination products in those cases where a straightforward substitution of media for product during aseptic processing is not feasible or would not simulate the actual aseptic process.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13408-1:2008, *Aseptic processing of health care products — Part 1: General requirements*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 13408-1 and the following apply.

3.1

extrinsic contamination

ingress of material of external origin during the manufacturing process

NOTE The focus of extrinsic contamination in this part of ISO 13408 is biological agents e.g. bacteria, mould, yeast.

3.2

process simulation

exercise that simulates the manufacturing process or portions of the process in order to demonstrate the capability of the aseptic process to prevent biological contamination

3.3

risk management

systematic application of quality management policies, procedures and practices to the tasks of analysing, evaluating, controlling and monitoring risk

[ISO 14971:2007, definition 2.22]

3.4

surrogate product

item designed to represent product in process simulations and which is comparable to the actual product