

Dentistry - Metallic materials for fixed and removable restorations and appliances (ISO 22674:2016)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 22674:2016 sisaldab Euroopa standardi EN ISO 22674:2016 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 22674:2016 consists of the English text of the European standard EN ISO 22674:2016.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 24.02.2016.	Date of Availability of the European standard is 24.02.2016.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.060.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Metallic materials for fixed and removable restorations and appliances (ISO 22674:2016)

Médecine bucco-dentaire - Matériaux métalliques pour les restaurations fixes et amovibles et les appareillages (ISO 22674:2016)

Zahnheilkunde - Metallische Werkstoffe für festsitzenden und herausnehmbaren Zahnersatz und Applikationen (ISO 22674:2016)

This European Standard was approved by CEN on 24 October 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 22674:2016) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2016, and conflicting national standards shall be withdrawn at the latest by August 2016.

This document supersedes EN ISO 22674:2006.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 22674:2016 has been approved by CEN as EN ISO 22674:2016 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	3
5 Requirements	4
5.1 Chemical composition	4
5.1.1 Reported composition	4
5.1.2 Permitted deviation from the reported composition for elements	4
5.2 Hazardous elements	4
5.2.1 Hazardous elements	4
5.2.2 Limits for the hazardous elements cadmium and beryllium	4
5.2.3 Limit for lead	4
5.2.4 Nickel	5
5.3 Biocompatibility	5
5.4 Mechanical properties	5
5.4.1 General	5
5.4.2 Proof stress of 0,2 % non-proportional extension	6
5.4.3 Elongation after fracture	6
5.5 Elastic modulus	7
5.5.1 Precision of measurement method	7
5.5.2 Determination of compliance with the requirements for Type 5 materials	7
5.5.3 Mean value	8
5.6 Density	8
5.7 Corrosion resistance for material integrity	8
5.8 Tarnish resistance	8
5.9 Solidus and liquidus temperatures (alloy) or melting point (commercially pure metal)	8
5.10 Thermal expansion coefficient	8
5.11 Information, instructions and marking	9
6 Sampling	9
7 Preparation of test specimens	9
7.1 General	9
7.2 Heat treatment	9
7.2.1 General conditions	9
7.2.2 Metallic materials for which a heat treatment is recommended in the manufacturer's instructions for use	9
7.2.3 Metallic material for metal-ceramic restorations	9
7.2.4 Metallic materials for which no heat-treatment is recommended in the manufacturer's instructions for use	9
7.3 Proof stress of 0,2 % non-proportional extension and the elongation after fracture: Metallic materials for which production of conventional specimens is possible	10
7.4 Type 0 metallic materials for which the production of conventional specimens is not possible	11
7.5 Elastic moduli	11
7.5.1 General	11
7.5.2 Tensile strain method	11
7.5.3 Flexure method (three or four point bending)	11
7.5.4 Acoustic resonance method	12
7.6 Density measurement	13
7.7 Corrosion resistance	13
7.8 Tarnish resistance and colour	13

7.9	Linear thermal expansion	13
8	Test methods	13
8.1	Information, instructions and marking	13
8.2	Chemical composition	14
8.3	Mechanical testing	14
8.3.1	Apparatus	14
8.3.2	Test procedure	14
8.4	Proof stress of 0,2 % non-proportional extension	15
8.4.1	Calculation	15
8.4.2	Mean value	15
8.5	Percentage elongation after fracture	15
8.5.1	Calculation	15
8.5.2	Mean value	15
8.6	Elastic modulus measurement	15
8.6.1	Tensile strain method	15
8.6.2	Flexure method in three- or four-point bending mode	16
8.6.3	Acoustic resonance method	18
8.7	Calculation of elasticity parameters from acoustic measurement	19
8.7.1	General	19
8.7.2	Elastic modulus	20
8.7.3	Shear modulus	20
8.7.4	Poisson ratio	21
8.8	Density	21
8.8.1	Preparation of test specimens	21
8.8.2	Reagents	21
8.8.3	Apparatus	21
8.8.4	Procedure	21
8.9	Corrosion resistance by the static immersion procedure of ISO 10271:2011	21
8.9.1	Preparation of test specimens	21
8.9.2	Reagents	21
8.9.3	Apparatus	21
8.9.4	Test solution	22
8.9.5	Test procedure	22
8.9.6	Analysis	22
8.9.7	Treatment of data	22
8.10	Sulfide tarnish test (cyclic immersion)	22
8.11	Sulfide tarnish test (static immersion)	22
8.12	Solidus and liquidus temperatures (dental casting alloys) or melting point (commercially pure metals)	22
8.12.1	Cooling curve method	22
8.12.2	Thermal analysis method	23
8.13	Linear thermal expansion	24
9	Information and instructions for use	24
9.1	Information	24
9.2	Processing instructions	25
10	Marking and labelling	25
10.1	Marking	25
10.2	Labelling of the package	25
Annex A (informative) Tensile testing of a non-cast Type 0 metallic material that is intended for use in a thickness between 0,1 mm and 0,5 mm		27
Annex B (normative) Calculation of uncertainty for elasticity measurement		30
Annex C (informative) Measurement of Poisson ratio		34
Bibliography		36

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthodontic materials*.

This second edition cancels and replaces the first edition (ISO 22674:2006), which has been technically revised with the following changes:

- Corrosion resistance measurement was referred to the recent standard ISO 10271:2011.
- A second tarnish test was included, referring to provisions in ISO 10271:2011.
- Clarification of the term “free of” was added to the requirements of composition and labelling;
- Lead was added as a hazardous element.
- Measurement of elasticity was revised. Beside the method of calculation of elastic modulus using an extensometer, other alternative methods were added, namely, the flexure method in three- and four-point bending and the acoustic resonance method.
- Informative [Annex A](#) was added, dealing with tensile testing of non-cast Type 0 metallic materials intended for use in a thickness between 0,1 and 0,5 mm.
- Normative [Annex B](#) was added, giving information on calculation of uncertainty for elastic measurement.
- Informative [Annex C](#) was added, giving information for measurement of Poisson ratio.

Introduction

Specific qualitative and quantitative requirements for freedom from biological hazard are not included in this International Standard, but it is recommended that, in assessing possible biological hazards, reference has to be made to ISO 10993-1 and ISO 7405.

Requirements for the performance of metals and alloys used for the metallic component of a metal-ceramic restoration contained in this International Standard supersede such requirements formerly contained in ISO 9693. The requirements for the performance of ceramic material and the metal-ceramic bond in metal-ceramic restorative systems continue to be specified in ISO 9693-1.

Requirements for the proof stress and minimum elongation after fracture for Type 0 metallic materials are not included in this International Standard, but it is recommended to adopt the test procedure given in [Annex A](#) when measuring these properties. Requirements will be included in a revision of this International Standard when information becomes available to Technical Committee ISO/TC 106/SC 2.

Dentistry — Metallic materials for fixed and removable restorations and appliances

1 Scope

This International Standard classifies metallic materials that are suitable for the fabrication of dental restorations and appliances, including metallic materials recommended for use either with or without a ceramic veneer, or recommended for both uses, and specifies their requirements. Furthermore, it specifies requirements with respect to packaging and marking the products and to the instructions to be supplied for the use of these materials, including products delivered for sale to a third party.

This International Standard does not apply to alloys for dental amalgam (ISO 24234), dental brazing materials (ISO 9333), or metallic materials for orthodontic appliances (ISO 15841) (e.g. wires, brackets, bands and screws).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5832-2, *Implants for surgery — Metallic materials — Part 2: Unalloyed titanium*

ISO 5832-3¹⁾, *Implants for surgery — Metallic materials — Part 3: Wrought titanium 6-aluminium 4-vanadium alloy*

ISO 6344-1, *Coated abrasives — Grain size analysis — Part 1: Grain size distribution test*

ISO 6892-1²⁾, *Metallic materials — Tensile testing — Part 1: Method of test at room temperature*

ISO 9513, *Metallic materials — Calibration of extensometer systems used in uniaxial testing*

ISO 9693-1, *Dentistry — Compatibility testing — Part 1: Metal-ceramic systems*

ISO 10271:2011, *Dentistry — Corrosion test methods for metallic materials*

ISO 15223-1:2012, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

3 Terms and definitions

For the purposes of this document, the terms and definitions of ISO 1942 and the following apply.

3.1

base metal

any metallic element with the exception of noble metals (i. e. gold and metals of the platinum group) and silver

1) To be published.

2) To be published.