

SÜDAME-VERESOOKONKONNA IMPLANTAADID.
SÜDAMEKLAPIPROTEESID. OSA 1: ÜLDNÕUDED

Cardiovascular implants - Cardiac valve prostheses -
Part 1: General requirements (ISO 5840-1:2015)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 5840-1:2015 sisaldab Euroopa standardi EN ISO 5840-1:2015 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 5840-1:2015 consists of the English text of the European standard EN ISO 5840-1:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 30.09.2015.	Date of Availability of the European standard is 30.09.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.40

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Cardiovascular implants - Cardiac valve prostheses - Part 1: General requirements (ISO 5840-1:2015)

Implants cardiovasculaires - Prothèses valvulaires -
Partie 1: Exigences générales (ISO 5840-1:2015)

Herz- und Gefäßimplantate - Herzklappenprothesen -
Teil 1: Allgemeine Anforderungen (ISO 5840-1:2015)

This European Standard was approved by CEN on 10 July 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 5840-1:2015) has been prepared by Technical Committee ISO/TC 150 “Implants for surgery” in collaboration with Technical Committee CEN/TC 285 “Non-active surgical implants” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2016, and conflicting national standards shall be withdrawn at the latest by March 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 5840-1:2015 has been approved by CEN as EN ISO 5840-1:2015 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC on medical devices

Clause(s)/sub-clause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
5, 6, 7	7.1	
5, 6.2.4, 6.4, 6.5	7.2	
6.2.2, 6.2.3, 7.3	7.3	
6.5	7.5	
6.2.2, 6.5	7.6	
5, 6.4, 6.5	8.1	
6.2.4, 6.4	8.3	
6.2.4	8.4	
6.4	8.5	
6.2.4	8.6	
6.2.4	8.7	
6.2.1, 6.3, 7	9.1	
6.2.1, 6.3, 6.4, 6.5, 7	9.2, 1. indent	
6.4, 6.5, 7	9.2, 2. indent	
6.4, 6.5, 7	9.2, 3. indent	
6.2.1, 6.3, 6.4, 6.5, 7	9.2, 4. indent	
6.2.4	13	

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Abbreviations.....	11
5 Fundamental requirements.....	12
6 Device description.....	12
6.1 Intended use.....	12
6.2 Design inputs.....	12
6.2.1 Operational specifications.....	12
6.2.2 Performance specifications.....	12
6.2.3 Implant procedure.....	12
6.2.4 Packaging, labelling, and sterilization.....	13
6.3 Design outputs.....	13
6.4 Design transfer (manufacturing verification/validation).....	14
6.5 Risk management.....	14
7 Design verification testing and analysis/design validation.....	15
7.1 General requirements.....	15
7.2 <i>In vitro</i> assessment.....	15
7.3 Preclinical <i>in vivo</i> evaluation.....	15
7.4 Clinical investigations.....	15
Annex A (informative) Rationale for the provisions of this part of ISO 5480.....	16
Annex B (normative) Packaging.....	19
Annex C (normative) Product labels, instructions for use, and training.....	20
Annex D (normative) Sterilization.....	23
Annex E (informative) <i>In vitro</i> test guidelines for paediatric devices.....	24
Annex F (informative) Statistical procedures when using <i>in vitro</i> performance criteria.....	28
Annex G (informative) Examples and definitions of some physical and material properties of heart valve systems.....	29
Annex H (informative) Examples of standards applicable to testing of materials and components of heart valve systems.....	40
Annex I (informative) Raw and post-conditioning mechanical properties for support structure materials.....	46
Annex J (informative) Corrosion assessment.....	48
Annex K (informative) Echocardiographic protocol.....	51
Bibliography.....	54

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary Information](#)

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This first edition of ISO 5840-1, together with ISO 5840-2 and ISO 5840-3, cancels and replaces ISO 5840:2005, which has been technically revised.

ISO 5840 consists of the following parts, under the general title *Cardiovascular implants — Cardiac valve prostheses*:

- *Part 1: General requirements*
- *Part 2: Surgically implanted heart valve substitutes*
- *Part 3: Heart valve substitutes implanted by transcatheter techniques*

Introduction

There is, as yet, no heart valve substitute which can be regarded as ideal.

The ISO 5840-series has been prepared by a group well aware of the issues associated with heart valve substitutes and their development. In several areas, the provisions of the ISO 5840-series deliberately have not been specified to encourage development and innovation. It does specify the types of tests, test methods, and/or requirements for test apparatus and requires documentation of test methods and results. The areas with which the ISO 5840-series are concerned are those which will ensure that associated risks to the patient and other users of the device have been adequately mitigated, facilitate quality assurance, aid the clinician in choosing a heart valve substitute, and ensure that the device will be presented at the operating table in convenient form. Emphasis has been placed on specifying types of *in vitro* testing, on preclinical *in vivo* and clinical evaluations, on reporting of all *in vitro*, preclinical *in vivo*, and clinical evaluations, and on the labelling and packaging of the device. Such a process involving *in vitro*, preclinical *in vivo*, and clinical evaluations is intended to clarify the required procedures prior to market release and to enable prompt identification and management of any subsequent problems.

With regard to *in vitro* testing and reporting, apart from basic material testing for mechanical, physical, chemical, and biocompatibility characteristics, the ISO 5840-series also covers important hydrodynamic and durability characteristics of heart valve substitutes. The ISO 5840-series does not specify exact test methods for hydrodynamic and durability testing, but it offers guidelines for the test apparatus.

The ISO 5840-series is incomplete in several areas. It is intended to be revised, updated, and/or amended as knowledge and techniques in heart valve substitute technology improve.