

Dentistry - Powered polymerization activators (ISO 10650:2015)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 10650:2015 sisaldab Euroopa standardi EN ISO 10650:2015 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 10650:2015 consists of the English text of the European standard EN ISO 10650:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 30.09.2015.	Date of Availability of the European standard is 30.09.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.060.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Powered polymerization activators (ISO 10650:2015)

Médecine bucco-dentaire - Activeurs électriques de polymérisation (ISO 10650:2015)

Zahnheilkunde - Polymerisationslampen (ISO 10650:2015)

This European Standard was approved by CEN on 22 August 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 10650:2015) has been prepared by Technical Committee ISO/TC 106 “Dentistry” in collaboration with Technical Committee CEN/TC 55 “Dentistry” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2016, and conflicting national standards shall be withdrawn at the latest by March 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10650-1:2005, EN ISO 10650-2:2007.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10650:2015 has been approved by CEN as EN ISO 10650:2015 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	2
5 Requirements	2
5.1 General	2
5.1.1 Design	2
5.1.2 Connection	2
5.1.3 Operating controls	2
5.1.4 Cleaning, disinfection, and sterilization	2
5.1.5 Excessive temperatures	2
5.2 Radiant exitance	3
5.2.1 Radiant exitance in the 385 nm to 515 nm (blue) wavelength region	3
5.2.2 Radiant exitance in the 200 nm to 385 nm wavelength region	3
5.2.3 Radiant exitance in the wavelength region above 515 nm	3
5.3 Electrical safety requirements	3
6 Sampling	3
7 Test methods	3
7.1 General	3
7.1.1 General provisions for tests	3
7.1.2 Atmospheric conditions	4
7.2 Radiant exitance	4
7.2.1 Apparatus	4
7.2.2 Procedures	6
7.2.3 Treatment of results	9
8 Information to be supplied by the manufacturer	11
8.1 Instructions for use	11
8.2 Technical description	12
9 Marking	12
10 Packaging	12
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

This first edition of ISO 10650 cancels and replaces ISO 10650-1:2004 and ISO 10650-2:2007, which have been technically revised with the following changes:

- limitation of blue wavelength region to: 200 nm to 385 nm;
- test procedure [7.2](#) radiant exitance was adopted to LED-diode lamps;
- information to be supplied by the manufacturer and marking requirements were updated.

Introduction

This International Standard specifies requirements and test methods in the 200 nm to 385 nm wavelength region and the wavelength region above 515 nm for powered polymerization activators. No minimum requirement value is given for the 385 nm to 515 nm wavelength region. The value in the 385 nm to 515 nm wavelength region is no less than the manufacturer's stated value.

This International Standard uses wavelength regions based on cut-off filters. Thus, the 200 nm to 385 nm region includes not only the ultraviolet region but also the near blue wavelength region of around 380 nm. The 385 nm to 515 nm region is taken as the region for powered polymerization activation. The region above 515 nm reaches approximately 1100 nm, which is the detection limit of the detector specified in this International Standard. The test methods described do not give absolute values nor do they reflect energy emitted as black body radiation. The measured values are not true radiant exitance but are values obtained using the methods described in this International Standard. Nevertheless, the values obtained using these test methods are used in conjunction with this International Standard.

This International Standard refers to IEC 60601-1, the basic International Standard on safety of medical electrical equipment, wherever relevant, by stating the respective clause numbers of IEC 60601-1.

Dentistry — Powered polymerization activators

1 Scope

This International Standard specifies requirements and test methods for powered polymerization activators in the 385 nm – 515 nm wavelength region intended for chairside use in polymerization of dental polymer-based materials. This International Standard applies to quartz-tungsten-halogen lamps and light-emitting diode (LED) lamps. Powered polymerization activators could have internal power supply (rechargeable battery powered) or be connected to external (mains) power supply. Lasers or plasma arc devices are not covered by this International Standard.

This International Standard does not cover powered polymerization activators used in laboratory fabrication of indirect restorations, veneers, dentures, or other oral dental appliances. This International Standard takes priority over IEC 60601-1 where specified in the individual clauses of this International Standard.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 9687, *Dentistry — Graphical symbols for dental equipment*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

IEC 60601-1, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance + Amendment 1:2012*

IEC 60601-1-2, *Medical electrical equipment — Part 1: General requirements for safety — 2. Collateral Standard: Electromagnetic compatibility — Requirements and test*

IEC 80601-2-60, *Medical electrical equipment — Part 2-60: Particular requirements for basic safety and essential performance of dental equipment*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and IEC 60601-1 apply.

NOTE The issue corresponds to IEC 60601-1:2005+A1:2012, Clause 3.

3.1

powered polymerization activator

device producing light primarily in the 385 nm to 515 nm region, intended for chairside use in polymerizing polymer-based filling, restorative, and luting materials

3.2

light-emitting diode lamps

semiconductor-based light emitting lamps