

Dentistry - Refractory investment and die material (ISO 15912:2016)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 15912:2016 sisaldab Euroopa standardi EN ISO 15912:2016 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 15912:2016 consists of the English text of the European standard EN ISO 15912:2016.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 10.02.2016.	Date of Availability of the European standard is 10.02.2016.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.060.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Refractory investment and die material (ISO
15912:2016)

Médecine bucco-dentaire - Revêtements et matériaux
pour modèles réfractaires (ISO 15912:2016)

Zahnheilkunde - Hochtemperaturbeständige
Einbettmassen und Stumpfmaterialien (ISO
15912:2016)

This European Standard was approved by CEN on 24 October 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 15912:2016) has been prepared by Technical Committee ISO/TC 106 “Dentistry” in collaboration with Technical Committee CEN/TC 55 “Dentistry” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2016, and conflicting national standards shall be withdrawn at the latest by August 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15912:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 15912:2016 has been approved by CEN as EN ISO 15912:2016 without any modification.

Contents

Page

Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	3
5 Requirements	3
5.1 General	3
5.2 Material consistency and freedom from contamination	3
5.3 Fluidity	3
5.4 Initial setting time	4
5.5 Compressive strength	4
5.6 Linear thermal dimensional change	4
5.7 Adequacy of expansion of Type 1 and Type 2 materials	4
6 Sampling, test conditions and mixing	4
6.1 Sampling	4
6.2 Test conditions	4
6.3 Mixing	5
6.3.1 Apparatus	5
6.3.2 Procedure	5
7 Test methods	5
7.1 Material consistency and freedom from contamination	5
7.1.1 Test procedure	5
7.1.2 Test report	5
7.2 Fluidity	5
7.2.1 Apparatus	5
7.2.2 Number of test-pieces	6
7.2.3 Test procedure	6
7.2.4 Evaluation of results	6
7.2.5 Test report	6
7.3 Initial setting time	7
7.3.1 Apparatus	7
7.3.2 Procedure	9
7.3.3 Number of determinations	9
7.3.4 Evaluation of results	10
7.3.5 Test report	10
7.4 Compressive strength	10
7.4.1 Apparatus	10
7.4.2 Number of test-pieces	10
7.4.3 Preparation of test-pieces	11
7.4.4 Test procedure	11
7.4.5 Evaluation of results	12
7.4.6 Test report	12
7.5 Linear thermal dimensional change	12
7.5.1 Apparatus	12
7.5.2 Number of test-pieces	13
7.5.3 Preparation of the test-piece	13
7.5.4 Test procedure for the measurement of linear thermal expansion of Types 1, 2 and 3 products	14
7.5.5 Test procedure for the measurement of linear thermal dimensional changes of a Type 4 product	15
7.5.6 Evaluation of results	15
7.5.7 Test report	16

7.6	Adequacy of expansion of Type 1 and Type 2 products	16
7.6.1	General	16
7.6.2	Type 1 and Type 2 products intended for casting of dental metallic materials	17
7.6.3	Type 1 products intended for dental pressable-ceramic products	18
7.6.4	Evaluation of results	20
7.6.5	Test report	20
8	Manufacturer's instructions	20
8.1	General	20
8.2	Information for use	20
8.3	Physical properties	21
8.4	Safety labelling and instructions for a product containing silica	22
9	Marking	22
9.1	General	22
9.2	Powder container	22
9.2.1	Outer container	22
9.2.2	Individual packets	23
9.3	Liquid container	23
10	Packaging	23
10.1	Powder	23
10.2	Liquid	24
	Bibliography	25

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106 *Dentistry*, Subcommittee SC2, *Prosthodontic Materials*.

This second edition cancels and replaces the first edition (ISO 15912:2006), which has been technically revised. It also incorporates the Amendment ISO 15912-1:2006/Amd 1:2011.

In this edition, dental pressable-ceramic investment materials are included in the Scope for the first time. These products are intended for the production of ceramic crowns and inlays and, as such, the same requirements as those for an investment product intended for the production of metallic crowns and inlays by casting are relevant (Type 1, according to the classification in this standard).

The previous edition contained requirements and test methods that had been developed for discontinued composition specific standards. In recent years products have been introduced that have other chemistries (for the binder and the refractory phase), specifically to minimize chemical reaction between the mould and the molten casting metallic material. A number of technical changes have been made to enable all dental casting investment products, regardless of their composition, to seek compliance with this International Standard and maintains the agreed philosophy that this International Standard should be inclusive, application-driven and not be limited by composition considerations.

Where appropriate, aspects of the test procedures have been changed to follow the manufacturer's instructions for use. The requirement for thermal dimensional change now takes into account the cooling of some products (after burn-out) to a lower casting temperature. The specification for the dilatometer has been changed for it to be compatible with the heating — and where relevant, the cooling after burn-out — of the product to the casting temperature.

The procedure for determining the initial setting time has been revised to harmonize with that present in the latest edition of the standard for dental gypsum products, ISO 6873:2013.^[1] Although substantially editorial, there are technical changes.

Information for use now requires a statement of the type of refractory phase(s) that is (are) present.

Labelling requirements for products that contain silica have been revised to comply with the current United Nations Globally Harmonized System for Classification and Labelling of Chemicals (UN GHS)^[2] and recommendations for silica as a hazardous material.

Containers of liquid must be marked to indicate the use to which the liquid is put.

This document is a preview generated by EVS

Dentistry — Refractory investment and die material

1 Scope

This International Standard gives requirements and test methods for determining the compliance of dental casting investment, dental brazing investment, dental pressable-ceramic investment and dental refractory die materials used in the dental laboratory, regardless of the composition of the refractory powder, the composition of the binder, or the particular application.

This International Standard classifies such products into types and classes, according to their intended use and the burn-out procedure recommended by the manufacturer.

It also gives requirements for marking, labelling and manufacturer's instructions.

It specifies requirements for the essential physical and mechanical properties of the products and the test methods to be used for determining them.

NOTE 1 Compliance with all of the requirements presented in [Clause 5](#) may not be necessary for some products, and a requirement might not be applicable to a product with a particular binder chemistry or be intended for an application in which that requirement is irrelevant. When this is the case, a clear statement to this effect is given according to [Clause 5](#).

NOTE 2 A specific quantitative requirement for setting expansion is not included in this International Standard. If the setting expansion of gypsum-bonded investment is measured, then the procedure given in ISO 6873[1] can be considered — a procedure not recommended, however, for investment materials with other binders.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 286-2, *Geometrical product specifications (GPS) — ISO code system for tolerances on linear sizes — Part 2: Tables of standard tolerance classes and limit deviations for holes and shafts*

ISO 1942, *Dentistry — Vocabulary*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 6344-1, *Coated abrasives — Grain size analysis — Part 1: Grain size distribution test*

ISO 6872, *Dentistry — Ceramic materials*

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

ISO 15854, *Dentistry — Casting and baseplate waxes*

ISO 22674, *Dentistry — Metallic materials for fixed and removable restorations and appliances*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.