

Dentistry - Oral care products - Manual interdental  
brushes (ISO 16409:2016)

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 16409:2016 sisaldab Euroopa standardi EN ISO 16409:2016 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 16409:2016 consists of the English text of the European standard EN ISO 16409:2016.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 30.11.2016.	Date of Availability of the European standard is 30.11.2016.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 97.170

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

English Version

Dentistry - Oral care products - Manual interdental  
brushes (ISO 16409:2016)

Médecine bucco-dentaire - Produits de soins bucco-  
dentaires - Broses interdentaires manuelles (ISO  
16409:2016)

Zahnheilkunde - Mundhygieneprodukte - Manuelle  
Interdentalbürsten (ISO 16409:2016)

This European Standard was approved by CEN on 14 August 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

## European foreword

This document (EN ISO 16409:2016) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2017, and conflicting national standards shall be withdrawn at the latest by May 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 16409:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Endorsement notice

The text of ISO 16409:2016 has been approved by CEN as EN ISO 16409:2016 without any modification.

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Classification</b>	<b>2</b>
<b>5 Requirements</b>	<b>2</b>
5.1 Defects or contamination	2
5.2 Brush size	2
5.3 Filament retention	3
5.4 Stem retention	3
5.5 Stem durability	3
<b>6 Sampling</b>	<b>3</b>
<b>7 Test methods</b>	<b>3</b>
7.1 General	3
7.2 Pass/fail criteria	3
7.3 Visual inspection	3
7.4 Passage hole diameter	3
7.4.1 Apparatus	3
7.4.2 Procedure	4
7.4.3 Evaluation	5
7.5 Filament retention	6
7.5.1 Apparatus	6
7.5.2 Procedure	6
7.5.3 Evaluation	6
7.6 Stem retention	6
7.6.1 Apparatus	6
7.6.2 Procedure	7
7.6.3 Evaluation	8
7.7 Stem durability	8
7.7.1 Apparatus	8
7.7.2 Procedure	10
7.7.3 Evaluation	11
<b>8 Test report</b>	<b>11</b>
<b>9 Accompanying information</b>	<b>11</b>
9.1 Manufacturer's instructions for use	11
9.2 Labelling on the package	11
<b>10 Packaging</b>	<b>12</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral care products*.

This second edition cancels and replaces the first edition (ISO 16409:2006), which has been technically revised with the following changes:

- addition of bigger brush size 8 in [Table 1](#);
- specification of number of test samples (eight manual interdental brushes) in [Clause 6](#);
- clarification of measuring procedure for the passage hole diameter for cylindrical and tapered brush heads.

It also incorporates the Amendment ISO 16409:2006/Amd 1:2010.

## Introduction

Manual interdental brushes are used for oral hygiene purposes intended to provide health benefits. Their main application purpose is mechanical plaque removal, primarily from the interdental regions.

# Dentistry — Oral care products — Manual interdental brushes

## 1 Scope

This International Standard specifies requirements and test methods for performance criteria for manual interdental brushes with a round cross-section of the brush head and consisting of a wired stem with inserted filaments. It also specifies the accompanying information such as manufacturer's instructions for use and labelling of the packaging.

Excluded are interdental brushes with a plastic core.

This International Standard is not applicable to powered interdental brushes, manual toothbrushes, dental floss, tapes, and strings and to interdental cleaners that do not include filaments.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

### 3.1

#### **manual interdental brush**

hand-powered device composed of *filaments* (3.5) emanating radially from a *stem* (3.4) intended for cleaning of interdental surfaces

### 3.2

#### **interdental brush head**

part of *manual interdental brush* (3.1), which passes in and out of spaces between the teeth to clean the exposed surfaces

Note 1 to entry: The brush head may be fixed or removable, but is fixed during use.

### 3.3

#### **interdental brush handle**

part of *manual interdental brush* (3.1), which holds the *stem* (3.4) of the manual interdental brush

### 3.4

#### **stem**

central support structure of the manual *interdental brush head* (3.2), usually composed of twisted wire, which secures the *filaments* (3.5)

Note 1 to entry: The stem is either secured

- in the handle,
- in a connector, or
- fulfils the function of a handle itself.