

**Acoustics - Requirements for the
performance and calibration of reference
sound sources used for the determination
of sound power levels**

Acoustics - Requirements for the performance and
calibration of reference sound sources used for the
determination of sound power levels

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 6926:2002 sisaldab Euroopa standardi EN ISO 6926:2001 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 14.02.2002 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 6926:2002 consists of the English text of the European standard EN ISO 6926:2001.</p> <p>This document is endorsed on 14.02.2002 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

<p>Käsitlusala:</p> <p>The standard specifies the acoustical performance requirements for reference sound sources:</p> <ul style="list-style-type: none">- temporal steadiness and repeatability of the sound power output,- spectral characteristics,- directivity index.	<p>Scope:</p> <p>The standard specifies the acoustical performance requirements for reference sound sources:</p> <ul style="list-style-type: none">- temporal steadiness and repeatability of the sound power output,- spectral characteristics,- directivity index.
-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

ICS 17.140.01

Võtmesõnad: acoustic tests, acoustics, calibration, determination, noise (sound), performance, reference sources, sound power, sound sources, specifications, tests

English version

Acoustics

**Requirements for the performance and calibration of reference
sound sources used for the determination of sound power levels
(ISO 6926 : 2000)**

Acoustique – Exigences relatives aux performances et à l'étalonnage des sources sonores de référence utilisées dans la détermination des niveaux de puissance sonore (ISO 6926 : 2000)

Akustik – Anforderungen an die Eigenschaften und die Kalibrierung von Vergleichsschallquellen für die Bestimmung von Schalleistungspegeln (ISO 6926 : 2000)

This European Standard was approved by CEN on 2001-06-09.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Management Centre: rue de Stassart 36, B-1050 Brussels

Foreword

International Standard

ISO 6926 : 2000 Acoustics – Requirements for the performance and calibration of reference sound sources used for the determination of sound power levels,

which was prepared by ISO/TC 43 'Acoustics' of the International Organization for Standardization, has been adopted by Technical Committee CEN/TC 211 'Acoustics', the Secretariat of which is held by DS, as a European Standard.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, and conflicting national standards withdrawn, by January 2002 at the latest.

In accordance with the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard:

Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 6926 : 2000 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to international publications are listed in Annex ZA (normative).

Contents

Page

Foreword.....	2
Introduction	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	5
4 Measurement uncertainty	7
5 Performance requirements	8
5.1 General	8
5.2 Temporal steadiness and repeatability of sound power output	8
5.3 Total broad-band sound power level	8
5.4 Spectral characteristics	8
5.5 Directivity index	9
5.6 Recalibration	9
6 Installation and operation of the reference sound source during calibration	9
6.1 General	9
6.2 Position of the reference source	10
7 Calibration procedure in hemi-anechoic rooms	10
7.1 Test environment	10
7.2 Microphone	10
7.3 Microphone positions	10
7.4 Measurements	11
7.5 Air absorption	11
7.6 Calculations	11
8 Calibration procedure in reverberation rooms	12
8.1 Test environment	12
8.2 Microphone	12
8.3 Microphone positions	12
8.4 Measurements	12
8.5 Calculations	12
9 Information to be recorded	13
10 Information to be reported	13
Bibliography	14

Introduction

Reference sound sources are used extensively in "comparison methods" for determining the noise emissions of physically stationary sound sources. A reference sound source, of known sound power output, is used to establish the numerical relationship between the sound power level of a source, in a given location in a given acoustical environment and the space- and time-averaged sound pressure level at a set of microphone positions. Once that relationship is established, it is straightforward to measure the average sound pressure level produced by an "unknown source" and to determine the sound power level produced by that source.

This International Standard defines the important physical and performance characteristics of reference sound sources and specifies procedures for their calibration, primarily to determine the sound power level of other sound sources.

This International Standard supplements a series of International Standards, the ISO 3740 series, that describes various methods for determining the sound power levels of machines and equipment. This series specifies the acoustical requirements for measurements that are appropriate for different test environments.

Five International Standards in the ISO 3740 series include procedures in which a reference sound source is used: ISO 3741, ISO 3743, ISO 3744, ISO 3746 and ISO 3747. ISO 3740 gives guidelines for the use of all the International Standards in the series.

It should be noted that the sound power output of reference sound sources will vary, in particular at low frequencies, with the distance from the source to nearby reflecting planes. Sound power data of reference sound sources are thus valid only for the position used during the calibration.

In addition to being useful for determining sound power levels by the comparison method, reference sound sources may be used for qualification tests on an acoustic environment and to estimate the influence of an acoustic environment on the sound pressure levels produced by one or more sound sources located in that environment. Examples of International Standards referring to reference sound sources with these applications are ISO 11690-3 and ISO 14257. Requirements other than those of this International Standard may be applicable in these cases.

1 Scope

This International Standard specifies the acoustical performance requirements for reference sound sources:

- temporal steadiness and repeatability of the sound power output,
- spectral characteristics,
- directivity index.

The stability of the sound power output and the directivity index, for those sources where directivity is needed, are normally only determined in connection with pattern evaluation of the reference sound source. Because of the directivity measurements (for an exception see 5.5), pattern evaluations can only be performed in a hemi-anechoic environment. For regular verification measurements, only the frequency band sound power levels are normally determined. In this case measurements may be made in either hemi-anechoic or reverberant conditions.

This International Standard also specifies procedures for calibrating a sound source intended for use as a reference sound source in terms of its sound power level under the reference condition that the characteristic impedance of air (ρc) is equal to 400 N s/m^3 in octave and in one-third-octave bands, and with frequency weighting A. Different procedures are specified for pattern evaluation and verification.

NOTE Reference sound sources may also be used for measurements in one-half-octave bands, e.g. for ISO 9295. However, under these circumstances the stability and reproducibility limits stated in this International Standard will not apply.

This International Standard specifies methods to calibrate reference sound sources not only in a free field over a reflecting plane but also in reverberation rooms at different distances from the boundary surfaces. For the position of the reference sound source on one reflecting plane, the two different test environments mentioned above are considered equivalent for frequency bands above or equal to 100 Hz. Below 100 Hz the measurement uncertainties are significantly different (see Table 1).

This International Standard is applicable to a sound source which is intended for use as a reference sound source. The sound source may either be placed directly on the floor or mounted on a stand to be used at a certain elevation above the floor. For floor-mounted sources, this International Standard is valid only for sources whose maximum vertical dimension is less than 0,5 m and whose maximum horizontal dimension is less than 0,8 m. According to this International Standard only floor-mounted reference sound sources may be used when carrying out measurements on a measurement surface. For reference sound sources to be used or calibrated under reverberant conditions, no such restrictions on maximum dimensions apply.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For

undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3741:1999, *Acoustics — Determination of sound power levels of noise sources using sound pressure — Precision methods for reverberation rooms.*

ISO 3744, *Acoustics — Determination of sound power levels of noise sources using sound pressure — Engineering method in an essentially free field over a reflecting plane.*

ISO 3745:1977, *Acoustics — Determination of sound power levels of noise sources — Precision methods for anechoic and semi-anechoic rooms.*

ISO 5725-1, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions.*

ISO 9613-1, *Acoustics — Attenuation of sound during propagation outdoors — Part 1: Calculation of the absorption of sound by the atmosphere.*

IEC 61183, *Electroacoustics — Random-incidence and diffuse-field calibration of sound level meters.*

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1

free field over a reflecting plane

sound field in a homogeneous, isotropic medium in the half-space above an infinite, rigid plane surface over which the source is located

3.2

hemi-anechoic room

test room with a reflecting plane (hard floor) meeting the requirements of ISO 3745

3.3

surface sound pressure level

L_{pf}

energy-average (see ISO 3744) of the time-averaged sound pressure levels at all the microphone positions on the measurement surface

NOTE It is expressed in decibels.

3.4

sound power level

L_W

ten times the logarithm to the base 10 of the ratio of the sound power radiated by the sound source under test to the reference sound power (10^{-12} W)

NOTE It is expressed in decibels.

3.5

measurement surface

hypothetical surface enveloping the source on which the sound pressure levels are measured

NOTE For the purposes of this International Standard, the measurement surface is either a hemisphere terminating on the reflecting plane or a sphere.