INTERNATIONAL STANDARD

ISO 7981-2

First edition 2005-06-15

Water quality — Determination of polycyclic aromatic hydrocarbons (PAH) —

Part 2:

Determination of six PAH by highperformance liquid chromatography with fluorescence detection after liquid-liquid extraction

Qualité de l'eau — Détermination des hydrocarbures aromatiques polycycliques (HAP) —

Partie 2: Dosage de six HAP par chromatographie de haute performance en phase liquide avec détection fluorimétrique à la suite d'une extraction liquide-liquide



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below

This document is a preview denetated by this

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org

Published in Switzerland

Contents Page 1 Scope 2 Normative references Principle...U. 3 4 Interferences 5 Reagents Apparatus. 6 7 Sampling... 8 Procedure ... 9 Calibration **Measurement of samples** 10 Determination of the recovery... 11 12 Blank measurement 13 Calculation..... 14 **Expression of results** 15 16 Annex B (informative) Examples for the construction of special apparatus Annex C (informative) Accuracy..... Bibliography

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in Maison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7981-2 was prepared by Technical Committee ISO/TC 147, Water quality, Subcommittee SC 2, Physical, chemical and biochemical methods.

ISO 7981 consists of the following parts, under the general title *Water quality* — *Determination of polycyclic aromatic hydrocarbons (PAH)*:

- Part 1: Determination of six PAH by high-performance thin-layer chromatography with fluorescence detection after liquid-liquid extraction
- Part 2: Determination of six PAH by high-performance liquid chromatography with fluorescence detection after liquid-liquid extraction

Introduction

Polycyclic aromatic hydrocarbons (PAH) are present in nearly all types of waters. These substances are adsorbed on solids (sediments, suspended matter) as well as dissolved in the liquid phase.

Some PAH are known or suspected to cause cancer. The maximum acceptable levels of PAH in waters intended for human consumption are given in European Legislation [1] [2] [3] [4].

The sum of the mess concentrations of the six PAH specified in this part of ISO 7981 usually is about 0,01 µg/l to 0,05 µg/l in ground water, up to 1 µg/l in surface water, and up to 1 000 µg/l in waste water.

© ISO 2005 - All rights reserved

Inis document is a preview denetated by EUS

Water quality — Determination of polycyclic aromatic hydrocarbons (PAH) —

Part 2:

Determination of six PAH by high-performance liquid chromatography with fluorescence detection after liquid-liquid extraction

WARNING — Some compounds being measured are presumed to be carcinogenic. Acetonitrile and hexane are harmful.

Persons using this part of ISO 7981 should be familiar with normal laboratory practise. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user of this part of ISO 7981 to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this part of ISO 7981 be carried out by suitably trained staff.

1 Scope

This part of ISO 7981 specifies the determination of six selected PAH in drinking, mineral and table waters and ground and surface waters in mass concentrations above $0,005 \,\mu\text{g/l}$, by high-performance liquid chromatography with fluorescence detection after liquid-liquid extraction. The six PAH are: fluoranthene, benzo[b]fluoranthene, benzo[a]pyrene, benzo[a]pyrene, benzo[a]pyrene, benzo[a]pyrene, and benzo[a]pyrene (see Table 1).

With some modification, this method is also applicable for the analysis of moderately polluted waste waters.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8466-1, Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function

3 Principle

Since PAH can to a large extent be adsorbed on particulate matter, the whole sample is analysed.

NOTE For the analysis of surface water, a differentiation between dissolved and undissolved PAH can be desirable, but this is not relevant for drinking water.

© ISO 2005 – All rights reserved