
**Water quality — Determination of
orthophosphate and total phosphorus
contents by flow analysis (FIA and
CFA) —**

**Part 2:
Method by continuous flow analysis
(CFA)**

*Qualité de l'eau — Dosage des orthophosphates et du phosphore total
par analyse en flux (FIA et CFA) —*

Partie 2: Méthode par analyse en flux continu (CFA)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Interferences	2
3.1 General interferences	2
3.2 Interferences in the determination of total-P	2
4 Principle	2
4.1 Determination of orthophosphate	2
4.2 Total phosphorus with manual digestion	3
4.3 Total phosphorus with integral UV digestion and hydrolysis	3
5 Reagents	3
6 Apparatus	7
6.1 Continuous-flow analysis (CFA)	7
6.2 Additional apparatus	7
6.3 Additional apparatus for the determination of total phosphorus after integral digestion	8
7 Sampling and sample preparation	8
8 Procedure	8
8.1 Preparation for analysis	8
8.2 Instrument performance check	8
8.3 Reagent blank check	9
8.4 Calibration	9
8.5 Check of UV digestion and hydrolysis for total-P determination (see Figure A.2)	9
8.6 Measurement	10
8.7 Closing down the system	10
9 Calculation of results	10
10 Expression of results	10
11 Test report	10
Annex A (informative) Examples of a CFA system	12
Annex B (informative) Precision and accuracy	14
Annex C (informative) Determination of orthophosphate-P and total-P by CFA and tin(II) chloride reduction	15
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15681-2 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

ISO 15681 consists of the following parts, under the general title *Water quality — Determination of orthophosphate and total phosphorus contents by flow analysis (FIA and CFA)*:

- *Part 1: Method by flow injection analysis (FIA)*
- *Part 2: Method by continuous flow analysis (CFA)*

Introduction

Methods of determining water quality using flow analysis automated wet chemical procedures and are particularly suitable for the processing of many analytes in water in large sample series at a high analysis frequency.

Analysis can be performed by flow injection analysis (FIA) ^{[1], [2]} or continuous flow analysis (CFA) ^[3]. Both methods share the feature of an automatic dosage of the sample into a flow system (manifold) where the analyte in the sample reacts with the reagent solutions on its way through the manifold. The sample preparation may be integrated in the manifold. The amount of reaction product is measured in a flow detector (e.g. flow photometer). This part of ISO 15681 describes the CFA method.

The user should be aware that particular problems could require the specification of additional marginal conditions.

This document is a preview generated by EVS

Water quality — Determination of orthophosphate and total phosphorus contents by flow analysis (FIA and CFA) —

Part 2: Method by continuous flow analysis (CFA)

WARNING — Persons using this part of ISO 15681 should be familiar with normal laboratory practice. This part of ISO 15681 does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions. Molybdate and antimony waste solutions should be disposed of properly. It is absolutely essential that tests conducted according to this part of ISO 15681 be carried out by suitably qualified staff.

1 Scope

This part of ISO 15681 specifies CFA methods for the determination of orthophosphate in the mass concentration range from 0,01 mg/l to 1,00 mg/l P, and total phosphorus in the mass concentration range from 0,10 mg/l to 10,0 mg/l P. The method includes the digestion of organic phosphorus compounds and the hydrolysis of inorganic polyphosphate compounds, performed either manually as described in ISO 6878 [5], [6] or with an integrated UV digestion and hydrolysis unit.

This part of ISO 15681 is applicable to various types of water (such as ground, drinking, surface, leachate and waste water). The range of application may be changed by varying the operating conditions.

This method is also applicable to the analysis of seawater, but with changes in sensitivity, by adaptation of the carrier and calibration solutions to the salinity of the samples.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specifications and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design on sampling programmes*

ISO 5667-2, *Water quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of water samples*

ISO 6878:—¹⁾, *Water quality — Determination of phosphorus — Ammonium molybdate spectrometric method*

1) To be published.

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

3 Interferences

3.1 General interferences

ISO 6878:—, Annex B gives a list of general interferences. In addition, or contrary to the cited standard, the following guidelines apply.

- a) Arsenate causes serious interference. 100 µg/l As, present as arsenate, results in a response comparable to approximately 30 µg/l P.
- b) If the silicate concentration in samples is not greater than 60 times the phosphorus concentration, interferences by silicate can be neglected.
- c) Fluoride interference is significant above 50 mg/l.
- d) Nitrite interference is significant above 5 mg/l. The interference can be eliminated by acidifying samples after collection.
- e) For samples containing high concentrations of oxidizing agents, the amount of added reduction reagent can be insufficient. In this case, remove the oxidizing material prior to digestion.
- f) The self-absorption of the sample can be compensated by measuring, in addition to the sample signal (8.6), the signal of the sample without the admixture of the reagents. In this case, the difference of the two responses is used for the evaluation (Clause 9).

3.2 Interferences in the determination of total-P

Samples containing solids or suspended particles can show low values when analysed by the UV method, if the particles are not completely transported into the UV unit. The error can be minimized by stirring the sample immediately before sampling, in order to ensure that a representative sample is delivered into the analyser, and by reducing the particle size.

The interferences from silicate, nitrite, fluoride and iron described for the orthophosphate determination are generally not observed in the UV method, due to the pre-digestion and the higher analytical range.

The efficiency of the UV digestion can be affected for water samples with chemical oxygen demand (COD) values of more than 10 times the highest concentrations of the calibration solutions (5.21). In this case, the sample should be diluted.

4 Principle

4.1 Determination of orthophosphate

The sample is mixed with a surfactant solution, followed by an acidic solution containing molybdate and antimony ions. The resulting phospho-antimony-molybdate complex is reduced by ascorbic acid to molybdenum blue [4], [5].