
Plastics — Determination of the transient extensional viscosity of polymer melts

*Plastiques — Détermination de la viscosité élongationnelle transitoire
des polymères à l'état fondu*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	1
4 General principles.....	3
5 Apparatus.....	3
6 Sampling and specimen preparation	6
7 Procedure.....	7
8 Analysis of extensional flow measurements	8
9 Precision	11
10 Test report.....	11
Annex A (informative) Checking for swelling of specimens due to immersion in silicone oil or other fluids.....	12
Annex B (informative) Uncertainties in transient extensional viscosity testing.....	13
Bibliography	18

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20965 was prepared by Technical Committee ISO/TC 61, *Plastics*, Subcommittee SC 5, *Physical-chemical properties*.

Plastics — Determination of the transient extensional viscosity of polymer melts

1 Scope

This International Standard specifies the general principles of a method for determining the transient extensional viscosity of polymer melts. The procedure details the measurement of polymer melt specimens stretched uniaxially under conditions of constant strain rate and constant temperature.

The method is capable of measuring the transient extensional viscosity of polymer melts at Hencky strain rates typically in the range $0,01 \text{ s}^{-1}$ to 1 s^{-1} , at Hencky strains up to approximately 4 and at temperatures up to approximately 250°C (see Notes 1 and 2). It is suitable for measuring transient extensional viscosity values typically in the range from approximately $10^4 \text{ Pa}\cdot\text{s}$ to $10^7 \text{ Pa}\cdot\text{s}$ (see Note 3).

NOTE 1 Hencky strains and strain rates are used (see Clause 3).

NOTE 2 Values of strain, strain rate and temperature outside these limiting values may be attained.

NOTE 3 The operating limit of an instrument in terms of the lowest transient extensional viscosity values that can be measured, is due to a combination of factors, including the ability of the specimen to maintain its shape during testing and the resolution of the instrument.

A list of documents related to this International Standard is given in the Bibliography.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 472, *Plastics — Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 472, together with the following, apply.

Definitions 3.1 to 3.5 are given by Whorlow^[1] for strains and strain rates, and by the Nomenclature Committee of the Society of Rheology for start-up flow in tensile uniaxial extension at constant Hencky strain rate^[2].

3.1

Hencky strain

ε

strain given by the natural logarithm of the elongation ratio

$$\varepsilon = \ln(l/l_0) \quad (1)$$

where l is the specimen length and l_0 is the original specimen length