

**VÄLISÕHK. ULTRAVIOLETTFLUORESTSENTSIL
PÕHINEV STANDARDMEETOD VÄÄVELDIOKSIIDI
KONTSETRATSIOONI MÕÕTMISEKS**

**Ambient air - Standard method for the measurement of
the concentration of sulphur dioxide by ultraviolet
fluorescence**

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 14212:2012 sisaldab Euroopa standardi EN 14212:2012 ingliskeelset teksti.	This Estonian standard EVS-EN 14212:2012 consists of the English text of the European standard EN 14212:2012.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 22.08.2012.	Date of Availability of the European standard is 22.08.2012.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.040.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Ambient air - Standard method for the measurement of the
concentration of sulphur dioxide by ultraviolet fluorescence**

Qualité de l'air ambiant - Méthode normalisée pour le
mesurage de la concentration en dioxyde de soufre par
fluorescence U.V.

Luftqualität - Messverfahren zur Bestimmung der
Konzentration von Schwefeldioxid mit Ultraviolett-
Fluoreszenz

This European Standard was approved by CEN on 10 May 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Foreword.....	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	6
4 Abbreviated terms	11
5 Principle.....	11
5.1 General.....	11
5.2 Measuring principle	11
5.3 Type approval test	12
5.4 Field operation and quality control.....	12
6 Sampling.....	13
6.1 General.....	13
6.2 Sampling location	13
6.3 Sampling system	13
6.4 Control and regulation of sample flow rate	14
6.5 Sampling pump for the manifold.....	14
7 Analyser equipment	15
7.1 General.....	15
7.2 Selective traps for interfering agents	15
7.3 Optical assembly	15
7.4 Pressure measurement	15
7.5 Flow rate indicator	15
7.6 Sampling pump for the analyser	16
7.7 Internal sulphur dioxide span source.....	16
7.8 Particle filter	16
8 Type approval of ultraviolet fluorescence sulphur dioxide analysers.....	16
8.1 General.....	16
8.2 Relevant performance characteristics and performance criteria	17
8.3 Design change	18
8.4 Procedures for determination of the performance characteristics during the laboratory test	19
8.5 Determination of the performance characteristics during the field test.....	29
8.6 Type approval and uncertainty calculation.....	33
9 Field operation and ongoing quality control	34
9.1 General.....	34
9.2 Suitability evaluation	34

9.3	Initial installation	36
9.4	Ongoing quality assurance/quality control.....	37
9.5	Calibration of the analyser.....	39
9.6	Checks	40
9.7	Maintenance	44
9.8	Data handling and data reports.....	45
9.9	Measurement uncertainty	45
10	Expression of results	46
11	Test reports and documentation.....	46
11.1	Type approval test	46
11.2	Field operation	47
Annex A	(normative) Test of lack of fit	49
Annex B	(informative) Sampling equipment.....	51
Annex C	(informative) Ultraviolet fluorescence analyser	53
Annex D	(informative) Manifold testing	54
Annex E	(normative) Type approval.....	56
Annex F	(informative) Calculation of uncertainty in field operation at the hourly limit value	75
Annex G	(informative) Calculation of uncertainty in field operation at the daily limit value.....	83
Annex H	(informative) Calculation of uncertainty in field operation at the annual critical level	93
Annex I	(informative) Significant technical changes.....	103
Bibliography	104

Foreword

This document (EN 14212:2012) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2013, and conflicting national standards shall be withdrawn at the latest by February 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14212:2005.

The technical changes made since EN 14212:2005 are listed in Annex I of this European Standard.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies a continuous measurement method for the determination of the concentration of sulphur dioxide present in ambient air based on the ultraviolet fluorescence measuring principle. This standard describes the performance characteristics and sets the relevant minimum criteria required to select an appropriate ultraviolet fluorescence analyser by means of type approval tests. It also includes the evaluation of the suitability of an analyser for use in a specific fixed site so as to meet the data quality requirements as specified in Annex I of Directive 2008/50/EC [1] and requirements during sampling, calibration and quality assurance for use.

The method is applicable to the determination of the mass concentration of sulphur dioxide present in ambient air up to 1000 $\mu\text{g}/\text{m}^3$. This concentration range represents the certification range for SO_2 for the type approval test.

NOTE 1 Other ranges may be used depending on the levels present in ambient air.

NOTE 2 When the standard is used for other purposes than for measurements required by Directive 2008/50/EC, the ranges and uncertainty requirements may not apply.

The method covers the determination of ambient air concentrations of sulphur dioxide in zones classified as rural areas, urban-background areas and traffic-orientated locations and locations influenced by industrial sources.

The results are expressed in $\mu\text{g}/\text{m}^3$ (at 20 °C and 101,3 kPa).

NOTE 3 1 000 $\mu\text{g}/\text{m}^3$ of SO_2 corresponds to 376 nmol/mol of SO_2 .

This standard contains information for different groups of users.

Clauses 5 to 7 and Annexes C and D contain general information about the principles of sulphur dioxide measurement by ultraviolet fluorescence analyser and sampling equipment.

Clause 8 and Annex E are specifically directed towards test houses and laboratories that perform type-approval testing of sulphur dioxide analysers. These sections contain information about:

- Type-approval test conditions, test procedures and test requirements;
- Analyser performance requirements;
- Evaluation of the type-approval test results;
- Evaluation of the uncertainty of the measurement results of the sulphur dioxide analyser based on the type-approval test results.

Clauses 9 to 11 and Annexes F and G are directed towards monitoring networks performing the practical measurements of sulphur dioxide in ambient air. These sections contain information about:

- Initial installation of the analyser in the monitoring network and acceptance testing;
- Ongoing quality assurance/quality control;
- Calculation and reporting of measurement results;
- Evaluation of the uncertainty of measurement results under practical monitoring conditions.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality — Certification of automated measuring systems — Part 1: General principles*

EN 15267-2, *Air quality — Certification of automated measuring systems — Part 2: Initial assessment of the AMS manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN ISO 6142, *Gas analysis — Preparation of calibration gas mixtures — Gravimetric method (ISO 6142)*

EN ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures (ISO 6143)*

EN ISO 6144, *Gas analysis — Preparation of calibration gas mixtures — Static volumetric methods (ISO 6144)*

EN ISO 6145-6, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 6: Critical orifices (ISO 6145-6)*

EN ISO 6145-7, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 7: Thermal mass-flow controllers (ISO 6145-7)*

EN ISO 6145-10, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 10: Permeation method (ISO 6145-10)*

EN ISO 14956, *Air quality — Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025)*

ENV 13005:1999, *Guide to the expression of uncertainty in measurement*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1 adjustment

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

Note 1 to entry: Types of adjustment of a measuring system include zero adjustment of a measuring system, offset adjustment, and span adjustment (sometimes called gain adjustment).

Note 2 to entry: Adjustment of a measuring system should not be confused with calibration, which is a prerequisite for adjustment.

[SOURCE: JCGM 200:2012 (VIM) [2]]

Note 3 to entry: In the context of this standard, adjustment is performed on measurement data rather than on the analyser.

3.2 ambient air

outdoor air in the troposphere, excluding workplaces as defined by Directive 89/654/EEC, where provisions concerning health and safety at work apply and to which members of the public do not have regular access

[SOURCE: 2008/50/EC [1]]