

ELEKTRILISED MEDITSIINISEADMED. OSA 2-16:
ERINÕUDED HEMODIALÜÜSI, HEMODIAFILTRATSIOONI
JA HEMOFILTRATSIOONISEADMETE ESMASELE
OHUTUSELE JA OLULISTELE TOIMIMISNÄITAJATELE

Medical electrical equipment - Part 2-16: Particular
requirements for basic safety and essential
performance of haemodialysis, haemodiafiltration and
haemofiltration equipment

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 60601-2-16:2015 sisaldab Euroopa standardi EN 60601-2-16:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 60601-2-16:2015 consists of the English text of the European standard EN 60601-2-16:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 12.06.2015.	Date of Availability of the European standard is 12.06.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040, 11.040.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Medical electrical equipment - Part 2-16: Particular requirements
for the basic safety and essential performance of haemodialysis,
haemodiafiltration and haemofiltration equipment
(IEC 60601-2-16:2012)**

Appareils électromédicaux - Partie 2-16: Exigences
particulières pour la sécurité de base et les performances
essentiels des appareils d'hémodialyse,
d'hémodiafiltration et d'hémofiltration
(IEC 60601-2-16:2012)

Medizinische elektrische Geräte - Teil 2-16: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Hämodialyse-,
Hämodiafiltrations- und Hämofiltrationsgeräten
(IEC 60601-2-16:2012)

This European Standard was approved by CENELEC on 2015-04-14. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

The text of document 62D/972/FDIS, future edition 4 of IEC 60601-2-16, prepared by SC 62D, "Electromedical equipment", of IEC TC 62, "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-16:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-01-14
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-04-14

This document supersedes EN 60601-2-16:1998.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive see informative Annex ZZ, which is an integral part of this document.

Endorsement notice

The text of the International Standard IEC 60601-2-16:2012 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-16:1998	NOTE Harmonized as EN 60601-2-16:1998 (not modified).
IEC 60601-2-39	NOTE Harmonized as EN 60601-2-39.
ISO 11197	NOTE Harmonized as EN ISO 11197.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006 applies with the following exceptions:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement :</i>				
IEC 60601-1-2	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr March	2007 2010
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6	2010
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corr March	2007 2010
<i>Addition:</i>				
IEC 60601-1-10	2007	Medical electrical equipment - Part 1-10: General requirements for basic safety and essential performance - Collateral Standard: Requirements for the development of physiologic closed-loop controllers	EN 60601-1-10	2008
IEC 60601-1-11	2010	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2010
IEC 62366	2007	Medical devices - Application of usability engineering to medical devices	EN 62366	2008

ISO 594-2	-	Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment - Part 2: Lock fittings		
ISO 3744	-	Acoustics - Determination of sound power levels and sound energy levels of noise sources using sound pressure - Engineering methods for an essentially free field over a reflecting plane	EN ISO 3744	-
ISO 8638	-	Cardiovascular implants and extracorporeal systems - Extracorporeal blood circuit for haemodialysers, haemodiafilters and haemofilters	EN ISO 8638	-

Annex ZZ (informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references	9
201.3 Terms and definitions	10
201.4 General requirements	12
201.5 General requirements for testing of ME EQUIPMENT	15
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	16
201.7 ME EQUIPMENT identification, marking and documents	16
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	19
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	20
201.10 Protection against unwanted and excessive radiation HAZARDS	21
201.11 Protection against excessive temperatures and other HAZARDS	21
201.12 * Accuracy of controls and instruments and protection against hazardous outputs.....	22
201.13 HAZARDOUS SITUATIONS and fault conditions	30
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	31
201.15 Construction of ME EQUIPMENT.....	31
201.16 * ME SYSTEMS	32
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	33
202 Electromagnetic compatibility – Requirements and tests.....	33
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	33
210 Process requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS.....	35
211 * Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT	35
Annexes	35
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures	36
Annex AA (informative) Particular guidance and rationale	37
Annex BB (informative) Examples of HAZARDS, foreseeable sequences of events, and HAZARDOUS SITUATIONS in HAEMODIALYSIS EQUIPMENT	55
Bibliography	63
Index of defined terms used in this particular standard	64
Figure 201.101 – Continuous air infusion test set-up with example dimensions	28
Figure AA.1 – Example of the HAEMODIALYSIS ME SYSTEM	51
Table 201.101 – ESSENTIAL PERFORMANCE requirements	12
Table AA.1: Possible ALARM CONDITION priorities according to 6.1.2 of IEC 60601-1-8:2006,.....	53
Table BB.1 – Hazardous situation list following ISO 14971:2007, Annex E	55

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT.

This document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Addition:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT, hereafter referred to as HAEMODIALYSIS EQUIPMENT.

This International Standard does not take into consideration the DIALYSIS FLUID control system of HAEMODIALYSIS EQUIPMENT using regeneration of DIALYSIS FLUID and CENTRAL DELIVERY SYSTEMS. It does however take into consideration the specific safety requirements of such HAEMODIALYSIS EQUIPMENT concerning electrical safety and PATIENT safety.

This International Standard specifies the minimum safety requirements for HAEMODIALYSIS EQUIPMENT. These devices are intended for use either by medical staff or for use by the PATIENT or other trained personnel under the supervision of medical expertise.

This International Standard includes all ME EQUIPMENT that is intended to deliver a HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION treatment to a PATIENT suffering from kidney failure.

The particular requirements in this International standard do not apply to:

- EXTRACORPOREAL CIRCUITS;
- DIALYSERS;
- DIALYSIS FLUID CONCENTRATES;
- water treatment equipment;
- equipment used to perform PERITONEAL DIALYSIS (see IEC 60601-2-39).

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of IEC 60601-1.

NOTE See also 4.2 of IEC 60601-1:2005.

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for HAEMODIALYSIS EQUIPMENT

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1, and Clause 201.2 of this International Standard.

IEC 60601-1-2, IEC 60601-1-8, IEC 60601-1-10 and IEC 60601-11 apply as modified in Clauses 202, 208, 210 and 211 respectively. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published..

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding, clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography beginning on page 63.

Clause 2 of the general standard applies, except as follows:

Amendment:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.*

Addition:

IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*

IEC 60601-1-11:2010, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 62366:2007, *Medical devices – Application of usability engineering to medical devices*

ISO 594-2, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment – Part 2: Lock fittings*

ISO 3744, *Acoustics – Determination of sound power levels of noise sources using sound pressure – Engineering method in an essentially free field over a reflecting plane*

ISO 8638, *Cardiovascular implants and artificial organs – Extracorporeal blood circuit for haemodialysers, haemodiafilters and haemofilters*