

**Lasertoodete ohutus. Osa 1: Seadmete klassifikatsioon ja nõuded**

**Safety of laser products - Part 1: Equipment classification and requirements**

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN 60825-1:2014 sisaldab konsolideeritult Euroopa standardi EN 60825-1:2014 ja selle paranduse EN 60825-1:2014/AC:2017-06 ingliskeelset teksti.	This Estonian standard EVS-EN 60825-1:2014 consists of the English text of the European standard EN 60825-1:2014 and its corrigendum EN 60825-1:2014/AC:2017-06.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 08.08.2014.	Date of Availability of the European standard is 08.08.2014.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 13.110, 31.260

### Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:  
Aru 10, 10317 Tallinn, Eesti; [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

### The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:  
Aru 10, 10317 Tallinn, Estonia; [www.evs.ee](http://www.evs.ee); phone 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

English Version

## Safety of laser products - Part 1: Equipment classification and requirements (IEC 60825-1:2014)

Sécurité des appareils à laser - Partie 1: Classification des matériels et exigences  
(CEI 60825-1:2014)

Sicherheit von Lasereinrichtungen - Teil 1: Klassifizierung von Anlagen und Anforderungen  
(IEC 60825-1:2014)

This European Standard was approved by CENELEC on 2014-06-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Foreword

The text of document 76/502/FDIS, future edition 3 of IEC 60825-1, prepared by IEC/TC 76 "Optical radiation safety and laser equipment" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60825-1:2014.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2015-03-19
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2019-06-19

This document supersedes EN 60825-1:2007.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

## Endorsement notice

The text of the International Standard IEC 60825-1:2014 was approved by CENELEC as a European Standard without any modification.

IEC 60027-1	NOTE	Harmonised in EN 60027-1.
IEC 60065	NOTE	Harmonised as EN 60065.
IEC 60079 (Series)	NOTE	Harmonised as EN 60079 (Series).
IEC 60204-1	NOTE	Harmonised as EN 60204-1.
IEC 60601-2-22	NOTE	Harmonised as EN 60601-2-22.
IEC 60825-2	NOTE	Harmonised as EN 60825-2.
IEC 60825-4	NOTE	Harmonised as EN 60825-4.
IEC 60825-12	NOTE	Harmonised as EN 60825-12.
IEC 60950 (Series)	NOTE	Harmonised as EN 60950 (Series).
IEC 61010-1	NOTE	Harmonised as EN 61010-1.
IEC 61508 (Series)	NOTE	Harmonised as EN 61508 (Series).
IEC 62115	NOTE	Harmonised as EN 62115.
IEC 62368-1	NOTE	Harmonised as EN 62368-1.
IEC/ISO 11553 (Series)	NOTE	Harmonised as EN ISO 11553 (Series).
ISO 11146-1	NOTE	Harmonised as EN ISO 11146-1.
ISO 12100	NOTE	Harmonised as EN ISO 12100.
ISO 13694	NOTE	Harmonised as EN ISO 13694.
ISO 13849 (Series)	NOTE	Harmonised as EN ISO 13849 (Series).
ISO 15004-2:2007	NOTE	Harmonised as EN ISO 15004-2:2007.
ISO 80000-1	NOTE	Harmonised as EN ISO 80000-1.

## CONTENTS

FOREWORD.....	6
1 Scope and object.....	8
2 Normative references .....	10
3 Terms and definitions .....	10
4 Classification principles .....	24
4.1 General.....	24
4.2 Classification responsibilities .....	24
4.3 Classification rules.....	24
4.4 Laser products designed to function as conventional lamps .....	29
5 Determination of the accessible emission level and product classification.....	29
5.1 Tests .....	29
5.2 Measurement of laser radiation.....	30
5.3 Determination of the class of the laser product.....	31
5.4 Measurement geometry.....	40
5.4.1 General.....	40
5.4.2 Default (simplified) evaluation.....	41
5.4.3 Evaluation condition for extended sources.....	42
6 Engineering specifications .....	44
6.1 General remarks and modifications .....	44
6.2 Protective housing .....	44
6.2.1 General .....	44
6.2.2 Service .....	45
6.2.3 Removable laser system.....	45
6.3 Access panels and safety interlocks.....	45
6.4 Remote interlock connector.....	46
6.5 Manual reset.....	46
6.6 Key control .....	46
6.7 Laser radiation emission warning .....	47
6.8 Beam stop or attenuator.....	47
6.9 Controls .....	47
6.10 Viewing optics.....	47
6.11 Scanning safeguard .....	47
6.12 Safeguard for Class 1C products .....	48
6.13 "Walk-in" access .....	48
6.14 Environmental conditions .....	48
6.15 Protection against other hazards.....	48
6.15.1 Non-optical hazards.....	48
6.15.2 Collateral radiation .....	49
6.16 Power limiting circuit.....	49
7 Labelling.....	49
7.1 General.....	49
7.2 Class 1 and Class 1M .....	51
7.3 Class 1C .....	52
7.4 Class 2 and Class 2M .....	53
7.5 Class 3R.....	53
7.6 Class 3B .....	54

7.7	Class 4 .....	54
7.8	Aperture label .....	55
7.9	Radiation output and standards information .....	55
7.10	Labels for access panels .....	56
	7.10.1 Labels for panels .....	56
	7.10.2 Labels for safety interlocked panels .....	57
7.11	Warning for invisible laser radiation .....	57
7.12	Warning for visible laser radiation .....	57
7.13	Warning for potential hazard to the skin or anterior parts of the eye .....	57
8	Other informational requirements .....	58
	8.1 Information for the user .....	58
	8.2 Purchasing and servicing information .....	59
9	Additional requirements for specific laser products .....	60
	9.1 Other parts of the standard series IEC 60825 .....	60
	9.2 Medical laser products .....	60
	9.3 Laser processing machines .....	60
	9.4 Electric toys .....	60
	9.5 Consumer electronic products .....	60
Annex A (informative)	Maximum permissible exposure values .....	61
	A.1 General remarks .....	61
	A.2 Limiting apertures .....	66
	A.3 Repetitively pulsed or modulated lasers .....	67
	A.4 Measurement conditions .....	68
	A.4.1 General .....	68
	A.4.2 Limiting aperture .....	68
	A.4.3 Angle of acceptance .....	68
	A.5 Extended source lasers .....	69
Annex B (informative)	Examples of calculations .....	70
	B.1 Symbols used in the examples of this annex .....	70
	B.2 Classification of a laser product – Introduction .....	71
	B.3 Examples .....	75
Annex C (informative)	Description of the classes and potentially associated hazards .....	80
	C.1 General .....	80
	C.2 Description of classes .....	80
	C.2.1 Class 1 .....	80
	C.2.2 Class 1M .....	80
	C.2.3 Class 1C .....	80
	C.2.4 Class 2 .....	81
	C.2.5 Class 2M .....	81
	C.2.6 Class 3R .....	81
	C.2.7 Class 3B .....	82
	C.2.8 Class 4 .....	82
	C.2.9 Note on nomenclature .....	82
	C.3 Limitations of the classification scheme .....	84
	C.4 References .....	85
Annex D (informative)	Biophysical considerations .....	86
	D.1 Anatomy of the eye .....	86
	D.2 The effects of laser radiation on biological tissue .....	87

D.2.1	General .....	87
D.2.2	Hazards to the eye .....	89
D.2.3	Skin hazards.....	92
D.3	MPEs and irradiance averaging .....	93
D.4	Reference documents .....	93
Annex E (informative)	MPEs and AELs expressed as radiance .....	95
E.1	Background.....	95
E.2	Radiance values .....	95
E.3	Rationale .....	96
Annex F (informative)	Summary tables.....	99
Annex G (informative)	Overview of associated parts of IEC 60825.....	102
Bibliography.....		104
Figure 1	– Measurement set-up to limit angle of acceptance by imaging the apparent source onto the plane of the field stop .....	43
Figure 2	– Measurement set-up to limit angle of acceptance by placing a circular aperture or a mask (serving as field stop) close to the apparent source .....	43
Figure 3	– Warning label – Hazard symbol.....	50
Figure 4	– Explanatory label .....	51
Figure 5	– Alternative label for Class 1 .....	52
Figure 6	– Alternative label for Class 1M.....	52
Figure 7	– Alternative label for Class 1C.....	52
Figure 8	– Alternative label for Class 2 .....	53
Figure 9	– Alternative label for Class 2M.....	53
Figure 10	– Alternative label for Class 3R .....	54
Figure 11	– Alternative label for Class 3B .....	54
Figure 12	– Alternative label for Class 4 .....	55
Figure 13	– Alternative label for laser aperture .....	55
Figure B.1	– Flowchart guide for the classification of laser products from supplied output parameters.....	72
Figure B.2	– Flowchart guide for the classification of Class 1M and Class 2M laser products.....	73
Figure B.3	– AEL for Class 1 ultra-violet laser products for selected emission durations from $10^{-9}$ s to $10^3$ s .....	74
Figure B.4	– AEL for Class 1 ultra-violet laser products for emission durations from $10^{-9}$ s to $10^3$ s at selected wavelengths .....	74
Figure B.5	– AEL for Class 1 visible and selected infra-red laser products (case $C_6 = 1$ ).....	75
Figure D.1	– Anatomy of the eye.....	86
Figure D.2	– Diagram of laser-induced damage in biological systems .....	88
Figure E.1	– Radiance as a function of wavelength .....	95
Table 1	– Additivity of effects on eye and skin of radiation of different spectral regions.....	25
Table 2	– Times below which pulse groups are summed .....	28
Table 3	– Accessible emission limits for Class 1 and Class 1M laser products and $C_6 = 1$ .....	34

Table 4 – Accessible emission limits for Class 1 and Class 1M laser products in the wavelength range from 400 nm to 1 400 nm (retinal hazard region): extended sources .....	35
Table 5 – Accessible emission limits for Class 2 and Class 2M laser products .....	36
Table 6 – Accessible emission limits for Class 3R laser products and $C_6 = 1$ .....	37
Table 7 – Accessible emission limits for Class 3R laser products in the wavelength range from 400 nm to 1 400 nm (retinal hazard region): extended sources .....	38
Table 8 – Accessible emission limits for Class 3B laser products .....	39
Table 9 – Correction factors and breakpoints for use in AEL and MPE evaluations .....	39
Table 10 – Measurement aperture diameters and measurement distances for the default (simplified) evaluation .....	41
Table 11 – Reference points for Condition 3 .....	42
Table 12 – Limiting angle of acceptance $\gamma_{ph}$ .....	43
Table 13 – Requirements for safety interlocking .....	45
Table A.1 – Maximum permissible exposure (MPE) for $C_6 = 1$ at the cornea expressed as irradiance or radiant exposure .....	62
Table A.2 – Maximum permissible exposure (MPE) at the cornea for extended sources in the wavelength range from 400 nm to 1 400 nm (retinal hazard region) expressed as irradiance or radiant exposure <sup>d</sup> .....	63
Table A.3 – Maximum permissible exposure (MPE) of Table A.1 ( $C_6 = 1$ ) for the wavelength range from 400 nm to 1 400 nm expressed as power or energy <sup>a, b</sup> .....	64
Table A.4 – Maximum permissible exposure (MPE) of Table A.2 (extended sources) for the wavelength range from 400 nm to 1 400 nm expressed as power or energy <sup>a, b, c, d, e, f, g</sup> .....	65
Table A.5 – Maximum permissible exposure (MPE) of the skin to laser radiation .....	66
Table A.6 – Aperture diameters for measuring laser irradiance and radiant exposure .....	67
Table D.1 – Summary of pathological effects associated with excessive exposure to light ...	90
Table D.2 – Explanation of measurement apertures applied to the eye MPEs .....	93
Table E.1 – Maximum radiance of a diffused source for Class 1 .....	96
Table F.1 – Summary of the physical quantities used in this Part 1 .....	99
Table F.2 – Summary of manufacturer's requirements (1 of 2) .....	100
Table G.1 – Overview of additional data in associated parts of IEC 60825 .....	103



## SAFETY OF LASER PRODUCTS –

### Part 1: Equipment classification and requirements

#### 1 Scope and object

IEC 60825-1 is applicable to safety of laser products emitting laser radiation in the wavelength range 180 nm to 1 mm.

Although lasers exist which emit at wavelengths less than 180 nm (within the vacuum ultraviolet), these are not included in the scope of the standard since the laser beam normally has to be enclosed in an evacuated enclosure, and, therefore, the potential optical radiation hazards are inherently minimal.

A laser product may consist of a single laser with or without a separate power supply or may incorporate one or more lasers in a complex optical, electrical, or mechanical system. Typically, laser products are used for demonstration of physical and optical phenomena, materials processing, data reading and storage, transmission and display of information, etc. Such systems have found use in industry, business, entertainment, research, education, medicine and consumer products.

Laser products that are sold to other manufacturers for use as components of any system for subsequent sale are not subject to IEC 60825-1, since the final product will itself be subject to this standard. Laser products that are sold by or for manufacturers of end products for use as repair parts for the end products are also not subject to IEC 60825-1. However, if the laser system within the laser product is operable when removed from the end product, the requirements of this Part 1 apply to the removable laser system.

NOTE 1 Operable equipment does not require a tool to prepare for operation.

Any laser product is exempt from all further requirements of this Part 1 if classification by the manufacturer of that product according to Clauses 4 and 5 shows that the emission level does not exceed the AEL (accessible emission limit) of Class 1 under all conditions of operation, maintenance, service and failure. Such a laser product may be referred to as an exempt laser product.

NOTE 2 The above exemption is to ensure that inherently safe laser products are exempt from Clauses 6,7,8 and 9.

In addition to the adverse effects potentially resulting from exposure to laser radiation, some laser equipment may also have other associated hazards, such as electricity, chemicals and high or low temperatures. Laser radiation may cause temporary visual impairment, such as dazzle and glare. Such effects depend on the task and ambient lighting level and are beyond the scope of this Part 1. The classification and other requirements of this standard are intended to address only the laser radiation hazards to the eyes and skin. Other hazards are not included within its scope.

This Part 1 describes the minimum requirements. Compliance with this Part 1 may not be sufficient to achieve the required level of product safety. Laser products may also be required to conform to the applicable performance and testing requirements of other applicable product safety standards.

NOTE 3 Other standards may contain additional requirements. For example, a Class 3B or Class 4 laser product may not be suitable for use as a consumer product.

Where a laser system forms a part of equipment which is subject to another IEC product safety standard, e.g. for medical equipment (IEC 60601-2-22), IT equipment (IEC 60950 series), audio and video equipment (IEC 60065), audio-video and IT equipment (IEC 62368-1), equipment for use in hazardous atmospheres (IEC 60079), or electric toys (IEC 62115), this Part 1 will apply in accordance with the provisions of IEC Guide 104<sup>2)</sup> for hazards resulting from laser radiation. If no product safety standard is applicable, then IEC 61010-1 may be applied.

For ophthalmic instruments, to ensure patient safety, ISO 15004-2 should be consulted and the principles of the limits provided there should be applied for laser radiation (see also Annex C and D).

In previous editions, light-emitting diodes (LEDs) were included in the scope of IEC 60825-1, and they may be still included in other parts of the IEC 60825 series. However, with the development of lamp safety standards, optical radiation safety of LEDs in general can be more appropriately addressed by lamp safety standards. The removal of LEDs from the scope of this Part 1 does not preclude other standards from including LEDs whenever they refer to lasers. IEC 62471 may be applied to determine the risk group of an LED or product incorporating one or more LEDs. Some other (vertical) standards may require the application of the measurement, classification, engineering specifications and labelling requirements of this standard (IEC 60825-1) to LED products.

Laser products with accessible radiance below the criteria specified in 4.4, designed to function as conventional light sources, and which satisfy the requirements specified in 4.4 may alternatively be evaluated under the IEC 62471 series of standards, "Photobiological safety of lamps and lamp systems". Such a product remains within the scope of this part of IEC 60825, except that the above-described optical radiation emission need not be considered for classification.

The MPE (maximum permissible exposure) values provided in Annex A were developed for laser radiation and do not apply to collateral radiation. However, if a concern exists that accessible collateral radiation might be hazardous, the laser MPE values may be applied to conservatively evaluate this potential hazard, or the exposure limit values in IEC 62471 should be consulted.

The MPE values in Annex A are not applicable to intentional human exposure to laser radiation for the purpose of medical or cosmetic/aesthetic treatment.

NOTE 4 Informative Annexes A to G have been included for purposes of general guidance and to illustrate many typical cases. However, the annexes are not regarded as definitive or exhaustive.

The objectives of this part of IEC 60825 are the following:

- to introduce a system of classification of lasers and laser products emitting radiation in the wavelength range 180 nm to 1 mm according to their degree of optical radiation hazard in order to aid hazard evaluation and to aid the determination of user control measures;
- to establish requirements for the manufacturer to supply information so that proper precautions can be adopted;
- to ensure, through labels and instructions, adequate warning to individuals of hazards associated with accessible radiation from laser products;
- to reduce the possibility of injury by minimizing unnecessary accessible radiation and to give improved control of the laser radiation hazards through protective features.

---

2) IEC Guide 104:2010, *The preparation of safety publications and the use of basic safety publications and group safety publications*

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050 (all parts), *International Electrotechnical Vocabulary* (available at <http://www.electropedia.org>)

IEC 62471 (all parts), *Photobiological safety of lamps and lamp systems*

## 3 Terms and definitions

For the purposes of this document, the definitions given in IEC 60050-845 as well as the following apply.

NOTE For convenience here, the definitions have been arranged in English alphabetical order. Departures from IEC 60050-845 are intentional and are indicated. In such cases, reference is made, between brackets, to the definition of Part 845 of IEC 60050, with the mention “modified”.

### 3.1 access panel

part of the protective housing which provides access to laser radiation when removed or displaced

### 3.2 accessible emission

level of radiation determined at a position and with aperture stops (when the AEL is given in units of Watts or Joules) or limiting apertures (when the AEL is given in units of  $W \cdot m^{-2}$  or  $J \cdot m^{-2}$ ) as described in Clause 5

Note 1 to entry: The accessible emission is determined where human access is considered, as specified in Definition 3.40. The accessible emission (determined during operation) is compared with the accessible emission limit (Entry 3.3) in order to determine the class of the laser product. In the body of the standard, whenever the term “emission level” is used, it is to be understood as accessible emission.

Note 2 to entry: When the beam is larger than the aperture stop, the accessible emission when given in units of watts or joules is less than the total emitted power or energy of the laser product. When the beam is smaller than the limiting aperture, the accessible emission when given in units of  $W \cdot m^{-2}$  or  $J \cdot m^{-2}$ , i.e. as irradiance or radiant exposure averaged over the limiting aperture, is smaller than the actual irradiance or radiant exposure of the beam. See also aperture stop (3.9) and limiting aperture (3.55).

### 3.3 accessible emission limit

#### AEL

maximum accessible emission permitted within a particular class

Note 1 to entry: Wherever the text refers to “emission level not exceeding the AEL” or similar wording, it is implicit that the accessible emission is determined following the measurement criteria specified in Clause 5.

### 3.4 administrative control

safety measures of a non-engineering type such as: key supervision; safety training of personnel; warning notices; count-down procedures; and range safety controls

Note 1 to entry: These may be specified by the manufacturer (see Clause 8).