

**MEDITSIINIS KASUTATAVAD LIIKLUSVAHENDID JA
NENDE VARUSTUS. AEROKIIRABI. OSA 2: AEROKIIRABI
TOIMIMIS- JA TEHNILISED NÕUDED**

**Medical vehicles and their equipment - Air ambulances
- Part 2: Operational and technical requirements for air
ambulances**

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

See Eesti standard EVS-EN 13718-2:2015 sisaldab Euroopa standardi EN 13718-2:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 13718-2:2015 consists of the English text of the European standard EN 13718-2:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 18.03.2015.	Date of Availability of the European standard is 18.03.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.01, 11.160, 49.020

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Medical vehicles and their equipment - Air ambulances - Part 2: Operational and technical requirements for air ambulances

Véhicules sanitaire et leur équipement - Ambulances
aérienne - Partie 2: Exigences techniques et
opérationnelles pour les ambulances aérienne

Medizinische Fahrzeuge und ihre Ausrüstung -
Luftfahrzeuge zum Patiententransport - Teil 2: Operationelle
und technische Anforderungen an Luftfahrzeuge zum
Patiententransport

This European Standard was approved by CEN on 26 December 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	4
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	9
4 General requirements for air ambulances	11
4.1 General.....	11
4.2 Environmental conditions in the patient compartment	11
4.2.1 Temperature and humidity.....	11
4.2.2 Air conditioning system (ACS).....	11
4.2.3 Variable atmospheric pressure	11
4.2.4 Interior light	11
4.2.5 Ventilation.....	12
4.2.6 Noise exposure	12
4.3 Requirements for electrical power supply for medical devices in the patient compartment	12
4.4 Electromagnetic interference	13
4.5 Rail systems	13
4.6 Mechanical vibration	13
4.7 Requirements for fixation of medical devices	14
4.8 Restraint systems in the patient compartment	14
4.9 Patient compartment	14
4.9.1 General.....	14
4.9.2 Hygienic	15
4.9.3 Patient loading and unloading	15
4.9.4 Communication systems	15
4.9.5 Fire safety requirements	15
4.9.6 Emergency exit	15
4.10 Patient treatment area	15
4.10.1 General.....	15
4.10.2 Dimensions.....	15
4.11 Lists of equipment.....	17
5 Air ambulances, operation and performance requirements	17
5.1 Personnel.....	17

5.1.1	Flight crew	17
5.1.2	Medical crew	17
5.2	Specific requirements for helicopters operated in Helicopter Emergency Medical Service (HEMS)	18
5.3	Specific requirements for Helicopter Intensive Care Medical Service (HICAMS)	18
5.4	Specific requirements for Fixed Wing Air Ambulances (FWAA)	18
6	Gas installations in air ambulances	19
6.1	System components	19
6.2	General requirements	19
6.2.1	Capacity and pressure operating range	19
6.2.2	Continuity of supply	20
6.3	Supply systems with gas cylinders	20
6.4	Supply systems for compressed medical air	20
6.5	Supply systems with air compressor	21
6.6	Pipeline distribution system	21
6.7	Marking and colour coding	22
6.8	Alarms	22
6.9	Testing	22
6.9.1	General	22
6.9.2	Test for mechanical integrity for compressed medical gas systems	22
6.9.3	Test for leakage on all pipeline systems and for mechanical integrity of vacuum pipeline systems	22
6.9.4	Leakage from the compressed medical gas pipelines	23
6.9.5	Test for cross connection	23
6.10	Maintenance	23
	Annex A (normative) Medical devices in air ambulances	24
A.1	Introduction	24
A.2	Additional equipment	28
	Annex B (normative) Medicinal products and equipment additional to medical devices in air ambulances	29
B.1	Introduction	29
	Annex C (informative) A–deviations	32
C.1	Deviation in Germany	32
C.1.1	Additional specifications for the medical crew	32
	Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC	33
	Bibliography	34

Foreword

This document (EN 13718-2:2015) has been prepared by Technical Committee CEN/TC 239 "Rescue systems", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2015, and conflicting national standards shall be withdrawn at the latest by September 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 13718-2:2008.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA which is an integral part of this document.

EN 13718-2:2008 has been technically revised. The following points represent the most important changes in the revision:

- a) clarified unclear issues in this part of the standard and between the two parts of the standard (for example requirements for patient's compartment illumination);
- b) changed text related to enhancing safety related to the risk from rotors on helicopters;
- c) clarified the requirements for the patient compartment;
- d) the standard has been modified/integrated to meet the Medical Devices Directive 93/42/EEC requirements.

EN 13718 consists of the following parts, under the general title: *Medical vehicles and their equipment – Air ambulance*:

- *Part 1: Requirements for medical devices used in air ambulances;*
- *Part 2: Operational and technical requirements for air ambulances.*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This part of EN 13718 provides requirements for air ambulances, and in particular covers requirements for the ambulance role of the aircraft.

Air ambulances are equipped with medical devices as well as drugs and rescue equipment to be used by medical personnel. Requirements for medical devices intended for use in air ambulances are provided in EN 13718-1. This standard is supplementary to several European Standards as well as laws and regulations providing the requirements for aircraft in order to provide continuous patient care and monitoring during transport in and between various ambulance types and hospitals. The requirements cover ambulance flights in general. Several national and regional rules and regulations apply to aircraft being used as ambulances. This part of EN 13718 gives information on these in the annexes and in notes throughout the text. Provisions for the safety and care both of the patient as well as of the crew and the medical personnel are contained in existing national and international laws, regulations and guidelines.

This part of EN 13718 provides some general requirements for the safe operation of aircraft being used as ambulances. These requirements are not covered by the scope of the Medical Devices Directive or by international agreements for craft, transportation and traffic. They are provided in order to secure the safe and secure handling of patients. In order to accommodate continuity of patient care between different kinds of ambulances, some specific requirements are given. Requirements are set in order to secure safe use and handling of medical devices.

Aircraft being used as ambulances are equipped with medical devices, medicinal products and rescue equipment to enable the medical personnel to provide continuous patient care. The minima for the medical devices are specified in Annex A. The requirements set out in this part of EN 13718 give the minimum provisions for an ambulance service to provide satisfactory care and medical attention to emergency patients as well as other patients during transportation. The requirements are based on the state of the art of today and common practice in Europe.

This European Standard gives minimum requirements for interfaces and compatibility of medical devices used in air ambulances. The standards work was called for by the EU Commission by a mandate linked with the Medical Devices Directive (see Annex ZA and Bibliography [1]).

This European Standard is supplementary to several other European Standards and gives requirements for medical devices when used in situations where the ambient conditions differ from the normal indoor conditions prevailing within the health care system. Several specific requirements are related to the conditions prevailing in air ambulances. The requirements set are carefully selected to ensure interoperability and continuous patient care.

Medical devices need to conform to the applicable essential requirements in the Medical Devices Directive, 93/42/EEC. The essential requirements are listed in Annex I of the Medical Devices Directive. Annex ZA in this European Standard lists the essential requirements that are covered by the identified clauses of this European Standard.

The environmental conditions for medical devices used in air ambulances are different from those expected in a normal hospital environment. In particular, this implies environmental conditions such as temperature and humidity, vibration and shock caused by movement of the air ambulances, variable atmospheric pressures and electromagnetic disturbances between the air ambulances and the medical device.

1 Scope

This part of EN 13718 specifies the requirements for performance and equipping for air ambulances, including requirements for interfaces to medical devices used for the transport and treatment of sick or injured persons. This part of EN 13718 is applicable to air ambulances capable of transporting at least one person on a stretcher.

NOTE Requirements are specified for categories of air ambulances based on the different intended use. These are the helicopter emergency medical service (HEMS) the helicopter intensive care medical service (HICAMS) and the fixed wing air ambulance (FWAA).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 3-7:2004+A1:2007, *Portable fire extinguishers — Part 7: Characteristics, performance requirements and test methods*

EN 3-8:2006, *Portable fire extinguishers — Part 8: Additional requirements to EN 3-7 for the construction, resistance to pressure and mechanical tests for extinguishers with a maximum allowable pressure equal to or lower than 30 bar*

EN 3-9:2006, *Portable fire extinguishers — Part 9: Additional requirements to EN 3-7 for pressure resistance of CO₂ extinguishers*

EN 3-10:2009, *Portable fire extinguishers — Part 10: Provisions for evaluating the conformity of a portable fire extinguisher to EN 3-7*

EN 143:2000, *Respiratory protective devices — Particle filters — Requirements, testing, marking*

EN 374-1:2003, *Protective gloves against chemicals and micro-organisms — Part 1: Terminology and performance requirements*

EN 455-1:2000, *Medical gloves for single use — Part 1: Requirements and testing for freedom from holes*

EN 455-2:2009+A2:2013, *Medical gloves for single use — Part 2: Requirements and testing for physical properties*

EN 455-3:2006, *Medical gloves for single use — Part 3: Requirements and testing for biological evaluation*

EN 794-3:1998+A2:2009, *Lung ventilators — Part 3: Particular requirements for emergency and transport ventilators*

EN 1618:1997, *Catheters other than intravascular catheters — Test methods for common properties*

EN 1707:1996, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Lock fittings*

EN 1865-1:2010, *Patient handling equipment used in road ambulances — Part 1: General stretcher systems and patient handling equipment*

EN 13718-1:2014, *Medical vehicles and their equipment — Air ambulances — Part 1: Requirements for medical devices used in air ambulances*

EN 13976-1:2011, *Rescue systems — Transportation of incubators — Part 1: Interface conditions*

EN 13976-2:2011, *Rescue systems — Transportation of incubators — Part 2: System requirements*

EN 14605:2005+A1:2009, *Protective clothing against liquid chemicals — Performance requirements for clothing with liquid-tight (Type 3) or spray-tight (Type 4) connections, including items providing protection to parts of the body only (Types PB [3] and PB [4])*

EN 20594-1:1993, *Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements (ISO 594-1:1986)*

EN 60601-2-4:2011, *Medical electrical equipment — Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators (IEC 60601-2-4:2010)*

EN 60601-2-12:2006, *Medical electrical equipment — Part 2-12: Particular requirements for the safety of lung ventilators — Critical care ventilators (IEC 60601-2-12:2001)*

EN 60601-2-24:1998, *Medical electrical equipment — Part 2-24: Particular requirements for the safety of infusion pumps and controllers (IEC 60601-2-24:1998)*

EN 60601-2-34:2014, *Medical electrical equipment — Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment (IEC 60601-2-34:2014)*

EN 80601-2-30:2010, *Medical electrical equipment — Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers (IEC 80601-2-30:2009)*

EN ISO 5356-1:2004, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets (ISO 5356-1:2004)*

EN ISO 5359:2014, *Anaesthetic and respiratory equipment — Low-pressure hose assemblies for use with medical gases (ISO 5359:2014)*

EN ISO 5361:2012, *Anaesthetic and respiratory equipment — Tracheal tubes and connectors (ISO 5361:2012)*

EN ISO 5364:2011, *Anaesthetic and respiratory equipment — Oropharyngeal airways (ISO 5364:2008)*

EN ISO 5366-1:2009, *Anaesthetic and respiratory equipment — Tracheostomy tubes — Part 1: Tubes and connectors for use in adults (ISO 5366-1:2000)*

EN ISO 5367:2014, *Anaesthetic and respiratory equipment — Breathing sets and connectors (ISO 5367:2014)*

EN ISO 6009:1994, *Hypodermic needles for single use — Colour coding for identification (ISO 6009:1992)*

EN ISO 7376:2009, *Anaesthetic and respiratory equipment — Laryngoscopes for tracheal intubation (ISO 7376:2009)*

EN ISO 7396-1:2007, *Medical gas pipeline systems — Part 1: Pipeline systems for compressed medical gases and vacuum (ISO 7396-1:2007)*

EN ISO 7396-2:2007, *Medical gas pipeline systems — Part 2: Anaesthetic gas scavenging disposal systems (ISO 7396-2:2007)*

EN ISO 7864:1995, *Sterile hypodermic needles for single use (ISO 7864:1993)*

EN ISO 7886-1:1997, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use (ISO 7886-1:1993, including Technical Corrigendum 1:1995)*

EN ISO 7886-2:1997, *Sterile hypodermic syringes for single use — Part 2: Syringes for use with power-driven syringe pumps (ISO 7886-2:1996)*

EN ISO 8537:2008, *Sterile single-use syringes, with or without needle, for insulin (ISO 8537:2007)*

EN ISO 9170-1:2008, *Terminal units for medical gas pipeline systems — Part 1: Terminal units for use with compressed medical gases and vacuum (ISO 9170-1:2008)*

EN ISO 9360-1:2009, *Anaesthetic and respiratory equipment — Heat and moisture exchangers (HMEs) for humidifying respired gases in humans — Part 1: HMEs for use with minimum tidal volumes of 250 ml (ISO 9360-1:2000)*

EN ISO 10079-1:2009, *Medical suction equipment — Part 1: Electrically powered suction equipment — Safety requirements (ISO 10079-1:1999)*

EN ISO 10079-2:2014, *Medical suction equipment — Part 2: Manually powered suction equipment (ISO 10079-2:2014)*

EN ISO 10079-3:2014, *Medical suction equipment — Part 3: Suction equipment powered from a vacuum or positive pressure gas source (ISO 10079-3:2014)*

EN ISO 10524-1:2006, *Pressure regulators for use with medical gases — Part 1: Pressure regulators and pressure regulators with flow-metering devices (ISO 10524-1:2006)*

EN ISO 10524-2:2006, *Pressure regulators for use with medical gases — Part 2: Manifold and line pressure regulators (ISO 10524-2:2005)*

EN ISO 10524-3:2006, *Pressure regulators for use with medical gases — Part 3: Pressure regulators integrated with cylinder valves (ISO 10524-3:2005)*

EN ISO 10524-4:2008, *Pressure regulators for use with medical gases — Part 4: Low-pressure regulators (ISO 10524-4:2008)*

EN ISO 10555-1:2013, *Intravascular catheters — Sterile and single-use catheters — Part 1: General requirements (ISO 10555-1:2013, Corrected version 2014-01-15)*

EN ISO 10555-3:2013, *Intravascular catheters — Sterile and single-use catheters — Part 3: Central venous catheters (ISO 10555-3:2013)*

EN ISO 10555-5:2013, *Intravascular catheters — Sterile and single-use catheters — Part 5: Over-needle peripheral catheters (ISO 10555-5:2013)*

EN ISO 11070:2014, *Sterile single-use intravascular introducers, dilators and guidewires (ISO 11070:2014)*

EN ISO 13688:2013, *Protective clothing — General requirements (ISO 13688:2013)*

EN ISO 15002:2008, *Flow-metering devices for connection to terminal units of medical gas pipeline systems (ISO 15002:2008)*

EN ISO 18777:2009, *Transportable liquid oxygen systems for medical use — Particular requirements (ISO 18777:2005)*

EN ISO 19054:2006, *Rail systems for supporting medical equipment (ISO 19054:2005)*

EN ISO 23328-1:2008, *Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to assess filtration performance (ISO 23328-1:2003)*

EN ISO 23328-2:2009, *Breathing system filters for anaesthetic and respiratory use — Part 2: Non-filtration aspects (ISO 23328-2:2002)*

EN ISO 80601-2-55:2011, *Medical electrical equipment — Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors (ISO 80601-2-55:2011)*

EN ISO 80601-2-56:2012, *Medical electrical equipment — Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement (ISO 80601-2-56:2009)*

EN ISO 80601-2-61:2011, *Medical electrical equipment — Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment (ISO 80601-2-61:2011)*

EN ISO 81060-1:2012, *Non-invasive sphygmomanometers — Part 1: Requirements and test methods for non-automated measurement type (ISO 81060-1:2007)*

EN ISO 81060-2:2014, *Non-invasive sphygmomanometers — Part 2: Clinical investigation of automated measurement type (ISO 81060-2:2013)*

IEC 60601-1-12:2014, *Medical electrical equipment — Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

European Aviation Safety Agency (EASA) Certification Specifications CS-23, *Normal, Utility, Aerobatic, and Commuter Category Aeroplanes*

European Aviation Safety Agency (EASA) Certification Specifications CS-25, *Large Aeroplanes*

European Aviation Safety Agency (EASA) Certification Specifications CS-27, *Small Rotorcraft*

European Aviation Safety Agency (EASA) Certification Specifications CS-29, *Large Rotorcraft*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

air ambulance

aircraft designed to be normally staffed by two medically trained personnel equipped and intended for the transportation of at least one stretcher patient who will receive medical treatment during transport

3.2

fixed wing air ambulance

FWAA

aircraft especially equipped for transportation, medical treatment and care of patients, including patients requiring intensive care treatment

3.3

flight crew

members of the crew intended to operate the aircraft

Note 1 to entry: See Commission Regulation (EU) No 1178/2011.