

MEDITSIINISEADMED. MEDITSIINISEADME MÄRGISEL,
MÄRGISTUSEL NING KAASUVAS TEABES KASUTATAVAD
TINGMÄRGID. OSA 1: ÜLDNÕUDED

Medical devices - Symbols to be used with medical
device labels, labelling and information to be supplied -
Part 1: General requirements (ISO 15223-1:2016,
Corrected version 2017-03)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 15223-1:2016 sisaldab Euroopa standardi EN ISO 15223-1:2016 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 15223-1:2016 consists of the English text of the European standard EN ISO 15223-1:2016.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 30.11.2016.	Date of Availability of the European standard is 30.11.2016.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 01.080.20, 11.040.01

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English version

Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements (ISO 15223-1:2016, Corrected version 2017-03)

Dispositifs médicaux - Symboles à utiliser avec les étiquettes, l'étiquetage et les informations à fournir relatifs aux dispositifs médicaux - Partie 1: Exigences générales (ISO 15223-1:2016, Version corrigée 2017-03)

Medizinprodukte - Bei Aufschriften von Medizinprodukten zu verwendende Symbole, Kennzeichnung und zu liefernde Informationen - Teil 1: Allgemeine Anforderungen (ISO 15223-1:2016, korrigierte Fassung 2017-03)

This European Standard was approved by CEN on 22 October 2016.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



**CEN-CENELEC Management Centre:
Avenue Marnix 17, B-1000 Brussels**

European foreword

This document (EN ISO 15223-1:2016) has been prepared by Technical Committee ISO/TC 210 “Quality management and corresponding general aspects for medical devices” in collaboration with Technical Committee CEN/CLC/TC 3 “Quality management and corresponding general aspects for medical devices” the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2017, and conflicting national standards shall be withdrawn at the latest by May 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15223-1:2012.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, Annex ZB and Annex ZC, which are integral parts of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard “within the meaning of Annex ZA/Annex ZB/Annex ZC”, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table 1 — Correlations between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 7000	—	ISO 7000:2014 ^a
ISO 8601	—	ISO 8601:2004
ISO 14971	EN ISO 14971:2012	ISO 14971:2007
ISO 15223-2	—	ISO 15223-2:2010
^a Available only in database format from ISO or IEC.		

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 15223-1:2016, Corrected version 2017-03 has been approved by CEN as EN ISO 15223-1:2016 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC [OJ L 169] on Medical Devices

This European Standard has been prepared under a Commission's standardization request 'M/023 concerning the development of European standards related to medical devices' to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced "as far as possible", "to a minimum", "to the lowest possible level", "minimized" or "removed", according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements (ERs) of Directive 93/42/EEC	Clause(s)/subclause(s) of this European Standard	Qualifying remarks/Notes
8.7	5.2.7	Provided that the symbol is provided according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC and only for non-sterile products.
13.2	4.2, 4.3	Only the first sentence of this ERs is covered, provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC.

13.3 (a)	5.1.1, 5.1.2	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC.
13.3 (c)	5.2.1, 5.2.2, 5.2.3, 5.2.4, 5.2.5, 5.2.9	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC.
13.3 (d)	5.1.5, 5.1.7	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC. If a Serial number is not provided the symbol for 'LOT' must precede the batch code.
13.3 (e)	5.1.4	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC, the "use-by" date must be expressed as, at least, the year and the month.
13.3 (f)	5.4.2	Only the first sentence of this ER is covered, provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC.
13.3 (i)	5.3.1, 5.3.2, 5.3.3, 5.3.4, 5.3.5, 5.3.6, 5.3.7, 5.3.8, 5.3.9	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC, the ER is only covered with respect to the conditions indicated by the symbols. For other conditions, other symbols or other means of indication may be needed.

13.3 (k)	5.2.6, 5.2.7, 5.2.8, 5.4.1, 5.4.4, 5.4.5	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC, the ER is only covered with respect to the warnings indicated by the symbols. For other warnings, other symbols or other means of indication may be needed.
13.3 (l)	5.1.3	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC. Active medical devices must be labelled with at least the year of manufacture unless a “use-by” date (5.1.4) is given. The date of manufacture may be included in the batch or serial number (5.1.5, 5.1.7).
13.3 (m)	5.2.2, 5.2.3, 5.2.4, 5.2.5	Provided that the symbol is provided on the label and according to the general requirements indicated in the ER 13.1 of Directive 93/42/EEC, the ER is only covered with respect to the conditions indicated by the symbols.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

Annex ZB (informative)

Relationship between this European standard and the essential requirements of Directive 90/385/EEC [OJ L 189] aimed to be covered

This European standard has been prepared under a Commission's standardisation request 'M/023 concerning the development of European standards related to medical devices' to provide one voluntary means of conforming to essential requirements of Council Directive of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices [OJ L 189].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZB.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 90/385/EEC as amended by 2007/47/EC. This means that risks have to be reduced "as far as possible", "to a minimum", "to the lowest possible level", "minimized" or "removed", according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with Essential Requirements 1, 4, 5, 8, 9 and 10 of the Directive.

NOTE 3 This Annex ZB is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZB.1, it means that it is not addressed by this European Standard.

Table ZB.1 — Correspondence between this European Standard and Directive 90/385/EEC on active implantable medical devices

Essential Requirements (ERs) of Directive 90/385/EEC	Clause(s)/subclause(s) of this European Standard	Qualifying remarks/Notes
11	5.1.5, 5.1.6, 5.1.7	ER is covered only for indication of batch code or serial number. Components are not covered".
14.1, 1st indent	5.2.2, 5.2.3, 5.2.4, 5.2.5	Provided that the symbol is provided on the sterile pack, This ER is only covered with respect to the conditions indicated by the symbols. For other warnings, other symbols or other means of indication may be needed.
14.1, 2nd indent	5.2.1, 5.2.2, 5.2.3, 5.2.4, 5.2.5	Provided that the symbol is provided on the sterile pack.