

**Loomasööt. Deoksünivalenooli määramine söötades.
HPLC meetod koos UV detektoriga ja puhastamisega
immunoafiinsuskolonnis**

**Animal feeding stuffs - Determination of Deoxynivalenol
in animal feed - HPLC method with UV detection and
immunoaffinity column clean-up**

EVS

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

S E	E	EVS-EN 15791:2009 EN 15791:2009	E E	EVS-EN 15791:2009 E	EN
S		EVS	S	E	
E E		02 09 2009	02 09 2009	E	
S		E S	S	E	

EVS-

S 5 120

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

E S
10 10 17 E 05 5050 - E S :

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

N
E S
10 10 17 E 05 5050 - S :

EVS

English Version

**Foodstuffs - Determination of Deoxynivalenol in animal feed -
HPLC method with immunoaffinity column clean-up**

Produits alimentaires - Dosage du désoxynivalénol dans les
aliments pour animaux - Méthode de chromatographie
liquide haute performance avec détection UV et purification
sur colonne d'immuno-affinité

Futtermittel - Bestimmung von Deoxynivalenol in
Futtermitteln - HPLC-Verfahren mit Reinigung an einer
Immunoaffinitätssäule

This European Standard was approved by CEN on 1 August 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	4
5 Apparatus	6
6 Procedure	7
7 HPLC determination	8
8 Calculation	9
9 Precision	10
10 Test report	10
Annex A (informative) Precision data	12
Annex B (informative) Chromatogram	14
Bibliography	15

EVS

Foreword

This document (EN 15791:2009) has been prepared by Technical Committee CEN/TC 327 “Animal feeding stuffs”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2010, and conflicting national standards shall be withdrawn at the latest by March 2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

WARNING — The use of this standard can involve hazardous materials, operations and equipment. This standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

EVS

1 Scope

This Standard is applicable to the determination of deoxynivalenol (DON) in animal compound feed at concentrations of 150 µg/kg up to at least 4 000 µg/kg.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Principle

Deoxynivalenol (DON) is extracted from the commodity using water. The aqueous extract is then cleaned up with an immunoaffinity column to remove impurities from the sample. Subsequently DON is quantitatively determined by HPLC with UV detection.

4 Reagents

During the analysis, unless otherwise stated, use only reagents of recognised analytical grade and only double-distilled water or water of grade 1 as defined in EN ISO 3696. Solvents shall be of quality for HPLC analysis.

4.1 Acetonitrile

WARNING — Acetonitrile is hazardous and handling shall be carried out inside a fume cupboard. Appropriate safety equipment (lab coat, goggles, gloves) shall be worn.

4.2 Deoxynivalenol (DON), with a minimum purity of 97 %

WARNING — Deoxynivalenol is highly toxic. Gloves and safety glasses shall be worn at all times and all standard and sample preparation stages shall be carried out in a fume cupboard.

4.3 Methanol

WARNING — Methanol is hazardous and handling shall be carried out inside a fume cupboard. Appropriate safety equipment (lab coat, goggles, gloves) shall be worn.

4.4 Glacial acetic acid

WARNING — Glacial acetic acid is hazardous and handling shall be carried out inside a fume cupboard. Appropriate safety equipment (lab coat, goggles, gloves) shall be worn.

4.5 Mobile Phase

Mix 15 parts per volume of methanol (4.3) with 84,9 parts per volume of water and 0,1 parts of glacial acetic acid (4.4). The exact amount of methanol used and whether acetic acid will have to be used depends on the HPLC column chosen for analysis and must be adjusted if necessary. Degas this solution before use.