

Ambient air - Standard method for measurement of **NO_3^- , SO_4^{2-} , Cl^- , NH_4^+ , Na^+ , K^+ , Mg^{2+} , Ca^{2+}** in **PM_{2,5}** as deposited on filters

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 16913:2017 sisaldab Euroopa standardi EN 16913:2017 ingliskeelset teksti.	This Estonian standard EVS-EN 16913:2017 consists of the English text of the European standard EN 16913:2017.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 07.06.2017.	Date of Availability of the European standard is 07.06.2017.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.040.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Ambient air - Standard method for measurement of NO_3^- ,
 SO_4^{2-} , Cl^- , NH_4^+ , Na^+ , K^+ , Mg^{2+} , Ca^{2+} in PM_{2,5} as deposited
on filters

Air ambiant - Méthode normalisée pour le mesurage de
 NO_3^- , SO_4^{2-} , Cl^- , NH_4^+ , Na^+ , K^+ , Mg^{2+} , Ca^{2+} dans la
fraction PM_{2,5} telle que déposée sur des filtres

Außenluft - Messverfahren zur Bestimmung von NO_3^- ,
 SO_4^{2-} , Cl^- , NH_4^+ , Na^+ , K^+ , Mg^{2+} , Ca^{2+} in PM_{2,5} nach
Abscheidung auf Filtern

This European Standard was approved by CEN on 27 February 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

European foreword	4
Introduction	5
1 Scope	6
2 Normative references	7
3 Terms and definitions	7
4 Symbols and abbreviations	7
5 Principle	8
6 Equipment	8
7 Sampling	9
8 Transport and storage	10
9 Analysis	10
10 Calculation of results	12
11 Quality control	14
12 Measurement uncertainty	17
13 Artefacts and interferences	18
14 Data recording	19
Annex A (informative) Statistical analysis of anion and cation concentrations collected on filters from field validation exercise	20
A.1 General	20
A.2 Analysis methodology	20
A.2.1 General	20
A.2.2 Calculating between- and within-laboratory variability	20
A.2.2.1 Notation	20
A.2.2.2 Data processing	21
A.2.2.3 Outlier rejection	21
A.2.2.4 Data normalization	22
A.2.2.5 Analysis of variance	22
A.2.2.6 Calculation of standard deviations	23
A.2.3 Calculating between-sampler variability	23
A.2.3.1 Notation	23
A.2.3.2 Data processing	24
A.2.4 Combined standard uncertainty	25
A.3 Remarks	25

A.4	Results	25
A.4.1	Data set 1 – Between laboratory and internal laboratory variability	25
A.4.2	Data set 2 – Between sampler variability	26
A.4.3	Data set 3 – Uncertainty over the measured concentration range	27
A.4.4	Detection limit	31
A.4.5	Field Blanks	31
Annex B (informative)	Uncertainty budget	33
Annex C (informative)	Reagents.....	36
C.1	General	36
C.2	Anion determination by ion chromatography	36
C.3	Cation determination by ion chromatography	37
C.4	Cation determination by inductively coupled plasma optical emission spectrometry (ICP-OES).....	38
C.5	Ammonium determination by photometry.....	39
Annex D (informative)	Other analysis methods used in the validation programme	41
D.1	Inductively coupled plasma optical emission spectrometer system (ICP-OES).....	41
D.2	Conductometry	41
D.3	Photometry	42
D.3.1	Equipment.....	42
D.3.2	Preparation of calibration curve.....	42
D.3.3	Analytical procedure	42
Annex E (informative)	Preparation of stock standard solution.....	43
Annex F (informative)	Sampling artefacts.....	44
F.1	General	44
F.2	Ammonium nitrate.....	44
F.3	Chloride.....	45
	Bibliography	46

European foreword

This document (EN 16913:2017) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2017, and conflicting national standards shall be withdrawn at the latest by December 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This European Standard describes how to measure a specified range of anions and cations in $PM_{2,5}$ as deposited on filters.

The EU Air Quality Directive 2008/50/EC [1] on ambient air quality and cleaner air for Europe requests the measurements of concentrations of NO_3^- , SO_4^{2-} , Cl^- , NH_4^+ , Na^+ , K^+ , Mg^{2+} , Ca^{2+} in $PM_{2,5}$ at rural background locations. In Annex IV of the Directive, guidance for these measurements is given.

Measurements of anions and cations in PM are being performed by the EMEP programme, mainly by using a filterpack with limited particle size selectivity. The cooperative programme for monitoring and evaluation of long-range transmission of air pollutants in Europe (EMEP) was launched in 1977 as a response to the growing concern over the effects on the environment caused by acid deposition. EMEP was organized under the auspices of the United Nations Economic Commission for Europe (UNECE). Today EMEP is an integral component of the cooperation under the Convention on Long-range Transboundary Air Pollution.

Directive 2008/50/EC requires that measurements at rural sites, where appropriate, are coordinated with the monitoring strategy and measurement programme of EMEP. Although, there are different sampling procedures involved, a common approach is used for the analytical procedure.

In order to keep the agreement between existing EMEP data and data to be produced using this European Standard as close as possible, the EMEP protocol has been taken as starting point for this European Standard. This European Standard differs from the EMEP protocol in the sense that measurement of anions and cations are done in $PM_{2,5}$, and that a number of critical parameters (e.g. choice of filter material) are fixed.

Additional attention is given to harmonizing these critical parameters with elemental carbon/organic carbon (EC/OC) measurements and with $PM_{2,5}$ measurements as well, as sampling is usually done simultaneously.

1 Scope

This European Standard specifies a method for the determination of the mass concentration of water soluble NO_3^- (nitrate), SO_4^{2-} (sulphate), Cl^- (chloride), NH_4^+ (ammonium), Na^+ (sodium), K^+ (potassium), Mg^{2+} (magnesium), Ca^{2+} (calcium) in $\text{PM}_{2,5}$ as deposited on filters.

This European Standard describes the analytical procedures for determining anions and cations as part of the $\text{PM}_{2,5}$ particulate phase, sample extraction and analysis of anions and cations by ion chromatography. Sampling onto filters will be done in accordance with EN 12341 for $\text{PM}_{2,5}$.

NOTE 1 Alternatively, cations, excluding ammonium, can be analysed by inductively coupled plasma optical emission spectrometry (ICP-OES). Ammonium can also be analysed by photometry or conductometry.

This European Standard can be used for the measurements of anions and cations as required by Directive 2008/50/EC. The method does not take into account the possible losses during sampling due to evaporation.

NOTE 2 NO_3^- , Cl^- , NH_4^+ are part of the volatile fraction of $\text{PM}_{2,5}$, and the concentrations determined using this standard can be used as minimum values for the concentrations of these ions in $\text{PM}_{2,5}$. NO_3^- , NH_4^+ , Cl^- are usually up to 30 % underestimated due to evaporational losses from the filter during sampling.

This European Standard may be used at rural and urban background sites and road sites that are in accordance with the siting criteria of Directive 2008/50/EC.

This European Standard is applicable to the measurement of anion/cations in $\text{PM}_{2,5}$ samples corresponding to $\text{PM}_{2,5}$ mass concentrations between approximately $1 \mu\text{g}/\text{m}^3$ (i.e. the limit of detection of the standard measurement method (EN 12341) expressed as its uncertainty) up to $120 \mu\text{g}/\text{m}^3$.

The validated range of the anion and cation concentrations based on the field validation measurements is presented in Table 1.

Table 1 — Validated range of anions and cations

Component	Minimum $\mu\text{g}/\text{m}^3$	Maximum $\mu\text{g}/\text{m}^3$
Cl^-	0,001	1,4
NO_3^-	0,002	29
SO_4^{2-}	0,05	13
Na^+	0,003	1,9
NH_4^+	0,04	11
K^+	0,003	0,49
Mg^{2+}	0,001	0,38
Ca^{2+}	0,002	0,72

See Annex A for the statistical analysis of the field validation measurements.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 12341:2014, *Ambient air — Standard gravimetric measurement method for the determination of the PM₁₀ or PM_{2,5} mass concentration of suspended particulate matter*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12341:2014 and the following apply.

NOTE In particular, the following terms of EN 12341 are used in this document: calibration, combined standard uncertainty, coverage factor, expanded uncertainty, PM_x, standard uncertainty, uncertainty (of measurement).

3.1

field filter blank

filter that is taken through the same procedure as a sample, including transport to and from, and storage in the field, and analysis, but is not used for sampling air

Note 1 to entry: The filter is taken from the same batch as used for sampling.

3.2

laboratory filter blank

unused filter that does not leave the laboratory and is taken through the same procedure as a sample

Note 1 to entry: The filter is taken from the same batch as used for sampling.

3.3

reagent blank

solution that contains all the reagents used during the analysis of the sample, but without the sample and filter matrix

4 Symbols and abbreviations

For the purposes of this document, the following abbreviations apply.

EMEP	Cooperative programme for monitoring and evaluation of long-range transmission of air pollutants in Europe
CD	Conductivity Detector
FEP	Fluorinated Ethylene Propylene
HDPE	High Density PolyEthylene
HPLC	High Performance Liquid Chromatography
ICP-MS	Inductively Coupled Plasma Mass Spectrometry
ICP-OES	Inductively Coupled Plasma Optical Emission Spectrometry
NIST	National Institute for Standards and Technology
PE	Polyethylene