Ophthalmic implants - Intraocular lenses - Part 2: Optical et.

October School Sch properties and test methods (ISO 11979-2:2014)



# **EESTI STANDARDI EESSÕNA**

#### **NATIONAL FOREWORD**

See Eesti standard EVS-EN ISO 11979-2:2014 sisaldab Euroopa standardi EN ISO 11979-2:2014 inglisekeelset teksti.	This Estonian standard EVS-EN ISO 11979-2:2014 consists of the English text of the European standard EN ISO 11979-2:2014.		
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.  Date of Availability of the European standard is 20.08.2014.		
,			
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.		

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.70

#### Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega: Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

#### The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation: Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

# EUROPEAN STANDARD NORME EUROPÉENNE

**EUROPÄISCHE NORM** 

**EN ISO 11979-2** 

August 2014

ICS 11.040.70

Supersedes EN ISO 11979-2:1999

#### **English Version**

# Ophthalmic implants - Intraocular lenses - Part 2: Optical properties and test methods (ISO 11979-2:2014)

Implants ophtalmiques - Lentilles intraoculaires - Partie 2: Propriétés optiques et méthodes d'essai (ISO 11979-2:2014) Ophthalmische Implantate - Intraokularlinsen - Teil 2: Optische Eigenschaften und Prüfverfahren (ISO 11979-2:2014)

This European Standard was approved by CEN on 25 July 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

## **Foreword**

This document (EN ISO 11979-2:2014) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2015, and conflicting national standards shall be withdrawn at the latest by February 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11979-2:1999.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

#### **Endorsement notice**

The text of ISO 11979-2:2014 has been approved by CEN as EN ISO 11979-2:2014 without any modification.

Coı	ntents	Page
Fore	eword	iv
Intro	oduction	v
1	Scope	1
2	Normative references	
3	Terms and definitions	1
4	Requirements	
	4.1 General 4.2 Dioptric power	
	4.3 Determination of imaging quality	
	4.4 Spectral transmittance	
	ex A (normative) Measurement of dioptric power	
	ex B (normative) Measurement of resolution efficiency	
	ex C (normative) Measurement of MTF	
Bibli	liography	22
	8	
	<u></u>	
	4	
		6
		$O_{\lambda}$
© ISO	O 2014 – All rights reserved	iii

# Introduction

This part of ISO 11979 initially addressed monofocal IOLs and now has been revised to include the requirements and test methods for spherical monofocal, aspheric monofocal, toric, multifocal, and accommodative IOLs. This part of ISO 11979 contains several test methods for which associated requirements are given and one test method for which no requirement is formulated. The former are directly connected to the optical functions of intraocular lenses. The latter, the test for spectral transmittance, has been provided for information about UV transmission and in specific situations, e.g. when using laser light sources for diagnosis and treatment.

For the original spherical monofocal IOLs, extensive interlaboratory testing was carried out before setting the limits specified. During this testing some basic problems were encountered as described in Reference [1]. The accuracy in the determination of dioptric power has an error that is not negligible in relation to the half dioptre steps in which intraocular lenses are commonly labelled. The dioptric power tolerances take this fact into account. Hence the limits set may lead to some overlap into the next labelled power, especially for high dioptre lenses. Reference [1] gives further discussion on this subject.

The majority of lenses hitherto implanted were qualified using the method described in Annex B or Annex C (model eye 1). The method in Annex B is limited in its applicability, however. The limits for the more general method in Annex C have been set in terms of MTF in a model eye, following two approaches. The first is by correlation to the method and limit in Annex B. Further discussion can be found in Reference [2]. The second is set as a percentage of what is calculated as theoretical maximum for the design, with the rationale that a minimum level of manufacturing accuracy be guaranteed. For common PMMA lenses, these two limits correspond well with each other. For lenses made of materials with lower refractive index, or with certain shape factors, or for extreme power lenses in general, the latter limit is lower than the former. However, such lenses are already in use, indicating clinical acceptance. The question of which is the absolute lowest limit that is compatible with good vision arises. No definite answer can be found, but following clinical data presented to the working group, an absolute lower limit has been set for the calculation method.

# Ophthalmic implants — Intraocular lenses —

# Part 2:

# Optical properties and test methods

# 1 Scope

This part of ISO 11979 specifies requirements and test methods for certain optical properties of intraocular lenses (IOLs) with any of spherical, aspheric, monofocal, toric, multifocal, and/or accommodative optics. The generic descriptor 'IOL' used throughout this document also includes phakic intraocular lenses (PIOL).

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6328:2000, Photography — Photographic materials — Determination of ISO resolving power

ISO 9334, Optics and photonics — Optical transfer function — Definitions and mathematical relationships

ISO 9335, Optics and photonics — Optical transfer function — Principles and procedures of measurement

ISO 11979-1, Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary

ISO 11979-3, Ophthalmic implants — Intraocular lenses — Part 3: Mechanical properties and test methods

ISO 11979-4, Ophthalmic implants — Intraocular lenses — Part 4: Labelling and information

#### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11979-1 and ISO 9334 apply.

## 4 Requirements

### 4.1 General

The manufacturer shall demonstrate that the entire range of available powers meets the specifications herein. All optical properties apply at *in situ* conditions, either by being measured at simulated *in situ* conditions, or being measured at other conditions and then corrected to *in situ* conditions.

For IOLs where the optic is intended to be deformed during implantation, it shall be demonstrated that dioptric power and imaging quality are retained at *in situ* conditions or equivalent following surgical manipulation and recovery. See ISO 11979-3 for more detail.

The test methods described in this standard are reference methods. Alternative methods that produce equivalent results to those obtained with the reference methods can be used if the manufacturer can demonstrate that the IOLs meet the minimum dioptric power and imaging quality requirements.