

First edition
2007-07-15

Corrected version
2009-01-15

**Dentistry — Implantable materials for
bone filling and augmentation in oral and
maxillofacial surgery — Contents of a
technical file**

*Art dentaire — Matériaux implantables de comblement et de
reconstruction osseuse en chirurgie orale et maxillofaciale — Contenu
d'un dossier technique*



Reference number
ISO 22794:2007(E)

© ISO 2007

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Implantable materials.....	3
5 Technical file.....	3
5.1 Contents.....	3
5.2 Chemical composition.....	3
5.3 Physical properties.....	4
5.4 Intended applications, precautions, warnings and instructions.....	4
5.5 Preclinical and clinical evaluation.....	5
5.6 Manufacture.....	6
5.7 Materials of animal origin.....	6
5.8 Sterilization.....	6
5.9 Packaging.....	7
5.10 Additional information supplied by the manufacturer.....	7
Bibliography.....	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22794 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 8, *Dental implants*.

In this corrected version of ISO 22794:2007 changes have been made to the list of Normative references (Clause 2) and to the Bibliography. Consequently cross references in subclauses 5.2, 5.5.5, 5.7, 5.8 and 5.9.2 have been altered.

Further, new subclauses 5.5.3 and 5.5.4 have been designated with the former 5.5.3 becoming 5.5.5.

Introduction

Different materials used for the preservation of masticatory function, such as dental restorative materials and dental implants are subject to standards and regulations, either in existence or in preparation, designed to evaluate the performance of these products.

Implantable materials for bone filling and augmentation in oral and maxillofacial surgery are not covered by the procedures for evaluating and testing dental restorative materials and dental implants; it is necessary to develop a new standard for these materials.

The aim of this International Standard is to define the content of a technical file that demonstrates safety and effectiveness of bone filling and augmentation materials used in oral and maxillofacial surgery.

This document is a preview generated by EVS

This document is a preview generated by EVS

Dentistry — Implantable materials for bone filling and augmentation in oral and maxillofacial surgery — Contents of a technical file

1 Scope

This International Standard applies to implantable materials, whether resorbable or non-resorbable, used as dental devices for filling and augmenting bones in oral and maxillofacial surgery. Products that are essentially pure (> 90 %) hydroxyapatite are not covered by this International Standard.

Evaluation includes the physico-chemical, mechanical, biological and clinical aspects and behaviour of these implantable dental materials.

Materials such as autografts, allografts and membranes, and products for which the primary intended use is to deliver a medicinal product, are not covered by this International Standard.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942 ¹⁾, *Dentistry — Vocabulary*

ISO 10993-1 ²⁾, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management system*

ISO 10993-17, *Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances*

ISO 11135-1 ³⁾, *Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1 ⁴⁾, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11607-2 ⁵⁾, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

1) To be published. (Revises and replaces ISO 1942 parts 1 to 5:1989)

2) To be published. (Revision of ISO 10993-1:2003)

3) Cancels and replaces ISO 11135:1994 and ISO 11135:1994/Cor.1:1994.

4) Cancels and replaces ISO 11137:1995, ISO 11137:1995/Cor.1:1997 and ISO 11137:1995/Amd.1:2001.

5) Cancels and replaces ISO 11607:2003.

ISO 13408-1, *Aseptic processing of health care products — Part 1: General requirements*

ISO 14155-1 ⁶⁾, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*

ISO 14155-2 ⁶⁾, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

ISO 17665-1 ⁷⁾, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

EN 1041, *Information supplied by the manufacturer of medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1 biocompatibility

- a) capability of a material to fulfil its function with an appropriate response for a specific application from the receiving host;
- b) quality of being accepted in a specific living environment without adverse or unwanted side effects

3.2 biomaterial

- a) material intended to interface with the biological system to evaluate, treat, augment or replace tissue, organ or function of the organism;
- b) material specially prepared and/or presented to exhibit bioacceptability, biocompatibility or positive biocompatibility

NOTE The implantable materials referred to in this document are all biomaterials.

3.3 filling

surgical placement of a biomaterial, resorbable or non-resorbable, into an intrabony cavity during oral and maxillofacial surgery

NOTE Intrabony cavity includes extraction socket.

3.4 augmentation

surgical placement of a biomaterial, resorbable or non-resorbable, to increase the volume of bone, usually on the sinus floor or the alveolar ridges

6) Cancels and replaces ISO 14155:1996.

7) Cancels and replaces ISO 11134:1994 and ISO 13683:1997.