

**Soil quality - Determination of carbon and nitrogen by  
near-infrared spectrometry (NIRS) (ISO 17184:2014)**

This document is a preview generated by EVS

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 17184:2014 sisaldab Euroopa standardi EN ISO 17184:2014 inglisekeelset teksti.	This Estonian standard EVS-EN ISO 17184:2014 consists of the English text of the European standard EN ISO 17184:2014.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 21.05.2014.	Date of Availability of the European standard is 21.05.2014.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 13.080.10

### **Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele**

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:  
Aru 10, 10317 Tallinn, Eesti; [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

### **The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation**

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:  
Aru 10, 10317 Tallinn, Estonia; [www.evs.ee](http://www.evs.ee); phone 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

ICS 13.080.10

English Version

## Soil quality - Determination of carbon and nitrogen by near-infrared spectrometry (NIRS) (ISO 17184:2014)

Qualité du sol - Dosage du carbone et de l'azote par spectrométrie proche infrarouge (SPIR) (ISO 17184:2014)

Bodenbeschaffenheit - Bestimmung von Kohlenstoff und Stickstoff durch Nahinfrarotspektroskopie (NIRS) (ISO 17184:2014)

This European Standard was approved by CEN on 24 April 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Foreword

This document (EN ISO 17184:2014) has been prepared by Technical Committee ISO/TC 190 "Soil quality" in collaboration with Technical Committee CEN/TC 345 "Characterization of soils" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2014, and conflicting national standards shall be withdrawn at the latest by November 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### Endorsement notice

The text of ISO 17184:2014 has been approved by CEN as EN ISO 17184:2014 without any modification.

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Principle</b> .....	<b>1</b>
<b>4 Apparatus</b> .....	<b>1</b>
<b>5 Procedure</b> .....	<b>2</b>
5.1 Preparation of samples.....	2
5.2 Instrument calibration.....	2
5.3 Sample measurement.....	4
<b>6 Test report</b> .....	<b>5</b>
<b>Annex A (informative) Precision data</b> .....	<b>6</b>
<b>Bibliography</b> .....	<b>9</b>

# Soil quality — Determination of carbon and nitrogen by near-infrared spectrometry (NIRS)

## 1 Scope

This International Standard specifies a method for the determination of carbon and nitrogen in soils by direct measurement of sample spectra in the near-infrared spectral region. The spectra are evaluated by a suitable calibration model derived from the results obtained by reference methods.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11464, *Soil quality — Pretreatment of samples for physico-chemical analysis*

## 3 Principle

Soil samples are measured by reflectance near-infrared (NIR) spectroscopy. Diffuse reflectance NIR spectroscopy offers a non-destructive means for measurement of soil properties based on reflectance spectra of illuminated soils. Spectral data are evaluated by a suitable calibrating model derived from the measurement of a sufficient number of representative soil samples with known content of carbon and/or nitrogen determined by reference methods. Calibration equations reflect the relationship between the constituents of the sample and NIR spectral information. The soil samples and the set of calibrating samples for the NIR measurement are prepared the same way.

NOTE 1 NIR spectrometry is a very fast non-destructive and environmentally friendly analytical technique when compared to the standard chemical methods used as reference methods.

NOTE 2 Soils generally have similar reflectance spectra in the 1 100 nm to 2 500 nm range. The absorption peaks for soils in the near-infrared region are difficult to assign to specific chemical components.

## 4 Apparatus

**4.1 Near-infrared instrument**, based on measurement of reflectance spectra in the near-infrared region (wavelength range from 900 nm to 2 500 nm is usually applied). The instrument should be equipped with a suitable measurement cell for pulverized solid samples. The instrument should measure sufficiently large sample volume to eliminate any significant influence of inhomogeneity of the sample. The software shall allow instrument tests, calibration, sample measurement and data evaluation.

Resolution of the instrument should be equal to 8 nm or better.

NOTE Wavelengths of spectra recorded in higher resolution may be averaged to reduce spectra noise and there is a risk of over fitting of the calibration model. Instruments with lower resolution may be used if their performance is verified for intended purposes.