

Laboratory glassware - Beakers (ISO 3819:2015)

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 3819:2015 sisaldab Euroopa standardi EN ISO 3819:2015 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 3819:2015 consists of the English text of the European standard EN ISO 3819:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 23.12.2015.	Date of Availability of the European standard is 23.12.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 17.060

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

English Version

## Laboratory glassware - Beakers (ISO 3819:2015)

Verrerie de laboratoire - Béchers (ISO 3819:2015)

Laborgeräte aus Glas - Becher (ISO 3819:2015)

This European Standard was approved by CEN on 5 December 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## European foreword

This document (EN ISO 3819:2015) has been prepared by Technical Committee ISO/TC 48 "Laboratory equipment" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2016, and conflicting national standards shall be withdrawn at the latest by June 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### Endorsement notice

The text of ISO 3819:2015 has been approved by CEN as EN ISO 3819:2015 without any modification.

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Types of beakers</b> .....	<b>1</b>
<b>4 Series of beakers</b> .....	<b>1</b>
<b>5 Capacity</b> .....	<b>1</b>
<b>6 Material</b> .....	<b>2</b>
<b>7 Dimensions</b> .....	<b>2</b>
7.1 General.....	2
7.2 Radius at base.....	2
7.3 Wall thickness.....	2
<b>8 Details of construction</b> .....	<b>3</b>
8.1 Base.....	3
8.2 Upper edge.....	3
8.3 Spout.....	3
8.4 Scale.....	4
8.5 Thermal shock resistance.....	4
<b>9 Marking</b> .....	<b>4</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 48, *Laboratory equipment* in collaboration with CEN/TC 332, *Laboratory equipment*.

This second edition cancels and replaces the first edition (ISO 3819:1985), which has been technically revised to include the following changes:

- additional series with thick-walled beakers has been added;
- dimensions and tolerances have been adjusted to state of manufacturing;
- beakers with a nominal capacity of 500 ml, 5 000 ml, and 10 000 ml have been added;
- thermal shock resistance has been modified.

# Laboratory glassware — Beakers

## 1 Scope

This International Standard specifies requirements for an internationally acceptable series of glass beakers for laboratory use.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 718, *Laboratory glassware — Thermal shock and thermal shock endurance — Test methods*

ISO 3585, *Borosilicate glass 3.3 — Properties*

## 3 Types of beakers

The following three types of beakers are specified:

- a) low-form beaker with spout;
- b) low-form beaker with spout, thick-walled;
- c) tall-form beaker with spout.

## 4 Series of beakers

The series of beakers covered by this International Standard and defined by type and nominal capacity shall be as follows:

- a) low-form beakers: 5 ml – 10 ml – 25 ml – 50 ml – 100 ml – 250 ml – 400 ml – 500 ml – 600 ml – 800 ml – 1 000 ml – 2 000 ml – 3 000 ml – 5 000 ml – 10 000 ml;
- b) thick-walled low-form beakers: 150 ml – 250 ml – 400 ml – 600 ml – 1 000 ml – 2 000 ml – 5 000 ml;
- c) tall-form beakers: 50 ml – 100 ml – 150 ml – 250 ml – 400 ml – 500 ml – 600 ml – 800 ml – 1 000 ml – 2 000 ml – 3 000 ml – 5 000 ml.

## 5 Capacity

The design of the beaker shall provide for the difference in volume between nominal capacity and overall capacity. The capacity of the beaker shall be determined by either one of the following relationships:

- a) overflow of capacity of a beaker shall exceed the nominal capacity by at least 5 %;
- b) distance between the levels corresponding to the nominal and overflow capacities shall be at least 10 mm.

Whichever of these two relationships produces the greater differential in capacity shall apply.