

Wheat and wheat flour - Gluten content - Part 2:  
Determination of wet gluten and gluten index by  
mechanical means (ISO 21415-2:2015)

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 21415-2:2015 sisaldab Euroopa standardi EN ISO 21415-2:2015 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 21415-2:2015 consists of the English text of the European standard EN ISO 21415-2:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 28.10.2015.	Date of Availability of the European standard is 28.10.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 67.060

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

English Version

**Wheat and wheat flour - Gluten content - Part 2:  
Determination of wet gluten and gluten index by  
mechanical means (ISO 21415-2:2015)**

Blé et farines de blé - Teneur en gluten - Partie 2:  
Détermination du gluten humide et du gluten index par  
des moyens mécaniques (ISO 21415-2:2015)

Weizen und Weizenmehl - Glutengehalt - Teil 2:  
Bestimmung von Feuchtgluten und Glutenindex durch  
mechanische Verfahren (ISO 21415-2:2015)

This European Standard was approved by CEN on 13 June 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## European foreword

This document (EN ISO 21415-2:2015) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 338 "Cereal and cereal products" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2016, and conflicting national standards shall be withdrawn at the latest by April 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 21415-2:2008.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### Endorsement notice

The text of ISO 21415-2:2015 has been approved by CEN as EN ISO 21415-2:2015 without any modification.

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Terms and definitions</b>	<b>1</b>
<b>3 Principle</b>	<b>1</b>
<b>4 Reagents</b>	<b>1</b>
<b>5 Apparatus</b>	<b>2</b>
<b>6 Sampling</b>	<b>3</b>
<b>7 Preparation of test sample</b>	<b>3</b>
<b>8 Procedure</b>	<b>3</b>
8.1 General	3
8.2 Testing	3
8.3 Paste preparation	3
8.4 Paste washing	4
8.4.1 Detection of starch	4
8.4.2 Flour	4
8.4.3 Ground wheat	4
8.4.4 Special case	4
8.5 Spinning and weighing the wet gluten	4
8.6 Determining the gluten index	4
8.7 Number of measurements	5
<b>9 Calculations and record of results</b>	<b>5</b>
<b>10 Precision</b>	<b>5</b>
10.1 Interlaboratory tests	5
10.2 Repeatability	6
10.3 Reproducibility	6
10.4 Critical difference	6
10.4.1 Comparison of two sets of measurements in the same laboratory	6
10.4.2 Comparison of two sets of measurements in two laboratories	7
<b>11 Test report</b>	<b>7</b>
<b>Annex A (informative) Washing chamber and mill of the Glutomatic unit and centrifuge</b>	<b>8</b>
<b>Annex B (normative) Preparation of ground wheat</b>	<b>11</b>
<b>Annex C (informative) Results of interlaboratory tests</b>	<b>12</b>
<b>Bibliography</b>	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*.

This second edition of ISO 21415-2 cancels and replaces the first edition (ISO 21415-2:2006) which has been technically revised.

ISO 21415 consists of the following parts, under the general title *Wheat and wheat flour — Gluten content*:

- *Part 1: Determination of wet gluten by a manual method*
- *Part 2: Determination of wet gluten and gluten index by mechanical means*
- *Part 3: Determination of dry gluten from wet gluten by using an oven-drying method*
- *Part 4: Determination of dry gluten from wet gluten by a rapid drying method*

## Introduction

The alternative techniques specified in this part of ISO 21415 and in ISO 21415-1 for isolation of wet gluten (i.e. manual extraction and mechanical extraction) do not generally yield equivalent results. The reason for this is that for full development of the gluten structure the dough needs to be allowed to rest. Hence, the result obtained by manual extraction is usually greater than that obtained by mechanical extraction, especially in the case of wheat with high gluten content. Therefore, the test report should always state the technique used.

# Wheat and wheat flour — Gluten content —

## Part 2:

## Determination of wet gluten and gluten index by mechanical means

### 1 Scope

This part of ISO 21415 specifies a method for determining the content of wet gluten and the gluten index for wheat flours (*Triticum aestivum* L. and *Triticum durum* Desf.) by mechanical means. This method is directly applicable to flours. It also applies to common and durum wheat after grinding, if their particular size distribution meets the specification given in [Table B.1](#).

### 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

#### 2.1

##### **wet gluten**

viscoelastic substance consisting mainly of two protein fractions (gliadin and glutenin) in hydrated form, obtained in the way indicated in this part of ISO 21415 or in ISO 21415-1

#### 2.2

##### **gluten index**

proportion of wet gluten remaining on the sieve after centrifugation

Note 1 to entry: The higher the index, the stronger the gluten is.

#### 2.3

##### **ground wheat**

result of experimental grinding of whole wheat with the granulometry cited in [Table B.1](#)

#### 2.4

##### **flour**

finely ground wheat endosperm with a granulometry of less than 250 µm

### 3 Principle

Preparation of a paste from a sample of flour or of ground wheat and a sodium chloride solution in the equipment's chamber; separation of the wet gluten by washing this paste with a sodium chloride solution, followed by removal of excess washing solution by centrifugation and weighing the residue. The gluten index is obtained after centrifuging to force the wet gluten through a special sieve. The percentage of wet gluten remaining on the sieve after centrifuging is defined as the gluten index.

### 4 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified and distilled or demineralized water, or water of equivalent purity.