

**Stomatoloogia. Meditsiinivahendid stomatoloogias.
Hambaimplantaadid**

Dentistry - Medical devices for dentistry - Dental implants

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN 1642:2011 sisaldab Euroopa standardi EN 1642:2011 ingliskeelset teksti.</p> <p>Standard on kinnitatud Eesti Standardikeskuse 31.10.2011 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 19.10.2011.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN 1642:2011 consists of the English text of the European standard EN 1642:2011.</p> <p>This standard is ratified with the order of Estonian Centre for Standardisation dated 31.10.2011 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.</p> <p>Date of Availability of the European standard text 19.10.2011.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

ICS 11.060.15

Standardite reprodutseerimis- ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

English Version

Dentistry - Medical devices for dentistry - Dental implants

Médecine bucco-dentaire - Dispositifs médicaux pour la
médecine bucco-dentaire - Implants dentaires

Zahnheilkunde - Medizinprodukte für die Zahnheilkunde -
Dentalimplantate

This European Standard was approved by CEN on 20 September 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	6
4 Requirements	6
4.1 General.....	6
4.2 Design and properties	6
4.2.1 Materials	6
4.2.2 Contents of technical file	6
4.2.3 Biocompatibility	6
4.2.4 Biophysical properties and modelling	6
4.3 Control of contamination	7
4.3.1 General.....	7
4.3.2 Dental implants supplied sterile.....	7
4.3.3 Dental implants supplied non-sterile.....	7
4.3.4 Dental implants which incorporate materials of animal origin	7
4.4 Dental implants used in combination	7
4.5 Clinical investigation and evaluation.....	7
4.6 Marking, labelling and information supplied by the manufacturer	8
4.6.1 General.....	8
4.6.2 Symbols	8
4.6.3 Label.....	8
4.6.4 Instructions for use	8
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC	10
Bibliography	11

Foreword

This document (EN 1642:2011) has been prepared by Technical Committee CEN/TC 55 “Dentistry”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2012, and conflicting national standards shall be withdrawn at the latest by April 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1642:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive 93/42/EEC.

For relationship with EU Directive 93/42/EEC, see informative Annex ZA, which is an integral part of this document.

The following changes were made:

a) normative references:

- 1) addition of new relevant product standards, issued after 2004: EN 1641, EN ISO 11135-1, EN ISO 11137-1, EN ISO 11607-1, EN ISO 11607-2, EN ISO 14801, EN ISO 14971, EN ISO 22794, EN ISO 22803;
- 2) deletion of the following withdrawn standards: EN 550, EN 552, EN ISO 14727;

b) 4.5 Clinical evaluation: clarification of requirement for a clinical evaluation;

c) 4.6.4 Instructions for use: clarification of requirement that information may be provided in an electronic format;

d) Annex ZA: actualisation of the annex.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Introduction

There are three levels of European Standards dealing with medical devices used in dentistry. These are as follows:

- Level 1: General requirements for medical devices;
- Level 2: Requirements for families of medical devices used in dentistry;
- Level 3: Specific requirements for types of medical devices used in dentistry.

There are no level 1 standards written exclusively in respect of medical devices used in dentistry.

This European Standard is a level 2 standard and details requirements that apply to dental implants (for surgically implantable dental materials included within the definition of restorative materials see EN 1641). It is also indicated that there are additional requirements in the level 3 standards. Where available, these are included as normative references. To cover all the requirements for a particular product, it is necessary to use a standard of the lowest available level.

In the bibliography a reference for guidance on the classification of dental devices and accessories [4] is given.

1 Scope

This European Standard specifies general requirements for dental implants and accessories. Surgically implantable dental materials defined as restorative materials are specifically excluded and described in EN 1641.

This European Standard includes requirements for intended performance, design attributes, components, sterilization, packaging, marking, labelling, and information supplied by the manufacturer.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 556-1, *Sterilization of medical devices — Requirements for medical devices to be designated "STERILE" — Part 1: Requirements for terminally sterilized medical devices*

EN 980, *Symbols for use in the labelling of medical devices*

EN 1041, *Information supplied by the manufacturer of medical devices*

EN 21942-1:1991, *Dental vocabulary — Part 1: General and clinical terms (ISO 1942-1:1989)*

EN 21942-2:1992, *Dental vocabulary — Part 2: Dental materials (ISO 1942-2:1989)*

EN ISO 1942-5:1994, *Dental vocabulary — Part 5: Terms associated with testing (ISO 1942-5:1989)*

EN ISO 7405, *Dentistry — Evaluation of biocompatibility of medical devices used in dentistry (ISO 7405:2008)*

EN ISO 10451, *Dental implant systems — Contents of technical file (ISO 10451:2002)*

EN ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing (ISO 10993-1:2003)*

EN ISO 11135-1, *Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices (ISO 11135-1:2007)*

EN ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices (ISO 11137-1:2006)*

EN ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2006)*

EN ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes (ISO 11607-2:2006)*

EN ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements (ISO 14155-1:2003)*

EN ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans (ISO 14155-2:2003)*

EN ISO 14801, *Dentistry — Implants — Dynamic fatigue test for endosseous dental implants (ISO 14801:2007)*

EN ISO 14971, *Medical devices — Application of risk management to medical devices (ISO 14971:2007)*

EN ISO 22794, *Dentistry — Implantable materials for bone filling and augmentation in oral and maxillofacial surgery — Contents of technical file (ISO 22794:2007)*

EN ISO 22803, *Dentistry — Membrane materials for guided tissue regeneration in oral and maxillofacial surgery — Contents of technical file (ISO 22803:2004)*

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 21942-1:1991, EN 21942-2:1992, EN ISO 1942-5:1994 and the following apply.

3.1 dental implants

device designed to be placed surgically within or on the mandibular or maxillary bone to provide resistance to displacement of a dental prosthesis or to provide orthodontic anchorage

NOTE The term dental implant includes transendodontic implants.

4 Requirements

4.1 General

4.1.1 Dental implants shall comply with the requirements which are applicable to them bearing in mind the intended purpose of the implant concerned. Conformity with these requirements shall be considered to be met by demonstrating compliance with the requirements of the following subclauses, if appropriate.

4.1.2 Risk management shall be carried out and documented. This shall include a risk analysis in accordance with EN ISO 14971.

4.2 Design and properties

4.2.1 Materials

Dental implants shall be manufactured from materials selected with regard to the properties required for the intended purpose.

4.2.2 Contents of technical file

The contents of the manufacturer's technical file shall be in accordance with EN ISO 10451, EN ISO 22794 and EN ISO 22803, if applicable.

4.2.3 Biocompatibility

Dental implants shall be assessed for biocompatibility. Guidance on the selection of tests is given in EN ISO 7405 and EN ISO 10993-1. EN ISO 7405 includes usage tests specific to dental materials.

Special attention shall be given to substances which are carcinogenic, mutagenic or toxic to reproduction.

NOTE Further information is given in Annex I of Council Directive 67/548/EEC of 27 June 1967.

4.2.4 Biophysical properties and modelling

Dental implants, other than those designed to provide orthodontic anchorage, shall be assessed for mechanical stability by fatigue testing according to EN ISO 14801.

NOTE The stability of dental implants designed to provide orthodontic anchorage shall be assessed by mechanical tests with regard to the properties required for the intended purpose.