
**Non-active surgical implants — Joint
replacement implants — Specific
requirements for hip-joint replacement
implants**

*Implants chirurgicaux non actifs — Implants de remplacement
d'articulation — Exigences spécifiques relatives aux implants de
remplacement de l'articulation de la hanche*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Intended performance	2
5 Design attributes.....	3
5.1 General.....	3
5.2 Tolerances and dimensions	3
5.3 Thickness of UHMWPE in acetabular components and bipolar heads.....	3
6 Materials	3
7 Design evaluation	4
7.1 General.....	4
7.2 Preclinical evaluation	4
8 Manufacture.....	4
9 Sterilization.....	5
10 Packaging	5
11 Information to be supplied by the manufacturer.....	5
11.1 General.....	5
11.2 Dimensions.....	5
11.3 Structural and functional compatibility of components	5
11.4 Marking	5
11.5 Information for the patient	6
Annex A (informative) Evaluation of relative angular motion of components.....	7
Bibliography	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21535 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 4, *Bone and joint replacements*.

This second edition cancels and replaces the first edition (ISO 21535:2002), which has been technically revised.

Introduction

There are three levels of International Standard dealing with non-active surgical implants. These are as follows, with level 1 being the highest:

- level 1: general requirements for non-active surgical implants and instrumentation used in association with implants;
- level 2: particular requirements for families of non-active surgical implants;
- level 3: specific requirements for types of non-active surgical implant.

This International Standard is a level 3 standard and contains requirements applying specifically to hip joint replacements.

The level 1 International Standard, ISO 14630, contains requirements that apply to all non-active surgical implants. It also indicates that there are additional requirements in the level 2 and level 3 standards. The level 2 standards apply to more restricted sets or families of implants such as those designed for use in osteosynthesis, cardiovascular surgery or joint replacement.

To address all requirements, it is recommended that a standard of the lowest available level be consulted first.

This document is a preview generated by EVS

Non-active surgical implants — Joint replacement implants — Specific requirements for hip-joint replacement implants

1 Scope

This International Standard provides specific requirements for hip joint replacement implants. With regard to safety, this International Standard specifies requirements for intended performance, design attributes, materials, design evaluation, manufacture, sterilization, packaging, information supplied by the manufacturer, and methods of test.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7206-1, *Implants for surgery — Partial and total hip joint prostheses — Part 1: Classification and designation of dimensions*

ISO 7206-2:1996, *Implants for surgery — Partial and total hip joint prostheses — Part 2: Articulating surfaces made of metallic, ceramic and plastics materials*

ISO 7206-4, *Implants for surgery — Partial and total hip joint prostheses — Part 4: Determination of endurance properties of stemmed femoral components*

ISO 7206-6:1992, *Implants for surgery — Partial and total hip joint prostheses — Part 6: Determination of endurance properties of head and neck region of stemmed femoral components*

ISO 7206-10, *Implants for surgery — Partial and total hip joint prostheses — Part 10: Determination of resistance to static load of modular femoral heads*

ISO 14630:—¹⁾, *Non-active surgical implants — General requirements*

ISO 21534:2007, *Non-active surgical implants — Joint replacement implants — Particular requirements*

3 Terms and definitions

For the purposes of this document the terms and definitions in ISO 21534 and ISO 7206-1 together with the following apply.

3.1

acetabular component

implant intended to be fixed to the prepared biological acetabulum

NOTE The component can be of monobloc or modular construction.

1) To be published. (Revision of ISO 14630:2005)