TECHNICAL REPORT RAPPORT TECHNIQUE **TECHNISCHER BERICHT**

CEN ISO/TR 13115

December 2011

ICS 19.100

English Version

Non-destructive testing - Methods for absolute calibration of acoustic emission transducers by the reciprocity technique (ISO/TR 13115:2011)

Essais non destructifs - Méthodes d'étalonnage absolu des capteurs d'émission acoustique par la technique de réciprocité (ISO/TR 13115:2011)

Zerstörungsfreie Prüfung - Methode zur Absolutkalibrierung von Schallemissionswandlern durch Reziproktechnik (ISO/TR 13115:2011)

This Technical Report was approved by CEN on 20 September 2011. It has been drawn up by the Technical Committee CEN/TC 138.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (CEN ISO/TR 13115:2011) has been prepared by Technical Committee ISO/TC 135 "Nondestructive testing" in collaboration with Technical Committee CEN/TC 138 "Non-destructive testing" the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of ISO/TR 13115:2011 has been approved by CEN as a CEN ISO/TR 13115:2011 without any modification.

Contents

Forev	word	iv
Intro	duction	v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4 4.1 4.2 4.3	Preparation for calibration Transfer medium and calibration signal Mounting of acoustic emission transducer Calculation of reciprocity parameters	3 3 4 5
5 5.1 5.2 5.3	Method for three-transducer calibration Apparatus to be used Method of measurement Method for determination of absolute sensitivity	6 6
6 6.1 6.2 6.3	Method for two-transducer calibration Apparatus to be used Method of measurement Method for determination of absolute sensitivity	9 10 10 10
7 7.1 7.2 7.3 7.4	Method for impulse response calibration Apparatus to be used Method of measurement Method for determination of frequency response Method for determination of impulse response	11 11 11 13 15
8 8.1 8.2 8.3	Method for representing calibration results Representation items for calibration results Method for representing frequency response of absolute sensitivity Method for representing impulse response of absolute sensitivity	16
Biblic	ography	18

Introduction

A standard method for primary calibration of acoustic emission transducers, ISO 12713:1998^[1], introduced the seismic surface pulse method for Rayleigh surface wave calibration, wherein the breaking of a glass capillary is employed for the sound source and a standard capacitive transducer is used for the measurement of dynamic displacements of the surface. In ISO 12714:1999^[2], on secondary calibration of acoustic emission sensors, a transducer which has been calibrated by the seismic surface pulse method is employed for comparison of reception sensitivity.

This Technical Report describes the methods for calibrating absolute sensitivity of acoustic emission transducers, both to Rayleigh surface waves and longitudinal waves, by means of a reciprocity technique. Since reciprocity parameters have been derived, absolute sensitivity can be determined by purely electrical measurements without the use of mechanical sound sources or reference transducers.

Procedures of the seismic surface pulse method and reciprocity technique differ from each other; however, there is a common theoretical basis in the two calibration methods. For the seismic surface pulse method, theoretical surface displacements were calculated on the basis of Lamb's theory (Reference [7]). For the reciprocity calibration, reciprocity parameters for the Rayleigh wave calibration were also derived from Lamb's theory. As for the Rayleigh surface wave calibration, a round robin experiment was carried out in a collaborative effort between the USA and Japan, and it was ascertained that absolute sensitivities as obtained by either method agreed well.

The aim of both methods is the same, namely, to establish uniformity of acoustic emission testing, to form a basis for data correlation, and to provide for the interpretation of results obtained by different laboratories at different times.

This Technical Report describes methods for three-transducer calibration, two-transducer calibration, and impulse response calibration, respectively. In three-transducer calibration, three acoustic emission transducers of the same kind, which are reversible transducers, are prepared to configure three independent pairs of transmitting and receiving transducers on a solid transfer medium. Transmission signal current and reception signal voltage are measured on each pair as a function of frequency, and frequency responses of amplitude of absolute sensitivity both to the Rayleigh surface waves and longitudinal waves are determined on each transducer. Once three-transducer calibration has been carried out, an optional transducer, which is not necessarily a reversible transducer, can be calibrated by a relatively simple procedure by using the calibrated transducer as a reference of transmission or reception. In two-transducer calibration, frequency responses of amplitude of absolute reception sensitivity are determined on an optional transducer by using one acoustic emission transducer, the transmission responses of which have been calibrated by the three-transducer calibration. In addition, by means of three-transducer calibration, impulse responses of each acoustic emission transducer can also be determined. In the impulse response calibration, frequency responses of phase angle, in addition to amplitude, of absolute sensitivity are measured by three-transducer calibration on the basis of complex reciprocity parameters, and impulse responses are determined through inverse Fourier transform of the frequency responses of amplitude and phase. 2

Non-destructive testing — Methods for absolute calibration of acoustic emission transducers by the reciprocity technique

1 Scope

This Technical Report describes the method of three-transducer calibration for calibrating frequency responses of absolute sensitivity by means of a reciprocity technique using three reversible acoustic emission transducers of the same kind, the method of two-transducer calibration for calibrating frequency responses of reception sensitivity of an optional acoustic emission transducer by using one acoustic emission transducer, the transmission responses of which have been calibrated by three-transducer calibration, the method for impulse response calibration for calibrating impulse responses of absolute sensitivity through inverse Fourier transform of the frequency responses measured by the three-transducer calibration, and the method for representing the calibration results.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 12716:2001, Non-destructive testing — Acoustic emission inspection — Vocabulary

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 12716 and the following apply.

3.1

reciprocity technique

calibration method on three reversible acoustic emission transducers of the same kind, wherein transducers are arranged on a solid transfer medium so that they configure three independent pairs of transmitting and receiving transducers, and absolute sensitivity is determined only by electrical measurements of transmission current and reception voltage on each pair

3.2

reversible transducer

transducer which can be used both for transmission and reception

3.3

absolute sensitivity

quantity of reception voltage sensitivity or transmission current response of an acoustic emission transducer

3.4

reception voltage sensitivity

ratio of the open-circuit output voltage of an acoustic emission transducer used for reception to the vertical component of displacement velocity at the position where the transducer is to be placed