

ICS 17.040.20

English Version

**Geometrical product specifications (GPS) - Cylindricity - Part 2:
Specification operators (ISO/TS 12180-2:2003)**

Spécification géométrique des produits (GPS) - Cylindricité
- Partie 2: Opérateurs de spécification (ISO/TS 12180-
2:2003)

Geometrische Produktspezifikation (GPS) - Zylindrizität -
Teil 2: Spezifikationsoperatoren (ISO/TS 12180-2:2003)

This Technical Specification (CEN/TS) was approved by CEN on 8 October 2007 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

The text of ISO/TS 12180-2:2003 has been prepared by Technical Committee ISO/TC 213 "Dimensional and geometrical product specifications and verification" of the International Organization for Standardization (ISO) and has been taken over as CEN ISO/TS 12180-2:2007 by Technical Committee CEN/TC 290 "Dimensional and geometrical product specification and verification" the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO/TS 12180-2:2003 has been approved by CEN as a CEN ISO/TS 12180-2:2007 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Complete specification operator	1
5 Compliance with specification	2
Annex A (informative) The harmonic content of a nominally cylindrical workpiece and extraction strategy	3
Annex B (informative) Extraction strategies	7
Annex C (informative) Relation to the GPS matrix model	10
Bibliography	12

Introduction

This part of ISO/TS 12180 is a geometrical product specification (GPS) Technical Specification and is to be regarded as a general GPS document (see ISO/TR 14638). It influences chain link 3 of the chain of standards on form of a surface (independent of a datum).

For more detailed information on the relation of this part of ISO/TS 12180 to other standards and the GPS matrix model, see Annex C.

This part of ISO/TS 12180 specifies the specification operators according to ISO/TS 17450-2 for cylindricity of integral features.

At the current state of development, ISO/TC 213 has not been able to reach a consensus on defaults for filter UPR, probe tip radius and method of association (reference cylinder). This means that a cylindricity specification should explicitly state which values are to be used for these specification operations in order for it to be unique.

Consequently, if a specification does not explicitly state which values are to be used for one or more of these operators, the specification is uncertain (see ISO/TS 17450-2) and a supplier can use any value for the operator(s) not specified when proving conformance.

Extracting data will always involve applying a certain filtering process. An additional filtering of the extracted data may or may not be applied. This additional filter can be a mean line filter (Gaussian, spline, wavelet, etc.) or a non-linear filter (e.g. morphological filter). The type of filtering will influence the definition of cylindricity and the specification operators and, therefore, needs to be stated unambiguously.

NOTE 1 Stylus filtering is not sufficient on its own to smooth a profile. In certain circumstances it can create spurious high-frequency content, thus giving incorrect values. To correct this, a longwave pass filter is employed. A Gaussian filter is used, since this is the current state-of-the-art in ISO standards. This filter has some shortcomings, e.g. it can distort, rather than eliminate some roughness features and it can distort, rather than transmit correctly some waviness features. It is envisioned that new filters under development within ISO will provide better solutions for several of these issues.

NOTE 2 If a smaller tip radius than the one specified is used for a given cut-off length, the resulting measured value will generally be higher. This effect is usually insignificant. If a larger tip radius is used, the resulting measured value will generally be lower. The amount of change is heavily dependent on the surface measured.

NOTE 3 The measuring force of zero N is chosen to eliminate effects of elastic deformation of the workpiece from the specification operator. On metal surfaces with adequate thickness, the effect of normally occurring measuring forces will be negligible.

NOTE 4 Aliasing and other problems during extraction (see Annex A) due to the higher harmonic content of the skin model, in the roundness and straightness directions, can cause specification uncertainty.

This part of ISO/TS 12180 is not intended to disallow any means of measuring cylindricity.

Geometrical Product Specifications (GPS) — Cylindricity —

Part 2: Specification operators

1 Scope

This part of ISO/TS 12180 specifies the complete specification operator for cylindricity of complete integral features only, i.e. geometrical characteristics of features of type cylinder.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/TS 12180-1:2003, *Geometrical Product Specifications (GPS) — Cylindricity — Part 1: Vocabulary*

ISO 14253-1:1998, *Geometrical Product Specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or non-conformance with specifications*

ISO/TS 17450-2:2002, *Geometrical Product Specifications (GPS) — General concepts — Part 2: Basic tenets, specifications, operators and uncertainties*

3 Terms and definitions

For the purposes of this part of ISO/TS 12180, the terms and definitions given in ISO/TS 12180-1 and ISO/TS 17450-2 apply.

4 Complete specification operator

4.1 General

The complete specification operator (see ISO/TS 17450-2) is a full ordered set of unambiguous specification operations in a well-defined order. The complete specification operator defines the transmission band for the cylindricity surface, together with an appropriate stylus tip geometry.

NOTE In practice it is unrealistic to hope to achieve comprehensive coverage of the cylindrical feature given by the theoretical minimum density of points (see Annex B) within an acceptable time span using current technology. Therefore more limited extraction strategies are employed that give specific rather than general information concerning the deviations from cylindrical form.