
***In vitro* diagnostic medical devices —
Information supplied by the manufacturer
(labelling) —**

Part 4:
***In vitro* diagnostic reagents for self-
testing**

*Dispositifs médicaux de diagnostic in vitro — Informations fournies par
le fabricant (étiquetage) —*

Partie 4: Réactifs de diagnostic in vitro pour auto-tests



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General	2
4.1 Essential requirements	2
4.2 Identification of kit components	2
4.3 Presentation of the instructions for use	2
5 Content of the outer container label	2
5.1 Manufacturer	2
5.2 Identification of the IVD reagent	2
5.3 Contents	3
5.4 Intended use	3
5.5 <i>In vitro</i> diagnostic use	3
5.6 Storage and handling conditions	3
5.7 Expiry date	3
5.8 Warnings and precautions	4
6 Content of the immediate container label	4
6.1 General provisions	4
6.2 Manufacturer	4
6.3 Identification of the IVD reagent	4
6.4 Contents	4
6.5 <i>In vitro</i> diagnostic use	4
6.6 Storage and handling conditions	5
6.7 Expiry date	5
6.8 Warnings and precautions	5
7 Content of the instructions for use	5
7.1 Manufacturer	5
7.2 Identification of the IVD reagent	5
7.3 Intended use	5
7.4 Principles of the examination method	6
7.5 Components	6
7.6 Additional required equipment	6
7.7 Reagent preparation	6
7.8 Storage and shelf life after first opening	6
7.9 Warnings and precautions	6
7.10 Primary sample collection, handling and storage	7
7.11 Examination procedure	7
7.12 Control procedure	7
7.13 Reading of examination results	7
7.14 Interpretation of results	7
7.15 Performance characteristics	7
7.16 Biological reference intervals	8
7.17 Limitations of examination procedure	8
7.18 Literature references	8
Bibliography	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18113-4 was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*

ISO 18113 consists of the following parts, under the general title *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling)*:

- *Part 1: Terms, definitions and general requirements*
- *Part 2: In vitro diagnostic reagents for professional use*
- *Part 3: In vitro diagnostic instruments for professional use*
- *Part 4: In vitro diagnostic reagents for self-testing*
- *Part 5: In vitro diagnostic instruments for self-testing*

Introduction

Manufacturers of *in vitro* diagnostic (IVD) reagents for self-testing supply users with information to enable the safe use and expected performance of their devices. The type and level of detail varies according to the intended uses and country-specific regulations.

The Global Harmonization Task Force (GHTF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions could allow patients earlier access to new technologies and treatments. See Reference [9]. This part of ISO 18113 provides a basis for harmonization of labelling requirements for IVD reagents for self-testing.

This part of ISO 18113 is concerned solely with information supplied with IVD reagents, calibrators and control materials intended for self-testing. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of general labelling concepts.

This part of ISO 18113 is based on EN 376:2002^[5]. The text has been modified to conform to Part 2 of the ISO/IEC Directives^[4], but the requirements, including those in ISO 18113-1, are substantially equivalent to the original European harmonized standard. This part of ISO 18113 is intended to support the essential labelling requirements of all the GHTF partners, as well as other countries that have enacted or plan to enact labelling regulations for IVD medical devices.

For IVD reagents, calibrators and/or control materials that are intended to be used as a system with an instrument provided by the same manufacturer, this part of ISO 18113 is also intended to be used together with ISO 18113-1 and ISO 18113-5^[3].

This document is a preview generated by EVS

***In vitro* diagnostic medical devices — Information supplied by the manufacturer (labelling) —**

Part 4:

***In vitro* diagnostic reagents for self-testing**

1 Scope

This part of ISO 18113 specifies requirements for information supplied by the manufacturer of IVD reagents for self-testing.

This part of ISO 18113 also applies to information supplied by the manufacturer with calibrators and control materials intended for use with IVD medical devices for self-testing.

This part of ISO 18113 can also be applied to accessories.

This part of ISO 18113 applies to the labels for outer and immediate containers and to the instructions for use.

This part of ISO 18113 does not apply to:

- a) IVD instruments or equipment,
- b) IVD reagents for professional use.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 18113-1, *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements*

EN 980, *Symbols for use in the labelling of medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18113-1 apply.