
**Medical devices utilizing animal
tissues and their derivatives —**

**Part 1:
Application of risk management**

Dispositifs médicaux utilisant des tissus animaux et leurs dérivés —

Partie 1: Application de la gestion des risques



This document is a preview generated by EBS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Risk management process	3
4.1 General.....	3
4.2 Risk analysis.....	3
4.2.1 Identification of qualitative and quantitative characteristics related to the safety of medical devices.....	3
4.2.2 Identification of hazards and hazardous situations.....	4
4.3 Risk evaluation.....	5
4.4 Risk control.....	5
4.4.1 General.....	5
4.4.2 Risk control for viruses and TSE agents.....	5
4.4.3 Risk control of other hazards.....	5
4.4.4 Residual risk evaluation.....	6
4.5 Evaluation of overall residual risk acceptability.....	6
4.5.1 General.....	6
4.5.2 Documentation.....	6
4.6 Production and post-production information system.....	7
Annex A (informative) Guidance on the application of this part of ISO 22442	8
Annex B (informative) Graphical representation of part of the risk management process for medical devices utilizing animal material	9
Annex C (normative) Special requirements for some animal materials considering the risk management for TSE agents	11
Annex D (informative) Information relevant to the management of TSE risk	16
Bibliography	25

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 194, *Biological and clinical evaluation of medical devices*, SC 1, *Tissue product safety*.

This second edition cancels and replaces the first edition (ISO 22442-1:2007), of which it constitutes a minor revision.

ISO 22442 consists of the following parts, under the general title *Medical devices utilizing animal tissues and their derivatives*:

- *Part 1: Application of risk management*
- *Part 2: Controls on sourcing, collection and handling*
- *Part 3: Validation of the elimination and/or inactivation of viruses and transmissible spongiform encephalopathy (TSE) agents*
- *Part 4: Principles for elimination and/or inactivation of transmissible spongiform encephalopathy (TSE) agents and validation assays for those processes [Technical Report]*

Introduction

Certain medical devices utilize materials of animal origin.

Animal tissues and their derivatives are used in the design and manufacture of medical devices to provide performance characteristics that have been chosen for advantages over non-animal based materials. The range and quantities of materials of animal origin in medical devices vary. These materials can comprise a major part of the device (e.g. bovine/porcine heart valves, bone substitutes for use in dental or orthopaedic applications, haemostatic devices), can be a product coating or impregnation (e.g. collagen, gelatine, heparin), or can be used in the device manufacturing process (e.g. tallow derivatives such as oleates and stearates, foetal calf serum, enzymes, culture media).

ISO 14971 is a general standard which specifies a process for a manufacturer by identifying hazards and hazardous situations associated with medical devices, including *in vitro* medical devices, to estimate and evaluate the risks associated with those hazards, to control these risks and to monitor the effectiveness of the control throughout the life cycle. This part of ISO 22442 provides additional requirements and guidance for the evaluation of medical devices manufactured utilizing animal tissues or derivatives which are non-viable or rendered non-viable.

This part of ISO 22442 is intended to cover medical devices including active implantable medical devices such as implantable infusion pumps.

This part of ISO 22442 does not apply to *in vitro* diagnostic devices.

This part of ISO 22442 can only be used in combination with ISO 14971 and is not a “standalone” standard.

To show compliance with this part of ISO 22442, its specified requirements should be fulfilled. The guidance given in the Notes and informative annexes is not normative and is not provided as a checklist for auditors.

Medical devices utilizing animal tissues and their derivatives —

Part 1: Application of risk management

1 Scope

This part of ISO 22442 applies to medical devices other than *in vitro* diagnostic medical devices manufactured utilizing materials of animal origin, which are non-viable or have been rendered non-viable. It specifies, in conjunction with ISO 14971, a procedure to identify the hazards and hazardous situations associated with such devices, to estimate and evaluate the resulting risks, to control these risks, and to monitor the effectiveness of that control. Furthermore, it outlines the decision process for the residual risk acceptability, taking into account the balance of residual risk, as defined in ISO 14971, and expected medical benefit as compared to available alternatives. This part of ISO 22442 is intended to provide requirements and guidance on risk management related to the hazards typical of medical devices manufactured utilizing animal tissues or derivatives such as

- a) contamination by bacteria, moulds or yeasts;
- b) contamination by viruses;
- c) contamination by agents causing Transmissible Spongiform Encephalopathies (TSE);
- d) material responsible for undesired pyrogenic, immunological or toxicological reactions.

For parasites and other unclassified pathogenic entities, similar principles can apply.

This part of ISO 22442 does not stipulate levels of acceptability which, because they are determined by a multiplicity of factors, cannot be set down in such an International Standard except for some particular derivatives mentioned in [Annex C](#). [Annex C](#) stipulates levels of TSE risk acceptability for tallow derivatives, animal charcoal, milk and milk derivatives, wool derivatives and amino acids.

This part of ISO 22442 does not specify a quality management system for the control of all stages of production of medical devices.

This part of ISO 22442 does not cover the utilization of human tissues in medical devices.

NOTE 1 It is not a requirement of this part of ISO 22442 to have a full quality management system during manufacture. However, attention is drawn to International Standards for quality management systems (see ISO 13485) that control all stages of production or reprocessing of medical devices.

NOTE 2 For guidance on the application of this part of ISO 22442, see [Annex A](#).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14971, *Medical devices — Application of risk management to medical devices*