

ICS 65.100.01; 67.050

English Version

**Food analysis - Determination of pesticide residues by GC-MS -
Retention times, mass spectrometric parameters and detector
response information**

Analyse des produits alimentaires - Détermination des
résidus de pesticides par GC-SM - Temps de rétention,
paramètres de spectrométrie de masse et information sur
la réponse des détecteurs

Lebensmitteluntersuchung - Bestimmung von
Pestizidrückständen mit GC-MS - Retentionszeiten,
Parameter für die Massenspektrometrie und
Detektionsempfindlichkeit

This Technical Report was approved by CEN on 3 December 2012. It has been drawn up by the Technical Committee CEN/TC 275.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

| | |
|---|----|
| Foreword..... | 3 |
| Introduction | 4 |
| 1 Scope..... | 5 |
| 2 Analyte specific parameters for gas chromatographic determination of pesticides | 5 |
| 2.1 General | 5 |
| 2.2 GC-MS, GC-ECD and GC-NPD Parameters | 5 |
| 2.3 Analyte specific retention times and MS parameters | 6 |
| Bibliography..... | 24 |

Foreword

This document (CEN/TR 16468:2013) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document is a preview generated by EVS

Introduction

Pesticide residue analysis employs multiple methods involving extraction of residues from foods and clean-up of the extract to obtain as many analytes as possible in the purified extracts. Afterwards the extracts can be analysed by different kinds of instruments.

The hyphenation of gas chromatography (GC) and mass spectrometry (MS) is one of the most often used universal and selective analysis techniques for identification and quantification of pesticide residues in food extracts.

For the ionisation of the analytes (pesticides and/or their metabolites) in GC-MS, electron impact ionisation (EI) is most commonly used. If the typical electron energy of 70 eV is used, very often molecular ions (cation radicals) and several fragment ions are formed simultaneously.

The selective determination of each target analyte is performed by simultaneous acquisition of typically three ions formed by the analyte in the selected ion monitoring (SIM) mode. The diagnostic value of an ion depends on its mass. Usually even-numbered, high mass ions are less frequently formed and their recording results in more specific chromatograms. A reduction of selectivity may be caused by background ions formed during 'column bleeding' or by sample matrix (e.g. typical ions from fatty or fruit acids or hydrocarbons).

1 Scope

This Technical Report lists mass spectrometric parameters which are useful for the application of European Standards for the determination of pesticide residues in foods that use GC-MS, such as the following standards:

- EN 1528 (all parts), *Fatty food — Determination of pesticides and polychlorinated biphenyls (PCBs)*;
- EN 12393 (all parts), *Foods of plant origin — Multiresidue methods for the gas chromatographic determination of pesticide residues*;
- EN 15662, *Foods of plant origin — Determination of pesticide residues using GC-MS and/or LC-MS/MS following acetonitrile extraction/partitioning and clean-up by dispersive SPE — QuEChERS-method*.

To facilitate the determination of pesticides and/or metabolites using GC-MS, Table 1 specifies the diagnostic ions suitable for quantification, which can be used.

2 Analyte specific parameters for gas chromatographic determination of pesticides

2.1 General

All values indicated in Table 1 were acquired using GC-MS, GC-ECD or GC-NPD systems under the experimental conditions as outlined in 2.2. Comparative investigations showed that these parameters can be transferred simply on instruments of other types of the same or other manufacturers [1].

2.2 GC-MS, GC-ECD and GC-NPD Parameters

The following GC operating conditions have been proven to be satisfactory. This is an example for appropriate experimental conditions. Equivalent conditions may be used if they can be shown to lead to the same results.

| | |
|-----------------------------|--|
| Gas chromatograph | Agilent model 6890 |
| Carrier gas | Helium, constant flow 1,0 ml/min (typical pressure at 70 °C is 61 kPa) |
| Injection technique | Pulsed splitless, 200 kPa; Pulse time 1,0 min |
| Injector temperature | 240 °C |
| Injection liner | single taper, splitless (Agilent 5181-3316) |
| Injection volume | 1 µl or 2 µl |
| Purge Gas | Helium; purge flow to split vent 50 ml/min; Purge time 1,5 min |
| Column | Fused silica capillary column HP-5MS, length 30 m, inner diameter 0,25 mm, film thickness 0,25 µm (Agilent Nr. 19091S-433) |
| Temperature programme | 2 min 70 °C, programmed to rise at 25 °C/min to 170 °C, at 3 °C/min to 210 °C, at 30 °C/min to 290 °C, then isothermal for 9 min |
| Transfer Line Temperatur | 280 °C |
| Mass spectrometer | Agilent MSD 5973N inert |
| MS temperatures | Quadrupole 150 °C; Source 230 °C; |
| Ionisation | Electron impact 70 eV |
| Solvent Delay | 4,0 min |
| Retention time of parathion | 15,85 min (parathion is used for calculation of relative retention times) |