

English Version

**Ambient air - Sampling and analysis of airborne pollen
grains and fungal spores for allergy networks - Volumetric
Hirst method**

Air ambiant - Échantillonnage et analyse des grains de
pollen et des spores fongiques aériens pour les réseaux
aérobiologiques - Méthode volumétrique de Hirst

Außenluft - Probenahme und Analyse luftgetragener
Pollen und Pilzsporen für Allergienetzwerke -
Volumetrische Hirst-Methode

This Technical Specification (CEN/TS) was approved by CEN on 15 September 2015 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions	5
4 Principle	8
5 Sampling.....	8
5.1 Equipment	8
5.1.1 Apparatus.....	8
5.1.2 Scrolling speed of support	11
5.1.3 Impaction support	11
5.1.4 Wind vane and rain shield	13
5.1.5 Complete sampling system	14
5.2 Operating procedure	14
5.2.1 Preparation of the coating medium [12]	14
5.2.2 Support preparation	15
5.2.3 Changing of the drum.....	16
6 Analysis.....	16
6.1 Equipment required.....	16
6.2 Operating procedure	17
6.2.1 Support	17
6.2.2 Mounting medium.....	17
6.3 Methodology for counting.....	18
6.3.1 Glass slide preparation for microscopy analysis for drum tape [13] [14]	18
6.3.2 Optical microscopy [14] [16]	20
6.3.3 Identification [15] [16].....	20
6.3.4 Counting method	20
6.3.5 Data recording.....	20
6.3.6 Conversion factor	21
7 Performance characteristics [10] [11] [17] [18] [19]	22
7.1 General.....	22
7.2 Performance requirements.....	22
7.2.1 Repeatability.....	22
7.3 Performance recommendations [18] [19]	22
7.3.1 Reproducibility and accuracy	22
7.3.2 Sensitivity and specificity.....	22
Annex A (informative) First type volumetric trap.....	24
Annex B (informative) Pictures of impaction support.....	25
Annex C (informative) Safety data sheet.....	27
Annex D (informative) Key of determination	28
Bibliography	32

European foreword

This document (CEN/TS 16868:2015) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

Biological particles (pollen and fungal spores) are present in the air, causing health impacts at various levels. In Europe, nearly 18 % to 20 % of people suffer from pollinosis due to pollen and/or fungal spores. Pollen grains and fungal spores are considered in some Member States as an air pollutant as well as particles suspended in the air (PM_{10,2,5}). In Europe, European Aerobiology Society (EAS) in coordination with International Association for Aerobiology (IAA) manage problems of sampling, analysis, quality control, development and information.

For the sampling and analysis of biological particles different methodology and operating procedures are used.

Sampling equipment is diversified (see Annex A). Analysis is based on optical light microscopy for identification and counting pollen grains and fungal spores.

Elements and reagents used during sampling and analysis have very specific properties and require to be handled carefully.

Given the close relationship between aerobiology and other sciences, one of the main aims is that information on airborne biological-particle counts should be of use in a wide range of disciplines and fields of application, including aerobiology, biodiversity, agriculture, forestry, phytopathology, meteorology, climatology, forensic science, bioterrorism, and health (sensitization and allergy).

1 Scope

This European Standard specifies the procedure to sample continuously and analyse the concentration of airborne pollen grains and fungal spores in ambient air using the volumetric Hirst type sampler [1] [2] [3] (see Annex A).

This European Standard describes both the sampling and the analysis procedures for the purpose of allergy networks. For the other tasks mentioned in the introduction, other specifications may be required.

2 Normative references

Not applicable.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

accuracy

closeness of agreement between a measured quantity value and a true quantity value of a measurement

3.2

bench

long work table in a workshop or laboratory

3.3

clockwork

mechanism with a spring and toothed gearwheels, used to drive a mechanical clock, toy, or other device

3.4

combined standard measurement uncertainty

obtained using the individual standard measurement uncertainties associated with the input quantities in a measurement model

3.5

defatted

surface conditions after clearing with a fat removing substance

3.6

drum

cylindrical device for the mounting of a sticky tape

3.7

exine

outer layer of the wall of a spore or pollen grain, also called an exosporium

3.8

eyepiece

lens or combination of lenses in an optical instrument through which the eye views the image formed by the objective lens or lenses; ocular