

MEDITSIINISEADMETE BIOLOOGILINE HINDAMINE. OSA
11: KATSED SÜSTEEMSE TOKSILISUSE HINDAMISEKS

Biological evaluation of medical devices - Part 11: Tests
for systemic toxicity (ISO 10993-11:2017)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 10993-11:2018 sisaldab Euroopa standardi EN ISO 10993-11:2018 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 10993-11:2018 consists of the English text of the European standard EN ISO 10993-11:2018.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 30.05.2018.	Date of Availability of the European standard is 30.05.2018.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.100.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Biological evaluation of medical devices - Part 11: Tests for
systemic toxicity (ISO 10993-11:2017)

Évaluation biologique des dispositifs médicaux - Partie
11: Essais de toxicité systémique (ISO 10993-11:2017)

Biologische Beurteilung von Medizinprodukten - Teil
11: Prüfungen auf systemische Toxizität (ISO 10993-
11:2017)

This European Standard was approved by CEN on 31 July 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 10993-11:2018) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2018, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10993-11:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA and ZB, which is an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard 'within the meaning of Annex ZA', the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table 1 — Correlations between undated normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 10993-1	EN ISO 10993-1:2009	ISO 10993-1:2009
ISO 10993-2	EN ISO 10993-2:2006	ISO 10993-2:2006

NOTE 2 This part of EN ISO 10993 refers to ISO 10993-1 which itself refers to ISO 14971. In Europe, it should be assumed that the reference to ISO 14971 is to EN ISO 14971:2012.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10993-11:2017 has been approved by CEN as EN ISO 10993-11:2018 without any modification.

This document is a preview generated by EVS

Annex ZA (informative)

Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered

This European Standard has been prepared under a Commission's joint standardization request M/BC/CEN/89/9 concerning harmonized standards relating to horizontal aspects in the field of medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
7.1 (First and second indent)	4, 5 and 6	<p>ER 7.1 is only partly covered by ISO 10993-11, since the standard does not provide requirements on design and manufacture. However, this part of ISO 10993 specifies test methods for the assessment of systemic toxicity of materials intended for use in medical devices. Therefore, this standard provides a means to evaluate systemic toxicity risks associated with the materials which are used.</p> <p>These tests are not intended to evaluate or determine the performance of the test sample in terms of mechanical or functional loading.</p> <p>Systemic toxicity studies conducted by</p>

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
		<p>implantation may satisfy the requirements of this part of ISO 10993. When conducting combined studies for evaluating local effects and systemic effects, the requirements of this part of ISO 10993 and ISO 10993-6 shall be fulfilled.</p> <p>For ER 7.1 (first and second indent), flammability is not covered</p>
7.2	4, 5 and 6	ER 7.2 is not covered by ISO 10993-11, since the standard does not provide requirements on design, manufacture and packaging and does not oblige to minimize risk. However, this part of ISO 10993 specifies test methods for the assessment of systemic effects arising from the exposure of users or patients to contaminants or residues present in medical devices. This assessment can be a preliminary step for risk minimization. However it does not address risks to persons involved in the transport or storage of medical devices.
7.5, first paragraph, first sentence only	4, 5 and 6	ER 7.5 is not covered by ISO 10993-11, since the standard does not provide requirements on design and manufacture and does not oblige to minimize risk. However, this part of ISO 10993 specifies test methods for the assessment of systemic effects arising from exposure to substances released by or leaching from medical devices. This evaluation can be a preliminary step for risk minimization. Other forms of toxicity are not dealt with in this standard.

General Note: Presumption of conformity depends on also complying with all relevant clauses/subclauses of ISO 10993-1.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

Annex ZB (informative)

Relationship between this European Standard and the essential requirements of Directive 90/385/EEC [OJ L 189] aimed to be covered

This European Standard has been prepared under a Commission's joint standardization request M/BC/CEN/89/9 concerning harmonized standards relating to horizontal aspects in the field of medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 90/385/EEC of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices [OJ L 189].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZB.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 90/385/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 4, 5, 8, 9 and 10 of the Directive.

NOTE 3 This Annex ZB is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZB.1, it means that it is not addressed by this European Standard.

Table ZB.1 — Correspondence between this European Standard and Annex I of Directive 90/385/EEC [OJ L 189]

Essential Requirements of Directive 90/385/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
9 (only first and second indent)	4, 5 and 6	<p>ER 9 is only partly covered by ISO 10993-11, since the standard does not provide requirements on design and manufacture. However, this part of ISO 10993 specifies test methods for the assessment of systemic toxicity of materials intended for use in medical devices. Therefore, this standard provides a means to evaluate systemic toxicity risks associated with the materials which are used.</p> <p>These tests are not intended to evaluate or determine the</p>

Essential Requirements of Directive 90/385/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
		<p>performance of the test sample in terms of mechanical or functional loading.</p> <p>Systemic toxicity studies conducted by implantation may satisfy the requirements of this part of ISO 10993. When conducting combined studies for evaluating local effects and systemic effects, the requirements of this part of ISO 10993 and ISO 10993-6 shall be fulfilled. Other forms of toxicity are not covered.</p>

General Note: Presumption of conformity depends on also complying with all relevant clauses/subclauses of ISO 10993-1.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General considerations	3
4.1 General	3
4.2 Selection of animal species	3
4.3 Animal status	3
4.4 Animal care and husbandry	3
4.5 Size and number of groups	4
4.5.1 Size of groups	4
4.5.2 Number of groups	4
4.5.3 Treatment controls	4
4.6 Route of exposure	5
4.7 Sample preparation	5
4.8 Dosing	5
4.8.1 Test sample administration	5
4.8.2 Dosage volumes	5
4.8.3 Dosage frequency	6
4.9 Body weight and food/water consumption	6
4.10 Clinical observations	6
4.11 Clinical pathology	6
4.12 Anatomic pathology	7
4.13 Study designs	7
4.14 Quality of investigation	7
5 Acute systemic toxicity	7
5.1 General	7
5.2 Study design	8
5.2.1 Preparations	8
5.2.2 Experimental animals	8
5.2.3 Test conditions	8
5.2.4 Body weights	9
5.2.5 Clinical observations	9
5.2.6 Pathology	9
5.3 Evaluation criteria	10
5.3.1 General	10
5.3.2 Evaluation of results	10
5.4 Final report	10
6 Repeated exposure systemic toxicity (subacute, subchronic and chronic systemic toxicity)	12
6.1 General	12
6.2 Study design	12
6.2.1 Preparations	12
6.2.2 Experimental animals	12
6.2.3 Test conditions	13
6.2.4 Body weights	13
6.2.5 Clinical observations	13
6.2.6 Pathology	13
6.3 Evaluation criteria	14
6.3.1 General	14
6.3.2 Evaluation of results	14